



EA MLA Signatory
Český institut pro akreditaci, o.p.s.
(Czech Accreditation Institute)
Hájkova 2747/22, Žižkov, 130 00 Praha 3

issues

according to section 16 of Act No. 22/1997 Coll., on technical requirements for products and on changes and amendments to some Acts, as amended

CERTIFICATE OF ACCREDITATION

No. 206/2025

Foster Bohemia s.r.o.
with registered office Krátká 1148/32, Strašnice, 100 00 Praha 10
Company Registration No. 28895576

for the Testing Laboratory No. 1150
Immission Measurement Laboratory

Scope of accreditation:

Determination of the concentration and presence of inorganic fibres with the focus on asbestos in the environment and materials, including sampling; determination of environmental parameters and air pressure difference to the extent as specified in the appendix to this Certificate.

This Certificate of Accreditation is a proof of accreditation issued on the basis of assessment of fulfillment of the accreditation criteria in accordance with

ČSN EN ISO/IEC 17025:2018

In its activities performed within the scope and for the period of validity of this Certificate, the abovementioned Accredited Body is entitled to refer to this Certificate, provided that the accreditation is not suspended and the Accredited Body meets the specified accreditation requirements in accordance with the relevant regulations applicable to the activity of an accredited conformity assessment body.

This Certificate of Accreditation replaces, to the full extent, Certificate No.: 662/2022 of 27/12/2022, and/or any administrative acts building upon it.

The Certificate of Accreditation is valid until: 05/05/2030

Prague: 05/05/2025



Signed in the Czech original:
Zdeňka Drdová on 05/05/2025

Jan Velíšek
Director of the Department
of Testing and Calibration Laboratories
Czech Accreditation Institute

This translation of the Czech original has been issued by: Eliška Frycová

**The Appendix is an integral part of
Certificate of Accreditation No. 206/2025 of 05/05/2025**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Foster Bohemia s.r.o.

CAB number 1150, Immission Measurement Laboratory
Krátká 1148/32, Strašnice, 100 00 Praha 10

Tests:

Ordinal number ¹	Test procedure / method name	Test procedure / method identification ²	Tested subject	Degrees of freedom ³
1*	Determination of numerical concentration and identification of respirable inorganic fibrous particles by SEM method with EDAX analyzer	SOP 01 (ČSN EN ISO 16000-7; Guideline VDI 3492, except chap. 5, 6; ISO 14966)	Outdoor, indoor and working air, exposed filters	-
2*	Determination of presence of inorganic fibres by SEM method with EDAX analyzer	SOP 04 (Guideline VDI 3866, Part 5)	Building, inorganic and organic materials, dust accumulations	-
3*	Determination of presence of inorganic fibres by Polarized Light Microscopy - PLM	SOP 05 (Guideline VDI 3866, Part 4)	Building, inorganic and organic materials, dust accumulations	-
4*	Determination of environmental parameters (air temperature, relative humidity of air, atmospheric pressure)	PZ 01 (ČSN EN ISO 7726)	Outdoor and indoor air, working environment	-
5*	Gravimetric determination of the weight of materials	PZ 02 (User manual of the device used)	Building and other materials	-
6*	Determination of air pressure difference by a barometer(level of tightness of a protected space)	PZ 03 (User manual of the device used)	Working and non-working environment	-

¹ asterisk at the ordinal number identifies the tests, which the laboratory is qualified to carry out outside the permanent laboratory premises

² if the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest valid edition of the specified procedure is used (including any changes)

³ the laboratory does not apply a flexible approach to the scope of accreditation

**The Appendix is an integral part of
Certificate of Accreditation No. 206/2025 of 05/05/2025**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Foster Bohemia s.r.o.

CAB number 1150, Immission Measurement Laboratory
Krátká 1148/32, Strašnice, 100 00 Praha 10

Sampling:

Ordinal number	Sampling procedure name	Sampling procedure identification ¹	Subject of sampling
1	Stationary sampling for the determination of inorganic fibrous particles, manual method using a measuring filter	SOP 02 (ČSN EN ISO 16000-1; ČSN EN ISO 16000-7; ČSN EN 482; ČSN EN 689+AC; NV No. 361/2007 Coll.; Guideline VDI 3492, chap. 5, 6; VDI 4280, Part 1; ISO 14966)	Outdoor and indoor air, working environment
2	Sampling for the determination of inorganic fibrous particles (manual method)	SOP 03 (Guideline VDI 3866, Part 1, Chap. 1-4)	Building, inorganic and organic materials, dust accumulations

¹ if the document identifying the sampling procedure is dated, only these specific procedures are used. If the document identifying the sampling procedure is not dated, the latest valid edition of the specified procedure is used (including any changes)

Explanations:

SEM	Scanning Electron Microscopy
EDAX	Energy-Dispersive Analyzer of X-rays
PLM	Polarized Light Microscopy
VDI	Guideline of Association of German Engineers
PZ. SOP	internal procedure
NV	Czech Government Regulation, as amended

"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and the certificate itself."