



**EA MLA Signatory**  
**Český institut pro akreditaci, o.p.s.**  
(Czech Accreditation Institute)  
**Hájkova 2747/22, Žižkov, 130 00 Praha 3**

issues

according to section 16 of Act No. 22/1997 Coll., on technical requirements for products and on changes and amendments to some Acts, as amended

# CERTIFICATE OF ACCREDITATION

No. **622/2025**

**PRIMA BILAVČÍK, s.r.o.**  
**with registered office 9. května 1182, 688 01 Uherský Brod**  
**Company Registration No. 26227631**

for the Calibration Laboratory No. **2318**  
Calibration Laboratory

Scope of accreditation:

Calibration in the field of length, plane angle, mass, force, pressure, temperature, electrical quantities, time quantities and frequency, physicochemical quantities to the extent as specified in the appendix to this Certificate.

This Certificate of Accreditation is a proof of accreditation issued on the basis of assessment of fulfillment of the accreditation criteria in accordance with

**ČSN EN ISO/IEC 17025:2018**

In its activities performed within the scope and for the period of validity of this Certificate, the abovementioned Accredited Body is entitled to refer to this Certificate, provided that the accreditation is not suspended and the Accredited Body meets the specified accreditation requirements in accordance with the relevant regulations applicable to the activity of an accredited conformity assessment body.

This Certificate of Accreditation replaces, to the full extent, Certificate No.: 462/2024 of 10/09/2024, and/or any administrative acts building upon it.

The Certificate of Accreditation is valid until: **01/12/2026**

Prague: 04/12/2025



Signed in the Czech original:  
Gor Petrosjan on 04/12/2025

**Jan Velíšek**  
Director of the Department  
of Testing and Calibration Laboratories  
Czech Accreditation Institute

This translation of the Czech original has been issued by: Eliška Frycová

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

**Calibration laboratory locations:**

- |    |                        |   |
|----|------------------------|---|
| 1. | <b>Uherský Brod</b>    | 9. května 1182, 688 01 Uherský Brod             |
| 2. | <b>Reserved</b>        |   |
| 3. | <b>Mladá Boleslav</b>  | Komenského náměstí 90/10, 293 01 Mladá Boleslav |
| 4. | <b>Reserved</b>        |   |
| 5. | <b>Uherský Brod II</b> | Antonína Dvořáka 1274, 688 01 Uherský Brod      |

**CMC for the field of measured quantity: Length**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
1*	Slide gauges	0 mm	to 2,000 mm	resol. 0.01 mm	(15·L +15) μm	Comparison with parallel gauge blocks	KP-PB-01	1,3
		1,000 mm	to 2,000 mm		(15·L +30) μm			
2*	Micrometers	0 mm	to 1,000 mm		(15·L +1.5) μm	Comparison with parallel gauge blocks	KP-PB-02	1,3
3*	Length gauges	0 mm	to 1,040 mm		(1.2·L +0.15) μm	Direct or comparative measurement by a distance meter	KP-PB-04	1,3
4	Parallel gauge blocks - 3rd order  - 4th order	0.5 mm	to 100 mm		(1·L +0.1) μm	Comparison with parallel gauge blocks	KP-PB-07	1  1,3 3 1
		0.5 mm	to 100 mm		(2·L +0.2) μm			
		100 mm	to 500 mm		(2·L +0.2) μm			
		100 mm	to 1,000 mm		(1.2·L +0.17) μm			
5*	Height gauges	0 mm	to 1,000 mm		(1.2·L +1) μm	Comparison with parallel gauge blocks	KP-PB-12	1
6*	Graduated scales, tape measures, measuring tapes	0 mm	to 50 m		(0.2·n) mm	Direct measurement on a measuring track, comparative	KP-PB-14	1,3

**The Appendix is an integral part of  
Certificate of Accreditation No. 622/2025 of 04/12/2025**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
						measurement with reference tape measure and ruler		
7*	Length gauges	0 mm	to 300 mm		$(0.8 \cdot L + 0.07) \mu\text{m}$	Comparison with parallel gauge blocks	KP-PB-16	1
8*	Gauges of length and geometric shapes by a laser interferometer	0 m	to 20 m		$(0.8 \cdot L + 0.03) \mu\text{m}$	Direct measurement by a laser interferometer	KP-PB-19	1
9*	Gauges of length, shape, geometric positions, angles with a laser interferometer and laser tracker	0.2 m	to 15 m		$(0.3 \cdot L + 0.2) \mu\text{m}$	Direct measurement by a LaserTracker laser interferometer	KP-PB-19	1
10	Angles	0 mm	to 1,000 mm		4 $\mu\text{m}$ 6 $\mu\text{m}$	Direct measurement using a coordinate measuring machine Comparison with a perpendicularity standard and parallel gauge blocks	KP-PB-26	1,3
11*	Internal gauges	0 mm	to 300 mm		$(3 \cdot L + 2) \mu\text{m}$	Comparison with setting rings	KP-PB-20	1,3
12*	Contact, optical, multisensor coordinate measuring machines	0 mm	to 1,600 mm		$(0.6 \cdot L + 0.2) \mu\text{m}$	Comparison with steel parallels and graduated scale	KP-PB-21	1
13*	Coordinate measuring machines – manual mobile hinged arms	0 mm	to 3,700 mm		$(2 \cdot L + 20) \mu\text{m}$	Comparison using step gauges	KP-PB-21	1
14*	Coordinate measuring machines – proximity scanners	0 mm 200 mm	to 200 mm to 1,500 mm		3 $\mu\text{m}$ 20 $\mu\text{m}$	Comparison with a measuring gauge	KP-PB-21	1
15*	Coordinate measuring machines	0 m	to 30 m		50 $\mu\text{m}$	Direct measurement by a LaserTracker laser interferometer	KP-PB-21	1
16*	Coordinate measuring machines – tomographs	0 mm	to 500 mm		3 $\mu\text{m}$	Comparison with a spherical standard	KP-PB-21	1

**The Appendix is an integral part of  
Certificate of Accreditation No. 622/2025 of 04/12/2025**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
17	Surface roughness standards	0.01 μm	to 100 μm		(0.06·L) μm	Direct measurement by a roughness meter	KP-PB-23	1
18*	Roughness meters	0.01 μm	to 100 μm		(0.035·L) μm	Comparison with a roughness standard	KP-PB-24	1
19	Ring gauges	0 μm	to 2,000 μm	roundness	0.1 μm	Comparison with a reference sphere	KP-PB-31	1
				straightness	0.11 μm	Comparison with a flat glass		
20	Reserved							
21*	Profilometers	0 mm	to 600 mm		(1.0·L +1.2) μm	Comparison with a contour standard	KP-PB-25	1
22	Gauges of length and geometric quantities	0 mm	to 1,000 mm		(3·L +0.5) μm	Direct measurement by a coordinate measuring machine	KP-PB-27	1,3
					(1.2·L +0.2) μm	Direct measurement by a ring gauge		1
23*	3D length gauges	0 mm	to 3,200 mm		43 μm	Direct measurement by a mobile hinged arm	KP-PB-27	1
24	Laser distance meters	0 m	to 20 m		0.9 mm	Direct measurement by a LaserTracker laser interferometer, comparison with a length standard	KP-PB-28	1
25	Surface layer thickness gauges	0 mm	to 2.0 mm		(20·L +1.3) μm	Comparative measurement by a thickness standard	KP-PB-29	1

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

L dimension in meters

n number of two-metre sections along the whole length

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

**CMC for the field of measured quantity: Plane angle**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
1*	Angle gauges	0 °	to 360 °		3′	Comparison with angle gauges	KP-PB-15	1, 3
2	Builder's levels					Direct measurement on an index head, comparison with an angle gauge	KP-PB-36	1
	Clinometers	-3 °	to 3 °		0.7 mm/m			
3	Rotation angle of torque wrench					Direct measurement by an angle sensor	KP-PB-40	1.3
					0.15°			
		0 °	to 270 °		0.32°			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

**CMC for the field of measured quantity: Mass**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Balances with non-automatic function			0 g				Loading using a class E2 reference weight (acc. to OIML R111-1:2004)	KP-PB-71	5
		1 mg	to	200 mg		0.018 mg				
		200 mg	to	1 g		0.026 mg				
		1 g	to	2 g		0.032 mg				
		2 g	to	10 g		0.050 mg				
		10 g	to	20 g		0.064 mg				
		20 g	to	50 g		0.078 mg				
		50 g	to	100 g		0.13 mg				
100 g	to	5 kg		1.3·10 <sup>-6</sup>						
		5 kg	to	55 kg		6·10 <sup>-6</sup>	Loading using a class F1 reference weight (acc. to OIML R111-1:2004)	KP-PB-71	5	
		55 kg	to	150 kg		2.2·10 <sup>-5</sup>	Loading using a class F2 reference weight (acc. to OIML R111-1:2004)	KP-PB-71	5	
		150 kg	to	1 t		7.3·10 <sup>-5</sup>	Loading using a class M1 reference weight (acc. to OIML R111-1:2004)	KP-PB-71	5	
2	Conventional mass of weights and bodies	1 mg	to	5 mg		0.0020 mg	Comparison with a reference weight (acc. to OIML R111-1:2004) on reference scales and comparators	KP-PB-75	5	
				10 mg		0.0022 mg				
				20 mg		0.0023 mg				
				50 mg		0.0029 mg				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
				100 mg		0.0035 mg				
				200 mg		0.0043 mg				
				500 mg		0,0053 mg				
				1 g		0.0064 mg				
				2 g		0.0083 mg				
				5 g		0.011 mg				
				10 g		0.013 mg				
				20 g		0.017 mg				
				50 g		0.024 mg				
				100 g		0.032 mg				
				200 g		0.062 mg				
				500 g		0.16 mg				
				1 kg		0.32 mg				
				2 kg		0.62 mg				
				5 kg		1.6 mg				
				10 kg		14 mg				
				20 kg		24 mg				

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

<sup>4</sup> The lowest calibration uncertainty for balances is stated without accounting for the effect of the calibrated meter.

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

**CMC for the field of measured quantity: Force**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Torque / torque wrenches, torque screwdrivers	0.2 N·m		to	1,000 N·m		0.4 %	Direct measurement by a torque sensor	KP-PB-40	1, 3
2*	Force / Working force-proving instruments	0.1 N		to	200 N	Tension, pressure	0,05 %	Suspension by means of load bodies	KP-PB-41	5
		10 N		to	1,000 N		0.07 %			
		1,000 N		to	2,500 N		0.13 % FS	Direct measurement by a force sensor		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

FS full scale

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

**CMC for the field of measured quantity: Pressure**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand		Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit						
1*	Mechanical pressure gauges and electromechanical manometers	-100 kPa	to	-2 kPa	positive gauge pressure	gas	0.025 % + 10 Pa 0.1 % + 1 digit <sup>5</sup> 0.025 % + 1 Pa 0.025 % + 10 Pa 0.025 % + 0.2 kPa 0.025 % + 0.6 kPa 1.4 kPa	Comparison with a reference pressure gauge	KP-PB-55 KP-PB-56	1	
		-2 kPa	to	2 kPa							
		2 kPa	to	10 kPa							
		10 kPa	to	100 kPa							
		0.1 MPa	to	2 MPa	positive gauge pressure	liquid	0.025 % +0.6 kPa 0.025 % +1.6 kPa 0.025 % + 6 kPa 25 kPa 250 kPa				
		2 MPa	to	6 MPa							
		6 MPa	to	60 MPa							
		60 MPa	to	100 MPa							
		100 MPa	to	250 MPa	absolute pressure	gas <sup>4</sup>	0.025 % + 60 Pa 0.025 % + 0.3 kPa 0.025 % + 0.7 kPa 1.5 kPa				
		70 kPa	to	100 kPa							
		0.1 MPa	to	2 MPa							
		2 MPa	to	6 MPa							
		6 MPa	to	7 MPa	absolute pressure	liquid <sup>4</sup>	0.025 % +0.7 kPa 0.025 % +1.7 kPa 0.025 % + 6.1 kPa				
		0.1 MPa	to	6 MPa							
		6 MPa	to	16 MPa							
		16 MPa	to	60 MPa							

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
		70 kPa		to	120 kPa	barometric pressure	0.05 kPa			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

<sup>4</sup> The resulting pressure is the sum of relative and barometric pressure. The resulting uncertainty is made up of the uncertainty of relative and barometric pressure.

<sup>5</sup> digit = last digit resolution

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

**CMC for the field of measured quantity: Temperature**

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
1*	Direct-indicating thermometers	-100 °C	to -60 °C		0.04 °C	Comparison using a reference resistance thermometer	KP-PB-51	3,5
		-60 °C	to 0 °C		0.04 °C			
		0 °C	to 50 °C		0.03 °C			
		50 °C	to 230 °C		0.04 °C			
		230 °C	to 660 °C		0.05 °C			
		660 °C	to 700 °C		1.1 °C			
		700 °C	to 1,000 °C		1.2 °C	Comparison with a reference thermocouple		
		1,000 °C	to 1,100 °C		1.2 °C			
		1,100 °C	to 1,200 °C		1.4 °C			
		1,200 °C	to 1,500 °C		2.4 °C			
		1,500 °C	to 1,600 °C		2.9 °C			
2*	Direct indicating thermometers – temperature measuring chains – externally	-100 °C	to 230 °C		0.2 °C	Comparison using a reference resistance thermometer	KP-PB-51	5
		230 °C	to 660 °C		0.3 °C			
		660 °C	to 700 °C		1.1 °C	Comparison with a reference thermocouple		
		700 °C	to 1,000 °C		2.1 °C			
		1,000 °C	to 1,100 °C		2.2 °C			
		1,100 °C	to 1,200 °C		2.3 °C			
		1,200 °C	to 1,500 °C		2.9 °C			
		1,500 °C	to 1,600 °C		3.4 °C			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workplace
		min. unit	max. unit					
3*	Temperature – contactless thermometers	-20 °C	to 0 °C		1.2 °C	Comparison with a reference black body	KP-PB-53	5
		0 °C	to 50 °C		1.3 °C			
		50 °C	to 100 °C		1.4 °C			
		100 °C	to 200 °C		1.9 °C			
		200 °C	to 300 °C		2.8 °C			
		300 °C	to 400 °C		3.4 °C			
		400 °C	to 500 °C		3.6 °C			
		500 °C	to 600 °C		3.9 °C			
		600 °C	to 800 °C		5.0 °C			
800 °C	to 1,000 °C	6.1 °C						
1,000 °C	to 1,200 °C	7.3 °C						
4*	Temperature calibrators, temperature indicators	-210 °C	to 0 °C	TC-J	0.15 °C	Direct generation of equivalent DC voltage for TC – without cold junction compensation	KP-PB-122	5
		0 °C	to 1,200 °C		0.08 °C			
		-270 °C	to 0 °C	TC-K	0.18 °C			
		0 °C	to 500 °C		0.09 °C			
		500 °C	to 1,372 °C	0.12 °C				
		-270 °C	to 0 °C	TC-T	0.17 °C			
0 °C	to 400 °C	0.05 °C						
-50 °C	to 250 °C	TC-R	0.70 °C					
250 °C	to 1,768 °C		0.24 °C					
-50 °C	to 250 °C	TC-S	0.64 °C					
250 °C	to 1,768 °C		0.24 °C					

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workplace
		min.	unit	max.	unit					
		250 °C	to	1,820 °C		TC-B	0.41 °C	Direct measurement of equivalent DC voltage for TC – without cold junction compensation	KP-PB-122	5
		-270 °C	to	0 °C		TC-N	0.27 °C			
		0 °C	to	1,300 °C			0.10 °C			
		-270 °C	to	0 °C		TC-E	0.12 °C			
		0 °C	to	1,000 °C			0.07 °C			
		-210 °C	to	0 °C		TC-J	0.08 °C			
		0 °C	to	1,200 °C			0.07 °C			
		-270 °C	to	0 °C		TC-K	0.09 °C			
		0 °C	to	500 °C			0.06 °C			
		500 °C	to	1,372 °C			0.07 °C			
		-270 °C	to	0 °C		TC-T	0.09 °C			
		0 °C	to	400 °C			0.03 °C			
-50 °C	to	250 °C		TC-R	0.32 °C					
250 °C	to	1,064 °C			0.09 °C					
1,064 °C	to	1,768 °C			0.10 °C					
-50 °C	to	250 °C		TC-S	0.29 °C					
250 °C	to	1,064 °C			0.10 °C					
1,064 °C	to	1,768 °C			0.12 °C					
250 °C	to	700 °C		TC-B	0.18 °C					
700 °C	to	1,820 °C			0.11 °C					
-270 °C	to	0 °C		TC-N	0.12 °C					
0 °C	to	600 °C			0.04 °C					
600 °C	to	1,300 °C			0.04 °C					

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workplace
		min. unit	max. unit					
		-270 °C	to 0 °C	TC-E	0.05 °C			
		0 °C	to 1,000 °C		0.03 °C			
			-100 °C	Pt 100 (3850)	0.03 °C	Direct generation of equivalent resistance for RTD	KP-PB-122	5
			0 °C		0.03 °C			
			30 °C		0.04 °C			
			60 °C		0.04 °C			
			100 °C		0.04 °C			
			200 °C		0.05 °C			
			400 °C		0.07 °C			
			800 °C		0.13 °C			
		-200 °C	to 0 °C	Pt 100 (3850)	0.15 °C			
		0 °C	to 800 °C		0.43 °C			
		-200 °C	to 260 °C	Pt 500 (3850)	0.12 °C			
		260 °C	to 800 °C		0.45 °C			
		-200 °C	to 0 °C	Pt 1000 (3850)	0.05 °C			
		0 °C	to 800 °C		0.45 °C			
		-60 °C	to 0 °C	Ni 100 (6180)	0.12 °C			
		0 °C			0.02 °C			
		0 °C	to 250 °C		0.21 °C			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
		-80 °C	to	0 °C	Ni 120 (6720)	0.08 °C				
		0 °C	to	260 °C		0.11 °C				
		-50 °C	to	0 °C	Ni 1000 (6180)	0.03 °C				
		0 °C	to	200 °C		0.15 °C				
		-200 °C	to	800 °C	Pt 100 (385)	0.03 °C	Direct measurement of equivalent resistance for RTD	KP-PB-122	5	
		-200 °C	to	800 °C	Pt 500 (385)	0.10 °C				
		-200 °C	to	800 °C	Pt 1000 (385)	0.16 °C				
		-60 °C	to	250 °C	Ni 100 (6180)	0.02 °C				
		-60 °C	to	250 °C	Ni 120 (6720)	0.02 °C				
		-60 °C	to	200 °C	Ni 1000 (6180)	0.02 °C				
		-60 °C	to	200 °C	Ni 1000 (6180)	0.02 °C				

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

**CMC for the field of measured quantity: Electrical quantities**

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	DC voltage / DC voltage sources	0 mV	to	120 mV		8.0 μV/V + 0.2 μV	Direct measurement by a reference multimeter	KP-PB-110	5	
		0.12 V	to	1.2 V		5.7 μV/V + 0.61 μV				
	1.2 V	to	12 V		5.5 μV/V + 6.1 μV					
		12 V	to	120 V		8.1 μV/V + 81 μV				
		120 V	to	1,050 V		8.1 μV/V + 1.2 mV				
		1.05 kV	to	10 kV		1.2 %	Measurement with a reference multimeter with HV probe	KP-PB-110	5	
1*	DC voltage / DC voltage meters	0 mV	to	200 mV		15 μV/V + 2.3 μV	Direct measurement on a reference calibrator or comparison with a reference multimeter	KP-PB-110	5	
		0.2 V	to	2 V		8.8 μV/V + 2.6 μV				
		2 V	to	20 V		7.5 μV/V + 25 μV				
		20 V	to	200 V		13 μV/V + 0.28 mV				
		200 V	to	1,020 V		13 μV/V + 1.8 mV				
2*	DC current / DC current sources	0 μA	to	120 μA		23 μA/A + 0.42 nA	Direct measurement by a reference multimeter	KP-PB-111	5	
		0.12 mA	to	1.2 mA		14 μA/A + 4.1 nA				
		1.2 mA	to	12 mA		16 μA/A + 41 nA				
		12 mA	to	120 mA		42 μA/A + 0.61 μA				
		0.12 A	to	1.2 A		0.020 % + 13 μA				
		1.2 A	to	10 A		0.048 % + 0.35 mA				
		10 A	to	30 A		0.064 % + 4.4 mA				

**The Appendix is an integral part of  
Certificate of Accreditation No. 622/2025 of 04/12/2025**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
		10 A	to	30 A			0.031 %	Indirect measurement using a reference shunt and multimeter.	KP-PB-111	5
		30 A	to	70 A			0.083 %			
		70 A	to	200 A			0.064 %			
		200	to	300 A			0.060 %			
		0 A	to	200 A			1.2 % +0.5 A	Direct measurement by a reference clamp multimeter or a current clamp	KP-PB-111	5
		200 A	to	2,000 A			1.0 % +4.5 A			
	DC current / DC current meters	0 µA	to	200 µA			91 µA/A + 10 nA	Direct measurement on a reference calibrator or comparison with a reference multimeter	KP-PB-111	5
		0.2 mA	to	2 mA			46 µA/A + 30 nA			
		2 mA	to	20 mA			46 µA/A + 0.21 µA			
		20 mA	to	200 mA			51 µA/A + 2.2 µA			
		0.2 A	to	2 A			87 µA/A + 32 µA			
		2 A	to	20 A			0.023 % + 0.30 mA			
		20 A	to	30 A			0.034 % + 0.94 mA			
		30 A	to	60 A			0.64 % + 0.10 A	Measurement of a current simulated by a calibrator with current coil	KP-PB-111	5
		60 A	to	300 A			0.78 % + 0.13 A			
		300 A	to	1,500 A			0.60 % + 0.42 A			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
3*	AC voltage / AC voltage sources	0 mV	to 100 mV	10 Hz to 40 Hz 40 Hz to 200 Hz 200 Hz to 2 kHz 2 kHz to 20 kHz 20 kHz to 100 kHz	0.061 % + 19 μV 0.027 % + 15 μV 0.023 % + 13 μV 0.032 % + 14 μV 0.081 % + 51 μV	Direct measurement by a reference multimeter	KP-PB-112	5
		0.1 V	to 1 V	10 Hz to 40 Hz 40 Hz to 200 Hz 200 Hz to 2 kHz 2 kHz to 20 kHz 20 kHz to 100 kHz 100 kHz to 1 MHz	0.048 % + 0.24 mV 0.023 % + 61 μV 0.018 % + 61 μV 0.030 % + 0.10 mV 0.072 % + 0.50 mV 1.2 % + 25 mV			
		1 V	to 10 V	10 Hz to 40 Hz 40 Hz to 200 Hz 200 Hz to 2 kHz 2 kHz to 20 kHz 20 kHz to 100 kHz 100 kHz to 200 kHz	0.048 % + 2.3 mV 0.021 % + 0.61 mV 0.018 % + 0.61 mV 0.030 % + 1.0 mV 0.072 % + 5.0 mV 1.2 % + 0.25 V			
		10 V	to 100 V	10 Hz to 40 Hz 40 Hz to 200 Hz 200 Hz to 2 kHz 2 kHz to 20 kHz 20 kHz to 50 kHz	0.060 % + 23 mV 0.024 % + 10 mV 0.022 % + 7.1 mV 0.033 % + 11 mV 0.11 % + 62 mV			
		100 V	to 1,000 V	10 Hz to 40 Hz 40 Hz to 200 Hz 200 Hz to 2 kHz 2 kHz to 10 kHz	0.060 % + 0.16 V 0.021 % + 0.10 V 0.022 % + 0.10 V 0.037 % + 0.11 V			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
		1 kV	to 10 kV	50 Hz	2.2 %	Measurement with a reference multimeter with HV probe	KP-PB-112	5
	AC voltage / AC voltage meters	0 mV	to 202 mV	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz 100 kHz to 500 kHz	0.071 % + 24 μV 0.018 % + 24 μV 0.020 % + 33 μV 0.13 % + 56 μV 0.35 % + 0.10 mV	Direct measurement on a reference calibrator or comparison with a reference multimeter	KP-PB-112	5
		0,202 V	to 2,02 V	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz 100 kHz to 500 kHz	0.053 % + 0.21 mV 0.019 % + 0.12 mV 0.020 % + 0.21 mV 0.067 % + 0.30 mV 0.29 % + 0.46 mV			
		2,02 V	to 20,2 V	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 20 kHz 20 kHz to 100 kHz	0.050 % + 2.0 mV 0.017 % + 1.3 mV 0.021 % + 1.8 mV 0.083 % + 3.8 mV			
		20,2 V	to 202 V	30 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 40 kHz	0.050 % + 24 mV 0.017 % + 13 mV 0.021 % + 19 mV 0.035 % + 31 mV			
		202 V	to 1020 V	30 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.060 % + 0.21 V 0.024 % + 0.08 V 0.039 % + 0.13 V			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location	
		min. unit	max. unit						
4*	AC current / AC current sources	0 μA	to	100 μA	10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 10 kHz	0.061 % + 24 nA 0.039 % + 16 nA 0.083 % + 32 nA	Direct measurement by a reference multimeter	KP-PB-113	5
		100 μA	to	1 mA	10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 10 kHz	0.058 % + 0.21 μA 0.037 % + 0.19 μA 0.082 % + 0.31 μA			
		1 mA	to	10 mA	10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 10 kHz	0.058 % + 3.1 μA 0.037 % + 1.9 μA 0.081 % + 3.1 μA			
		10 mA	to	100 mA	10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 10 kHz	0.059 % + 27 μA 0.035 % + 12 μA 0.082 % + 30 μA			
		100 mA	to	1 A	10 Hz to 40 Hz 40 Hz to 1 kHz 1 kHz to 10 kHz	0.074 % + 0.34 mA 0.054 % + 0.17 mA 0.081 % + 0.5 mA			
		1 A	to	10 A	10 Hz to 40 Hz 40 Hz to 1 kHz	0.10 % + 5.8 mA 0.091 % + 3.2 mA			
		10 A	to	30 A	10 Hz to 40 Hz 40 Hz to 1 kHz	0.093 % + 13 mA 0.082 % + 9.1 mA			
		0 A	to	4 A	30 Hz to 60 Hz	0.72 % + 0.003 A			
4 A	to	40 A	30 Hz to 60 Hz	0.71 % + 0.03 A					
40 A	to	200 A	30 Hz to 60 Hz	1.2 % + 0.5 A					
200 A	to	400 A	30 Hz to 60 Hz	1.1 % + 1.3 A					
400 A	to	1,500 A	30 Hz to 60 Hz	1.3 % + 1.6 A					
1,500 A	to	2,000 A	30 Hz to 60 Hz	1.1 % + 4.6 A					

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
	AC current / AC current meters	10 μA	to 202 μA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.20 % + 0.25 μA 0.071 % + 0.15 μA 0.96 % + 0.20 μA	Direct measurement on a reference calibrator or comparison with a reference multimeter	KP-PB-113	5
		0,202 mA	to 2,02 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.19 % + 0.31 μA 0.055 % + 0.27 μA 0.50 % + 0.35 μA			
		2,02 mA	to 20,2 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.19 % + 4.1 μA 0.048 % + 2.5 μA 0.25 % + 3.4 μA			
		20,2 mA	to 202 mA	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 10 kHz	0.19 % + 41 μA 0.049 % + 25 μA 0.50 % + 43 μA			
		0,202 A	to 2,02 A	10 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz	0.19 % + 0.36 mA 0.064 % + 0.28 mA 0.50 % + 0.45 mA			
		2,02 A	to 30 A	30 Hz to 45 Hz 45 Hz to 100 Hz 100 Hz to 1 kHz	0.19 % + 3.5 mA 0.059 % + 2.5 mA 0.50 % + 4.3 mA			
		30 A	to 60 A	30 to 60 Hz	0.43 % + 0.07 A			
		60 A	to 300 A	30 to 60 Hz	0.50 % + 0.07 A	Measurement of a current simulated by a calibrator with current coil	KP-PB-113	5
		300 A	to 1500 A	30 to 60 Hz	0.48 % + 0.08 A			
5*	DC resistance / Resistors and resistance boxes	0 Ω	to 1 Ω		20 μΩ/Ω + 6.1 μΩ	Direct measurement by a reference multimeter	KP-PB-114	5

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
		1 Ω	to 10 Ω		11 μΩ/Ω + 31 μΩ			
		10 Ω	to 100 Ω		10 μΩ/Ω + 0.10 mΩ			
		100 Ω	to 1 kΩ		9.1 μΩ/Ω + 0.81 mΩ			
		1 kΩ	to 10 kΩ		11 μΩ/Ω + 8.1 mΩ			
		10 kΩ	to 100 kΩ		12 μΩ/Ω + 81 mΩ			
		0.1 MΩ	to 1 MΩ		13 μΩ/Ω + 2 Ω			
		1 MΩ	to 10 MΩ		18 μΩ/Ω + 80 Ω			
		10 MΩ	to 90 MΩ		0.019 %			
		90 MΩ	to 900 MΩ		0.051 %			
		0.9 GΩ	to 9 GΩ		0.20 %			
		9 GΩ	to 2000 GΩ		2.5 %			
	DC resistance / DC resistance meters					Direct measurement of fixed resistance standards or direct measurement on a calibrator or comparison with a reference multimeter	KP-PB-114	5
			0 Ω		5.5 mΩ			
			0.0001 Ω		0.0073 %			
			0.001 Ω		0.0029 %			
			0.01 Ω		0.0022 %			
			0.1 Ω		0.0020 %			
			1 Ω		0.0020 %			
			10 Ω		0.0005 %			
			100 Ω		0.0004 %			
			1 kΩ		0.0039 %			
			10 kΩ		0.0019 %			
			100 kΩ		0.0031 %			
			1 MΩ		0.0046 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
			10 MΩ 100 MΩ 1 GΩ		0.018 % 0.35 % 1.0 %			
		0 Ω	to 1 kΩ		0,020 % + 50 mΩ	Direct measurement on a calibrator	KP-PB-114	5
		1 kΩ	to 10 kΩ		0,020 % + 58 mΩ			
		10 kΩ	to 100 kΩ		0.020 % + 0.30 Ω			
		0.1 MΩ	to 1 MΩ		0.020 % + 4.0 Ω			
		1 MΩ	to 10 MΩ		0.020 % +0.13 kΩ			
6*	AC resistance / AC resistance meters					Direct measurement on resistance standards	KP-PB-115	5
			10 Ω	1 kHz	0.013 %			
			10 Ω	100 kHz	0.043 %			
			50 Ω	1 kHz	0.013 %			
			100 Ω	1 kHz	0.013 %			
			50 Ω	100 kHz	0.021 %			
			100 Ω	100 kHz	0.021 %			
			1 kΩ	1 kHz	0.011 %			
			1 kΩ	100 kHz	0.021 %			
			10 kΩ	1 kHz	0.011 %			
			10 kΩ	100 kHz	0.041 %			
			100 kΩ	1 kHz	0.03 %			
			100 kΩ	100 kHz	0.12 %			
			1 MΩ	1 kHz	0.03 %			
			1 MΩ	100 kHz	0.38 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
7*	DC Power / DC Power Meters (for voltage from 1 V to 1,000 V and current from 0.5 mA to 30 A)	0.5 mW	to 30 kW		0.072 %	Direct measurement on a reference calibrator	KP-PB-123	5
	(for voltage 1 V to 1,000 V and current 30 A to 1,500 A)	30 kW	to 1,500 kW		0.74 %	Direct measurement on a reference calibrator with current coil		
8*	AC active power single phase/ AC active power meters (for voltage from 1 V to 1,000 V and current from 0.5 mA to 30 A, frequency from 40 Hz to 400 Hz), capacitive and inductive	0.5 mW	to 30 kW	cos φ 1 0.8 to 0.9 0.7 to 0.8 0.6 to 0.7 0.5 to 0.6 0.4 to 0.5 0.3 to 0.4 0.2 to 0.3 0.1 to 0.2 0.1	0.13 % 0.32 % 0.46 % 0.62 % 0.8 % 1.0 % 1.4 % 1.9 % 2.9 % 5.9 %	Direct measurement on a reference calibrator	KP-PB-123	5
	(for voltage 1 V to 1,000 V and current 30 A to 1,500 A, frequency 30 Hz to 60 Hz)	30 kW	to 1500 kW	cos φ 1	0.95 %	Direct measurement on a reference calibrator with current coil		
9	Reserved							

**The Appendix is an integral part of  
Certificate of Accreditation No. 622/2025 of 04/12/2025**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
10*	Capacity / Capacity meters					Direct measurement on a reference calibrator or direct measurement on capacity standards	KP-PB-115	5
				10 Hz to 1 kHz	0.30 %			
				1 kHz	0.30 %			
				100 kHz	1.5 %			
				10 Hz to 1 kHz	0.30 %			
				10 Hz to 1 kHz	0.30 %			
				10 Hz to 1 kHz	0.30 %			
				10 Hz to 1 kHz	0.49 %			
				10 Hz to 1 kHz	0.72 %			
11	Reserved							
12*	Inductance / Inductance meters					Direct measurement on a reference calibrator	KP-PB-115	5
				1 kHz	0.61 %			
				1 kHz	0.61 %			
				1 kHz	0.61 %			
				1 kHz	0.61 %			
				1 kHz	0.61 %			
				1 kHz	0.61 %			
				1 kHz	0.67 %			
				1 kHz	0.78 %			
				1 kHz	5.1 %			
13	Reserved							
14*	Inspection equipment / Insulation resistance					Direct measurement on a calibrator of inspection instruments or a reference resistance box	KP-PB-120	5
		0.01 MΩ	to	5 MΩ	Measuring voltage up to 5,000 V			
		5 MΩ	to	121 MΩ		0.12 %		
						0.22 %		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
 CAB number 2318, Calibration Laboratory  
 9. května 1182, 688 01 Uherský Brod

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
		122 MΩ	to 1,221 GΩ		0.52 %			
		1,22 GΩ	to 12,221 GΩ		1.2 %			
	Mains impedance at 50 Hz	Z <sub>0</sub> =0 Ω	to 2 Ω		20 mΩ	Direct measurement on a calibrator of inspection instruments		
		50 mΩ	+ Z <sub>0</sub>		2.4 mΩ			
		100 mΩ	+ Z <sub>0</sub>		3.8 mΩ			
		220 mΩ	+ Z <sub>0</sub>		6.6 mΩ			
		330 mΩ	+ Z <sub>0</sub>		9.6 mΩ			
		500 mΩ	+ Z <sub>0</sub>		12 mΩ			
		1 Ω	+ Z <sub>0</sub>		16 mΩ			
		5 Ω	+ Z <sub>0</sub>		32 mΩ			
		10 Ω	+ Z <sub>0</sub>		54 mΩ			
		100 Ω	+ Z <sub>0</sub>		0.50 Ω			
		1 kΩ	+ Z <sub>0</sub>		5.0 Ω			
	Transient resistance		0.01 Ω	50 Hz	0.60 %	Direct measurement on impedance standards		
			0.1 Ω	50 Hz	0.20 %			
			1 Ω	50 Hz	0.011 %			
	Leakage current		2 mA		1.7 %	Direct measurement on a calibrator of inspection instruments or direct measurement with a multimeter		
			4.7 mA		1.7 %			
			7.7 mA		1.7 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min. unit	max. unit					
	Tripping current of residual current circuit breakers	2 mA	to 3,000 mA		1.4 %	Direct measurement on a calibrator of inspection instruments		
	Tripping time of residual current circuit breakers	10 ms	to 400 ms		0.40 ms	Direct measurement on a calibrator of inspection instruments		
15	Reserved							
16	Oscilloscope / Vertical amplifier	2 mV	to 50 V	0 Hz	0.012 % + 20 μV	Direct measurement on an oscilloscope calibrator	KP-PB-121	5
		2 mV	to 50 V	1 kHz	0.13 % + 40 μV			
	/Time base	2 ns	to 5 s		0.010 %			
	/Limit frequency	5 MHz	to 600 MHz	50 Ω	600 mV <sub>p-p</sub>			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

**CMC for the field of measured quantity: Time and frequency quantities**

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Frequency / Frequency generators	0.001 Hz		to	20 GHz		$9.3 \cdot 10^{-8}$	Direct measurement by a reference counter	KP-PB-100	5
	Frequency / frequency meters	1 Hz		to	10 MHz	sine wave signal	$20 \cdot 10^{-6}$	Direct measurement on a reference calibrator		
2*	Time interval / Manually operated time meters	0:00:00 h:min:s		to	23:59:59 h:min:s		0.10 s	Comparison with a time interval standard	KP-PB-100	5
3*	Time base of digital stopwatch	10 Hz		to	10 MHz		$4.7 \cdot 10^{-6}$	Indirect measurement by a reference counter	KP-PB-100	5

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**PRIMA BILAVČÍK s.r.o.**  
CAB number 2318, Calibration Laboratory  
9. května 1182, 688 01 Uherský Brod

**CMC for the field of measured quantity: Physicochemical quantities**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Relative humidity	10 % RH	to	40 % RH	(10 to 60) °C	0.9 % RH	Comparison with a humidity standard	KP-PB-99	3,5	
		40 % RH	to	70 % RH		1.0 % RH				
		70 % RH	to	80 % RH		1.1 % RH				
		80 % RH	to	90 % RH		1.2 % RH				
		90 % RH	to	95 % RH		1.6 % RH				

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

---

*"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and the certificate itself. "*