

**The Appendix is an integral part of
Certificate of Accreditation No. 159/2023 of 04/04/2023**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

ALWAID s.r.o.
CAB No. 1452, Testing Laboratory
679 22 Milonice č.p. 69

Tests:

Ordinal number ¹	Test procedure/method name	Test procedure/method identification ²	Tested object	Degrees of freedom ³
1*	Test of error of indication	ČSN EN ISO 10360-2	Cartesian CMMs used for measuring linear dimensions	-
2*	Test of error of indication	ČSN EN ISO 10360-12	Articulated arms for measuring linear dimensions	-
3*	Test of error of indication	ČSN EN ISO 10360-8	Cartesian CMMs for measuring linear dimensions with optical distance sensors	-
4*	Test of error of indication	ČSN EN ISO 10360-4	Cartesian CMMs used for measuring linear dimensions in scanning mode	-
5*	Test of error of indication	ČSN EN ISO 10360-10	Laser tracker for measuring linear dimensions	-
6*	Determination of the error of single and multiple stylus contacting probing systems	ČSN EN ISO 10360-5	CMM stylus contacting probing systems	-
7*	Measurement of shape, position and dimensions on CMMs according to a CAD model or a drawing	ČSN EN ISO 5459; SW Metrolog	Industrial products and parts	-

¹ Asterisk at the ordinal number identifies the tests, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² If the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes).

³ The laboratory does not apply a flexible approach to the scope of accreditation.

Explanatory notes:

CMM Coordinate Measuring Machine

CAD Computer Aided Design

SW Software