

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Badger Meter Czech Republic s.r.o.
Calibration Laboratory
Maříkova 2082/26, 621 00 Brno

CMC for the field of measured quantity: Volume, flow rate

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max.	unit					
1	Water flow meters (water meters) and flow meters	0.010 m ³ /h	to	0.10 m ³ /h		water temperature 10 °C to 30 °C	0.15 % 0.10 %	Mass method	KP01	
	Mass and mass flow rate	10 kg/h	to	100 kg/h		water temperature 10 °C to 30 °C	0.15 % 0.10 %			
		0.10 m ³ /h	to	1,500 m ³ /h						
		0.10 t/h	to	1,500 t/h						

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).