

**The Appendix is an integral part of
Certificate of Accreditation No. 221/2022 of 12/05/2022**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Recovera Facility a.s.
Immission Measurement Laboratory
Drčkova 2798/7, 628 00 Brno

Tests:

Ordinal number¹	Test procedure/method name	Test procedure/method identification²	Tested object
1*	Determination of the concentration of suspended particles (PM ₁ , PM _{2.5} , PM ₁₀) by optoelectronic method	SOP 01 (Grimm, Palas manual)	Ambient air (immissions)
2*	Determination of sulphur dioxide (SO ₂) concentration by UV fluorescence method	SOP 02 (ČSN EN 14212)	Ambient air (immissions)
3*	Determination of nitrogen oxides (NO, NO ₂ and NO _x) concentration by chemiluminescence method	SOP 03 (ČSN EN 14211)	Ambient air (immissions)
4*	Determination carbon monoxide (CO) concentration by non-dispersive infrared spectrometry method	SOP 04 (ČSN EN 14626)	Ambient air (immissions)
5*	Determination of ozone (O ₃) concentration by UV absorption photometry	SOP 05 (ČSN EN 14625)	Ambient air (immissions)

¹ asterisk at the ordinal number identifies the tests, which the Laboratory is qualified to carry out outside the permanent laboratory premises

² if the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes)