

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

**CMC for the field of measured quantity: Length**

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min	unit	max	unit					
1	Parallel gauge blocks	0.5 mm	to	100 mm			(1 L + 0.1) µm	Mechanical comparison with a standard using a comparator	MKGD KM.2	
2	Slide gauges	0 mm	to	500 mm			(8 L + 10) µm	Comparison with parallel gauge blocks and rings	MKGD PM.2	
3	Micrometers	0 mm	to	100 mm			(1 L + 2.5) µm	Comparison with parallel gauge blocks	MKGD MM.2	
4	Deviation meters	0 mm	to	30 mm	division 0.01 mm		(1 L + 2.5) µm	Direct measurement on a calibration instrument for indicators	MKGD UM.2	
5	Cylindrical gauges	0.5 mm	to	50 mm			(1 L + 2.5) µm	Direct measurement by a passameter	MKGD VK.1	
6	Feeler gauges	0.02 mm	to	2 mm			(1 L + 2.5) µm	Direct measurement by a micrometer	MKGD LS.1	
7	Steel tape measures	0 mm	to	3,000 mm			(20 L + 100) µm	Comparison with a steel rule	MKGD SM.1	
	Steel rules	3,000 mm	to	5,000 mm			(50 L + 200) µm			
		0 mm	to	1,000 mm			(20 L + 100) µm			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min	unit	max	unit					
8	Tape measures	0 mm	to	5,000 mm			0.4 mm	Comparison with a tape measure	MKGD SM.1	
		5,000 mm	to	20,000 mm			0.6 mm			
		20,000 mm	to	30,000 mm			0.8 mm			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

L = length in metres

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**  
 CAB number 2249, MEROS Calibration Laboratory  
 Starozuberská 1453, 756 54 Zubří

**CMC for the field of measured quantity: Rotational speed**

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work -place
		min	unit	max	unit					
1*	Analogue, digital and proximity revolution counters, induction revolution counters, rpm sensors, stroboscopes	1 min <sup>-1</sup> to 100,000 min <sup>-1</sup>					1.0·10 <sup>-6</sup>	Comparison with a reference electronic speed simulator	MKF-O1	
		1 min <sup>-1</sup>	to	60 min <sup>-1</sup>			0.00020 min <sup>-1</sup>	Comparison with a reference standard – non-contact method	MKF-O1	
		60 min <sup>-1</sup>	to	100,000 min <sup>-1</sup>			1.5·10 <sup>-6</sup>	Comparison with a reference standard – contact and non-contact method	MKF-O1	
		1 min <sup>-1</sup>	to	5 min <sup>-1</sup>			0.25 %	Comparison with a reference stopwatch	MKF-O1	
		5 min <sup>-1</sup>	to	20 min <sup>-1</sup>			0.070 %			
		20 min <sup>-1</sup>	to	10,000 min <sup>-1</sup>			0.0060 %			
		1 min <sup>-1</sup>	to	60 min <sup>-1</sup>	measurement time at least 4 min		0.11 %			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

**CMC for the field of measured quantity: Pressure, mechanical stress**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work -place
		min	unit	max	unit					
1	Deformation and digital manometers, pressure transducers, differential pressure measuring chains									
2*	Deformation and digital manometers, pressure transducers, differential pressure measuring chains	0.1 MPa	to	0.35 MPa		Gauge pressure - liquid	0.00007 MPa	Comparison with a piston manometer	MKMP TL.2	
		0.35 MPa	to	3.5 MPa			0.02 %			
		3.5 MPa	to	7 MPa			0.0028 MPa			
		7 MPa	to	60 MPa			0.04 %			
		0 kPa	to	1 kPa		Gauge pressure - gas	0.1 % + 0.0015 kPa			
		1 kPa	to	10 kPa			0.04 % + 0.004 kPa			
		10 kPa	to	200 kPa			0.05 % + 0.03 kPa			
		200 kPa	to	600 kPa			0.06 % + 0.05 kPa			
		600 kPa	to	2,000 kPa			0.07 % + 0.11 kPa			
		0 kPa	to	1 kPa		Vacuum - gas	0.1 % + 1.5 Pa			
		1 kPa	to	90 kPa			0.05 % + 13 Pa			
		0 MPa	to	2 MPa		Gauge pressure - liquid	0.05 % + 0.3 kPa			
		2 MPa	to	10 MPa			0.06 % + 1 kPa			
		10 MPa	to	25 MPa			0.05 % + 5 kPa			
		25 MPa	to	60 MPa			0.06 % + 9 kPa			
		500 hPa	to	1,100 hPa		Barometric pressure	0.4 hPa			
		5 kPa	to	2,000 kPa		Absolute pressure - gas	0.05 % + 0.3 kPa			
		2,000 kPa	to	10,000 kPa		Absolute pressure - liquid	0.07 % + 1.2 kPa			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work -place
		min	unit	max	unit					
		10,000	kPa	to	25,000	kPa		0.06 % + 4 kPa		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

**CMC for the field of measured quantity: Temperature**

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place	
		min	unit	max	unit						
1	Resistance thermometers	-80 °C	to	-30 °C			0.12 °C	Comparison with a resistance temperature sensor	MKT T OT.2		
		-30 °C	to	0.01 °C			0.08 °C				
				0 °C			0.05 °C				
		0.01 °C	to	150 °C			0.07 °C				
		150 °C	to	230 °C			0.08 °C				
		230 °C	to	420 °C			0.15 °C				
	Thermocouple temperature sensors	420 °C	to	660 °C			0.30 °C				
		-80 °C	to	420 °C			0.5 °C	Comparison with a resistance temperature sensor	MKT T TE.2		
		420 °C	to	650 °C			0.7 °C				
		650 °C	to	1,100 °C			1.5 °C	Comparison with a thermoelectric temperature sensor			
	Electronic thermometers	1,100 °C	to	1,300 °C			2.0 °C				
		-80 °C	to	-20 °C			0.12 °C	Comparison with a resistance temperature sensor	MKT T ET.2		
		-20 °C	to	0.01 °C			0.08 °C				
				0 °C			0.05 °C				
		0.01 °C	to	150 °C			0.07 °C				
		150 °C	to	230 °C			0.08 °C				
		230 °C	to	420 °C			0.15 °C				
		420 °C	to	650 °C			0.3 °C				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
	Glass thermometers	650 °C	to	1,000 °C			1.4 °C	Comparison with a thermoelectric temperature sensor	MKT ET.2	
		1,000 °C	to	1,100 °C			1.5 °C			
		1,100 °C	to	1,300 °C			2.0 °C			
	Non-contact thermometers	-40 °C	to	-20 °C			0.12 °C	Comparison with a resistance temperature sensor in liquid bath.	MKT ST.2	
		-20 °C	to	200 °C			0.07 °C			
		200 °C	to	300 °C			0.15 °C			
2*	Electronic thermometers, resistance temperature sensors, thermoelectric temperature sensors, glass thermometers	-20 °C	to	300 °C			1.2 °C	Comparison with a reference standard	MKT PR.1	
		300 °C	to	1,100 °C			0.6 %			
		-80 °C	to	420 °C			0.2 °C			
		420 °C	to	1,100 °C			2.0 °C	Comparison with a thermoelectric temperature sensor		

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**  
 CAB number 2249, MEROS Calibration Laboratory  
 Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
	Temperature measuring chains	-200 °C	to	-100 °C			0.1 %	Simulation of sensor electrical input signal	MKT TX.1	
		-100 °C	to	100 °C			0.1 °C			
	Non-contact thermometers	100 °C	to	1,800 °C			0.1 %	Comparison with a reference standard	MKT PR.1	
		-20 °C	to	0 °C			2.2 °C			
		0 °C	to	300 °C			1.6 °C			
		300 °C	to	550 °C			0.5 % + 0.6 °C			
		550 °C	to	1,100 °C			0.6 % + 0.5 °C			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

**CMC for the field of measured quantity: Electrical quantities**

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place	
		min	unit	max	unit						
1*	DC Voltage / Power supplies, calibrators, voltmeters, multimeters, inspection instruments	0 mV	to	1 mV			0.50 µV	Direct measurement using a multimeter	MKE-USS1, MKE-SRP.1		
		1 mV	to	10 mV							
		10 mV	to	100 mV							
		100 mV	to	1 V							
		1 V	to	1.9 V							
		1.9 V	to	7 V							
		7 V	to	19 V							
		19 V	to	50 V							
		50 V	to	190 V							
		190 V	to	1,000 V							
				1,000 V							
		1 mV	to	10 mV			0.060 %	Direct generation with a calibrator	MKE-USS1, MKE-SRP.1		
		10 mV	to	100 mV							
		100 mV	to	220 mV							
		220 mV	to	2.2 V							
		2.2 V	to	22 V							
		22 V	to	1,000 V							

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
2*	AC Voltage / Power supplies, calibrators, voltmeters, multimeters, inspection instruments	2 mV	to	10 mV		10 Hz to 10 kHz	0.40 %	Direct measurement using a multimeter	MKE-UST.1, MKE-SRP.1	
						10 kHz to 30 kHz	0.80 %			
						30 kHz to 100 kHz	1.7 %			
		10 mV	to	20 mV		10 Hz to 100 Hz	0.080 %			
						100 Hz to 2 kHz	0.065 %			
						2 kHz to 10 kHz	0.080 %			
		20 mV	to	50 mV		10 kHz to 30 kHz	0.14 %			
						30 kHz to 100 kHz	0.32 %			
						10 Hz to 100 Hz	0.050 %			
		50 mV	to	100 mV		100 Hz to 2 kHz	0.045 %			
						2 kHz to 10 kHz	0.050 %			
						10 kHz to 30 kHz	0.090 %			
		100 mV	to	190 mV		30 kHz to 100 kHz	0.20 %			
						10 Hz to 100 Hz	0.035 %			
						100 Hz to 10 kHz	0.030 %			
						10 kHz to 30 kHz	0.060 %			
						30 kHz to 100 kHz	0.13 %			
						10 Hz to 100 Hz	0.030 %			
						100 Hz to 10 kHz	0.025 %			
						10 kHz to 30 kHz	0.050 %			
						30 kHz to 100 kHz	0.11 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
		190 mV	to	500 mV		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz	0.026 % 0.023 % 0.020 % 0.022 % 0.047 % 0.17 % 1.4 %			
		500 mV	to	1 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz	0.019 % 0.015 % 0.013 % 0.014 % 0.030 % 0.10 % 0.75 %			
		1 V	to	1.9 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz	0.016 % 0.013 % 0.010 % 0.012 % 0.026 % 0.082 % 0.55 %			
		1.9 V	to	5 V		10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz	0.026 % 0.022 % 0.019 % 0.022 % 0.045 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit					
				30 kHz to 100 kHz 100 kHz to 300 kHz	0.17 % 1.4 %			
		5 V	to	10 V	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz	0.018 % 0.015 % 0.013 % 0.017 % 0.032 % 0.11 % 0.80 %		
		10 V	to	19 V	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz 100 kHz to 300 kHz	0.016 % 0.013 % 0.010 % 0.013 % 0.027 % 0.080 % 0.55 %		
		19 V	to	50 V	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz 10 kHz to 30 kHz 30 kHz to 100 kHz	0.026 % 0.022 % 0.020 % 0.021 % 0.045 % 0.16 %		
		50 V	to	100 V	10 Hz to 40 Hz 40 Hz to 100 Hz 100 Hz to 2 kHz 2 kHz to 10 kHz	0.019 % 0.015 % 0.013 % 0.015 %		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
						10 kHz to 30 kHz	0.031 %			
						30 kHz to 100 kHz	0.11 %			
		100 V to 190 V				10 Hz to 40 Hz	0.017 %			
						40 Hz to 100 Hz	0.013 %			
						100 Hz to 2 kHz	0.011 %			
						2 kHz to 10 kHz	0.012 %			
						10 kHz to 30 kHz	0.026 %			
						30 kHz to 100 kHz	0.078 %			
		190 V to 500 V				40 Hz to 10 kHz	0.032 %			
						10 kHz to 30 kHz	0.070 %			
		500 V				40 Hz to 10 kHz	0.023 %			
						10 kHz to 30 kHz	0.055 %			
		500 V to 1,000 V				40 Hz to 10 kHz	0.040 %			
						10 kHz to 30 kHz	0.16 %			
		1 mV to 2.2 mV				10 Hz to 50 kHz	0.60 %			
						50 kHz to 100 kHz	0.80 %			
		2.2 mV to 10 mV				10 Hz to 100 kHz	0.32 %			
		10 mV to 22 mV				10 Hz to 40 Hz	0.090 %			
						40 Hz to 20 kHz	0.067 %			
						20 kHz to 50 kHz	0.082 %			
		22 mV to 100 mV				50 kHz to 100 kHz	0.13 %			
						10 Hz to 40 Hz	0.095 %			
						40 Hz to 20 kHz	0.050 %			
						20 kHz to 50 kHz	0.066 %			
						50 kHz to 100 kHz	0.15 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit					
				100 kHz to 300 kHz	0.23 %			
		100 mV	to	220 mV	10 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	0.048 % 0.019 % 0.033 % 0.076 % 0.14 %		
		220 mV	to	1 V	10 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	0.053 % 0.011 % 0.016 % 0.033 % 0.093 %		
		1 V	to	2.2 V	10 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	0.039 % 0.008 % 0.012 % 0.019 % 0.070 %		
		2.2 V	to	22 V	10 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz 100 kHz to 300 kHz	0.060 % 0.008 % 0.015 % 0.023 % 0.065 %		
		22 V	to	220 V	10 Hz to 40 Hz 40 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.055 % 0.010 % 0.015 % 0.031 %		
		220 V	to	1,000 V	50 Hz to 1 kHz	0.011 %		

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
	Peak-to-peak value of square wave AC voltage / Oscilloscopes					10 Hz to 100 kHz				
3*	AC current / Power supplies, calibrators, ammeters, multimeters, clamp meters, inspection instruments	1 mV	to	2 mV		10 Hz to 100 kHz	0.30 % + 10 µV	Direct generation with a calibrator	MKE-UST1	
		2 mV	to	100 mV			0.20 % + 10 µV			
		100 mV	to	220 V			0.15 %			
		1 nA	to	2 nA		10 Hz to 100 kHz	0.40 %	Direct measurement with a picoammeter	MKE-ISS1, MKE-SRP.1	
		2 nA	to	50 nA			0.30 %			
		50 nA	to	200 nA			0.20 %			
		200 nA	to	1 µA			0.25 %			
		1 µA	to	20 µA		10 Hz to 100 kHz	0.011 %	Direct measurement with a multimeter or indirect measurement with a shunt	MKE-ISS1, MKE-SRP.1	
		20 µA	to	20 mA			0.0040 %			
		20 mA	to	0.1 A			0.011 %			
		0.1 A	to	1 A			0.0040 %			
		1 A	to	2 A			0.0060 %			
		2 A	to	60 A			0.011 %			
		60 A	to	200 A			0.015 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
	DC current / Clamp meters, inspection instruments	200 A	to	600 A			0.040 %	Indirect measurement using a shunt and multimeter	MKE-ISS1, MKE-SRP.1	
		600 A	to	2,000 A			0.10 %			
		22 µA	to	100 µA			0.040 %	Direct generation with a calibrator	MKE-ISS1, MKE-SRP.1	
		100 µA	to	2.2 mA			0.0090 %			
		2.2 mA	to	22 mA			0.0070 %			
		22 mA	to	220 mA			0.0090 %			
		220 mA	to	2.2 A			0.016 %			
4*	AC current / Power supplies, calibrators, ammeters, multimeters, inspection instruments	200 A	to	2,000 A			0.2 %	Indirect measurement with a current coil	MKE-ISS1, MKE-SRP.1	
		10 µA	to	200 µA			40 Hz to 1 kHz			
		200 µA	to	2 mA			40 Hz to 1 kHz	Direct measurement using a multimeter	MKE-SRP.1, MKE-IST1	
		2 mA	to	20 mA			40 Hz to 1 kHz			
		20 mA	to	200 mA			40 Hz to 1 kHz			
		200 mA	to	2 A			40 Hz to 1 kHz			
		2 A	to	20 A			40 Hz to 1 kHz			

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
		200 mA	to	200 A		50 to 60 Hz	0.10 %	Indirect measurement using a current transformer and multimeter	MKE-IST1, MKE-SRP.1	
		200 A	to	2,500 A		50 Hz	0.20 %			
		22 µA	to	100 µA		40 Hz to 1 kHz	0.060 %	Direct generation with a calibrator	MKE-IST1, MKE-SRP.1	
		100 µA	to	5 mA		40 Hz to 1 kHz	0.035 %			
		5 mA	to	22 mA		40 Hz to 1 kHz	0.025 %			
		22 mA	to	220 mA		40 Hz to 1 kHz	0.028 %			
		220 mA	to	2.2 A		40 Hz to 1 kHz	0.050 %			
5*	DC power / Wattmeters, network analyzers, inspection instruments (30 V to 500 V, 0.02 A to 200 A)	1 W	to	105 kW			0.060 %	Direct measurement with multimeters or indirect measurement with a shunt and multimeter	MKE-W1	
	AC power / Wattmeters, network analyzers, inspection instruments (30 V to 500 V, 0.02 A to 200 A, 50 Hz to 60 Hz, cosφ 0.5 to 1)	1 W	to	105 kW			0.060 %	Direct measurement with an energy meter	MKE-W1	
	Electrical resistance / Multimeters, resistors, ohmeters, resistance boxes, calibrators, inspection instruments			0.1 mΩ			0.0080 %	Direct generation by resistance standards	MKE-RSS1	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
		1 mΩ		0.01 Ω			0.0020 %			
		0.01 Ω		0.1 Ω			0.0025 %			
		0.1 Ω		1 Ω			0.0025 %			
		1 Ω		10 Ω			0.0020 %	Direct generation by resistance standards	MKE-RSS1, MKE-SRP.1	
		10 Ω		100 Ω			0.0025 %			
		100 Ω		1 kΩ			0.0020 %			
		1 kΩ		10 kΩ			0.0015 %			
		10 kΩ		100 kΩ			0.0015 %			
		100 kΩ		1 MΩ			0.0020 %			
		1 MΩ		10 MΩ			0.0035 %			
		10 MΩ		100 MΩ			0.0050 %	Direct generation by resistance standards	MKE-RSS1, MKE-SRP.1	
		100 MΩ		1 GΩ			0.0050 %			
		1 GΩ		0.1 mΩ	to	0.2 mΩ	0.010 %			
		0.2 Ω		0.2 mΩ	to	0.5 mΩ	0.040 %	Indirect measurement using a shunt and multimeter	MKE-RSS1, MKE-SRP.1	
		0.5 mΩ		0.5 mΩ	to	0.9 mΩ	0.0080 %			
		0.9 mΩ		0.9 mΩ	to	5 Ω	0.0060 %			
		5 Ω		5 Ω	to	200 kΩ	0.0040 %			
		200 kΩ		200 kΩ	to	2 MΩ	0.0030 %	Direct measurement using a multimeter	MKE-RSS1, MKE-SRP.1	
		2 MΩ		2 MΩ	to	20 MΩ	0.0040 %			
		20 MΩ					0.0060 %	Direct measurement using a multimeter	MKE-RSS1, MKE-SRP.1	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
		20 MΩ	to	200 MΩ			0.040 %			
		200 MΩ	to	300 MΩ			0.070 %			
		300 MΩ	to	500 MΩ			0.050 %			
		500 MΩ	to	1 GΩ			0.035 %			
		1 GΩ	to	2 GΩ			0.020 %			
		2 GΩ	to	10 GΩ			0.65 %			
		10 GΩ	to	20 GΩ			0.20 %			
		20 GΩ	to	250 GΩ		up to 10,000 V	0.50 %	Indirect measurement using a picoammeter and HF voltmeter	MKE-RSS, MKE-SRP.1	
		250 GΩ	to	1 TΩ			0.70 %			
		1 Ω					0.010 %	Direct generation with a calibrator	MKE-RSS1, MKE-SRP.1	
		1.9 Ω					0.010 %			
		10 Ω					0.0026 %			
		19 Ω					0.0026 %			
		100 Ω					0.0014 %			
		190 Ω					0.0014 %			
		1 kΩ					0.0011 %			
		1.9 kΩ					0.0011 %			
		10 kΩ					0.0011 %			
		19 kΩ					0.0011 %			
		100 kΩ					0.0015 %			
		190 kΩ					0.0015 %			
		1 MΩ					0.0026 %			
		1.9 MΩ					0.0026 %			

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**

CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
		10 MΩ		19 MΩ			0.0050 %			
		100 MΩ					0.0050 %			
							0.015 %			
7*	Impedance module / Clamp meters, inspection instruments	0.1 mΩ	to	1 mΩ		50 Hz to 60 Hz	0.20 %	Indirect measurement using a shunt and multimeter	MKE-Z1, MKE-SRP.1	
		1 mΩ	to	0.1 Ω		50 Hz to 400 Hz	0.20 %			
		0.1 Ω	to	10 kΩ		50 Hz to 1 kHz	0.15 %			
		10 kΩ	to	10 MΩ		50 Hz to 60 Hz	0.30 %			
8*	Impedance / RLC bridges, resistance standards	10 mΩ		100 Hz to 1 kHz 1 kHz to 100 kHz 100 kHz to 500 kHz 500 kHz to 1 MHz		100 Hz to 1 kHz 1 kHz to 100 kHz 100 kHz to 500 kHz 500 kHz to 1 MHz	1.6 % 5.0 % 7.0 % 9.0 %	Direct measurement on impedance standards	MKE-Z2	
		100 mΩ		100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz		100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 300 kHz 300 kHz to 500 kHz 500 kHz to 1 MHz	0.16 % 0.60 % 0.80 % 1.2 % 2.3 %	Direct measurement on impedance standards	MKE-Z2	
		1 Ω		100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 1 MHz		100 Hz 100 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 1 MHz	0.030 % 0.070 % 0.12 % 0.25 %	Direct measurement on impedance standards	MKE-Z2	
		10 Ω		100 Hz 100 Hz to 1 kHz		100 Hz 100 Hz to 1 kHz	0.015 % 0.060 %	Direct measurement on impedance standards	MKE-Z2	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
						1 kHz to 10 kHz	0.12 %			
						10 kHz to 1 MHz	0.25 %			
		100 Ω		100 Hz		100 Hz	0.015 %			
				100 Hz to 1 kHz		100 Hz to 1 kHz	0.025 %			
				1 kHz to 10 kHz		1 kHz to 10 kHz	0.060 %			
				10 kHz to 300 kHz		10 kHz to 300 kHz	0.15 %			
				300 kHz to 1 MHz		300 kHz to 1 MHz	0.60 %			
		1 kΩ		100 Hz		100 Hz	0.015 %			
				100 Hz to 1 kHz		100 Hz to 1 kHz	0.025 %			
				1 kHz to 10 kHz		1 kHz to 10 kHz	0.060 %			
				10 kHz to 300 kHz		10 kHz to 300 kHz	0.15 %			
				300 kHz to 1 MHz		300 kHz to 1 MHz	0.60 %			
		10 kΩ		100 Hz		100 Hz	0.015 %			
				100 Hz to 1 kHz		100 Hz to 1 kHz	0.025 %			
				1 kHz to 10 kHz		1 kHz to 10 kHz	0.060 %			
				10 kHz to 300 kHz		10 kHz to 300 kHz	0.15 %			
				300 kHz to 1 MHz		300 kHz to 1 MHz	0.60 %			
		100 kΩ		100 Hz to 1 kHz		100 Hz to 1 kHz	0.025 %			
				1 kHz to 100 kHz		1 kHz to 100 kHz	0.12 %			
				100 kHz to 300 kHz		100 kHz to 300 kHz	0.40 %			
				300 kHz to 1 MHz		300 kHz to 1 MHz	0.60 %			
		1 MΩ		100 Hz		100 Hz	0.060 %			
				100 Hz to 10 kHz		100 Hz to 10 kHz	0.15 %			
				10 kHz to 100 kHz		10 kHz to 100 kHz	0.60 %			
		10 MΩ		100 Hz		100 Hz	0.060 %			
				100 Hz to 10 kHz		100 Hz to 10 kHz	0.15 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
Capacity / RLC bridges, capacity standards				10 kHz to 100 kHz		0.60 %				
		100 MΩ		1 kHz		0.50 %				
	10 pF			1 kHz		0.015 %				
				1 kHz to 10 kHz		0.12 %				
				10 kHz to 100 kHz		0.25 %				
				100 kHz to 1 MHz		0.60 %				
				100 Hz to 1 kHz		0.30 %				
	100 pF			1 kHz		0.015 %				
				1 kHz to 10 kHz		0.060 %				
				10 kHz to 100 kHz		0.15 %				
				100 kHz to 1 MHz		0.25 %				
				100 Hz to 1 kHz		0.12 %				
	1 nF			1 kHz		0.015 %				
				1 kHz to 100 kHz		0.060 %				
				100 kHz to 1 MHz		0.15 %				
				100 Hz to 1 kHz		0.060 %				
	10 nF			1 kHz		0.015 %				
				1 kHz to 100 kHz		0.060 %				
				100 kHz to 1 MHz		0.15 %				
				100 Hz to 1 kHz		0.060 %				
	100 nF			1 kHz		0.015 %				
				1 kHz to 10 kHz		0.060 %				
				10 kHz to 1 MHz		0.15 %				
				100 Hz to 1 kHz		0.060 %				
	1 μF			1 kHz		0.015 %				
				1 kHz to 10 kHz		0.060 %				
				10 kHz to 1 MHz		0.15 %				
				100 Hz to 1 kHz		0.070 %				
				1 kHz		0.030 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
Inductance / RLC bridges, inductance standards	10 µF			1 kHz to 10 kHz		0.070 %				
				10 kHz to 100 kHz		0.15 %				
				100 Hz to 1 kHz		0.15 %				
				1 kHz to 10 kHz		0.25 %				
	100 µF			100 Hz		0.15 %				
				100 Hz to 1 kHz		0.25 %				
	1,000 µF			1 kHz to 10 kHz		0.70 %				
				100 Hz		0.70 %				
				100 Hz to 1 kHz		1.2 %				
	Impedance / Resistance	10 µH		100 Hz		1.2 %				
				1 kHz		0.40 %				
				10 kHz		0.15 %				
		100 µH		100 Hz		0.60 %				
				1 kHz		0.15 %				
				10 kHz		0.15 %				
		1 mH		100 Hz		0.25 %				
				1 kHz		0.12 %				
				10 kHz		0.10 %				
		2 mH 10 mH 100 mH 1 H 10 H		1 kHz		0.070 %				
				1 kHz		0.050 %				
				1 kHz		0.050 %				
				1 kHz		0.050 %				
				1 kHz		0.050 %				

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
standards, resistance boxes, resistance meters						20 kHz	1.0 %	using a RLC bridge		
		0.2 Ω	to	2 Ω		100 Hz to 10 kHz	0.30 %			
						20 kHz	0.50 %			
		2 Ω	to	20 kΩ		100 Hz to 10 kHz	0.10 %			
						20 kHz	0.20 %			
		20 kΩ	to	200 kΩ		100 Hz to 10 kHz	0.20 %			
						20 kHz	0.50 %			
		0.2 MΩ	to	2 MΩ		100 Hz to 20 kHz	1.0 %			
		2 MΩ	to	20 MΩ		100 Hz to 4 kHz	2.0 %			
						10 kHz	3.0 %			
						20 kHz	5.0 %			
Capacity / Capacity standards, capacity decades, capacity meters						10 kHz	0.30 %			
		2 pF	to	20 pF		20 kHz	1.5 %			
						1 kHz to 4 kHz	0.30 %			
		20 pF	to	200 pF		10 kHz	0.20 %			
						20 kHz	1.0 %			
		0.2 nF	to	2 nF		100 Hz to 400 Hz	0.30 %			
						1 kHz to 4 kHz	0.20 %			
		0.2 nF	to	2 nF		10 kHz	0.10 %			
						20 kHz	0.50 %			
		2 nF	to	20 nF		100 Hz to 400 Hz	0.20 %			
						1 kHz to 10 kHz	0.10 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
		20 nF	to	200 nF		20 kHz	0.50 %			
						100 Hz to 10 kHz	0.10 %			
		0.2 µF	to	2 µF		20 kHz	0.50 %			
						100 Hz to 4 kHz	0.10 %			
						10 kHz	0.50 %			
		2 µF	to	20 µF		20 kHz	1.0 %			
						100 Hz to 400 Hz	0.10 %			
						1 kHz to 4 kHz	0.50 %			
						10 kHz	1.0 %			
						20 kHz	2.0 %			
		20 µF	to	200 µF		100 Hz to 400 Hz	0.50 %			
						1 kHz to 4 kHz	1.0 %			
						10 kHz	2.0 %			
						20 kHz	5.0 %			
		0.2 mF	to	2 mF		100 Hz to 400 Hz	1.0 %			
						1 kHz to 4 kHz	2.0 %			
						10 kHz	5.0 %			
						20 kHz	10 %			
	Inductance / Inductance standards, inductance decades, inductance meters	0.2 µH	to	2 µH		10 kHz	1.0 %			
						20 kHz	2.0 %			
		2 µH	to	20 µH		1 kHz to 4 kHz	1.0 %			
						10 kHz	0.50 %			
						20 kHz	1.0 %			

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
		20 µH	to	200 µH		100 Hz to 400 Hz 1 kHz to 20 kHz	1.0 % 0.50 %			
		0.2 mH	to	2 mH		100 Hz to 400 Hz 1 kHz to 10 kHz 20 kHz	0.50 % 0.10 % 0.50 %			
		2 mH	to	20 mH		100 Hz to 10 kHz 20 kHz	0.10 % 0.50 %			
		20 mH	to	200 mH		100 Hz to 4 kHz 10 kHz 20 kHz	0.10 % 0.50 % 1.0 %			
		0.2 H	to	2 H		100 Hz to 400 Hz 1 kHz to 4 kHz 10 kHz 20 kHz	0.10 % 0.50 % 2.0 % 5.0 %			
		2 H	to	20 H		100 Hz to 400 Hz 1 kHz to 4 kHz 10 kHz 20 kHz	0.50 % 2.0 % 5.0 % 10 %			
9*	HF voltage peak-to-peak value / HF voltage meters, oscilloscopes, HF millivoltmeters, frequency analyzers, function generators, signal generators	5 mV	to	5 V		100 kHz to 300 MHz 300 MHz to 550 MHz	3.7 % 4.2 %	Direct generation in the plane of the connecting BNC connector of 50 Ω	MKE-UVF1	

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
HF voltage effective value / HF voltage meters, oscilloscopes, HF millivoltmeters, frequency analyzers, function generators, signal generators		5 mV	to	3 V		550 MHz to 1.1 GHz	5.2 %			
		5 mV	to	2 V		1.1 GHz to 2.5 GHz	5.9 %			
						2.5 GHz to 3.2 GHz	5.9 %			
		1 mV	to	2 mV		100 kHz to 100 MHz	3.6 %			
		2 mV	to	10 mV			2.8 %			
		10 mV	to	10 V			2.5 %			
		1 mV	to	2 mV		100 MHz to 200 MHz	4.2 %			
		2 mV	to	10 mV			3.4 %			
		10 mV	to	1 V			3.0 %			
		1 V	to	10 V			4.2 %			
		1 mV	to	2 mV		200 MHz to 500 MHz	4.5 %			
		2 mV	to	10 mV			3.7 %			
		10 mV	to	1 V			3.3 %			
		1 V	to	10 V			4.9 %			
10*	High AC voltage / Meters, inspection instruments	1 kV	to	48 kV		50 Hz	0.060 %	Direct generation, comparison with a HV transformer	MKE-UVNST1, MKE-SRP.1	
	High AC voltage / Power supplies, inspection instruments	1 kV	to	48 kV		50 Hz	0.060 %	Direct measurement with a HV probe	MKE-UVNST1, MKE-SRP.1	
		48 kV	to	50 kV			1.2 %			

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
		50 kV	to	70 kV			1.3 %			
11*	High DC voltage / Power supplies, inspection instruments	70 kV	to	100 kV			1.2 %			
	High DC voltage / Meters, inspection instruments	1 kV	to	100 kV			0.20 %	Direct measurement with a HV probe	MKE-UVNSS1, MKE-SRP.1	
		1 kV	to	80 kV			0.20 %	Direct generation, comparison with a HV probe	MKE-UVNSS1, MKE-SRP.1	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

**CMC for the field of measured quantity: Light quantities**

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
1	Illumination / Luxmeters	1 Lx	to	30,000 Lx			2.5 %	Comparison with a luxmeter	MK-LUX1	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

**CMC for the field of measured quantity: Time and frequency quantities**

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place				
		min	unit	max	unit									
1*	Frequency / Frequency meters, counters, frequency analyzers, function generators, signal generators, frequency references, reference oscillators, multimeters, calibrators	0.001 Hz	to	50 MHz		$U \geq 2.2 \text{ V TTL} / 50 \Omega$	$3.5 \cdot 10^{-11}$	Direct generation by a GPS receiver or generator	MKF-f1					
			50 MHz	to	1 GHz									
			1 GHz	to	3.2 GHz									
		1 MHz	$U_{vst} \geq 1 \text{ V}_{ef}$ , signal to noise ratio $\geq 60 \text{ dB}$	$\tau$ from 100 s to 200 s	$6.0 \cdot 10^{-12}$	Indirect measurement with a GPS receiver and counter	MKF-f1							
			5 MHz	$U_{vst} \geq 1 \text{ V}_{ef}$ , signal to noise ratio $\geq 60 \text{ dB}$	$\tau$ from 100 s to 200 s	$4.0 \cdot 10^{-11}$								
			10 MHz											

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
		0.001 Hz	to	100 Hz		pulse signals	5·10 <sup>-10</sup>	Direct measurement by a counter	MKF-f1, MKE-SRP.1	
		100 Hz	to	350 MHz		τ over 10 s τ over 1 s	5·10 <sup>-10</sup>			
		1 Hz	to	30 Hz		U <sub>vst</sub> ≥ 1 V <sub>ef</sub> , signal to noise ratio ≥ 60 dB	5·10 <sup>-5</sup> Hz			
		30 Hz	to	300 Hz		τ over 10 s	3·10 <sup>-4</sup> Hz			
		300 Hz	to	100 kHz		τ over 1 s	5·10 <sup>-5</sup> Hz			
		100 kHz	to	300 kHz			6·10 <sup>-10</sup>	Direct measurement by a counter	MKF-f1	
		300 kHz	to	350 MHz			5·10 <sup>-10</sup>			
		350 MHz	to	6 GHz			3·10 <sup>-10</sup>			
		1 ns	to	4 ns		U <sub>ss</sub> ≥ 1 V	50 Ω	0.70 ns 4.5% +0.5 ns 5.5% +0.5 ns	Direct generation with a calibrator	MKF-t1
		4 ns	to	10 ns						
		10 ns	to	100 ns						
		100 ns	to	1 μs		U ≥ 2.2 V TTL	50 Ω	2.0 ns 4.0 ns 3.5·10 <sup>-11</sup>	Direct generation by a GPS receiver	
		1 μs	to	100 s						
		100 s	to	10 <sup>5</sup> s						
		2.85 ns	to	10 ms		U <sub>inp</sub> ≥ 1 V	τ over 10 s	5·10 <sup>-10</sup>	Direct measurement by a counter	MKF-t1
Pulse signal period / Time meters, counters, oscilloscopes, function generators, pulse										

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**  
 CAB number 2249, MEROS Calibration Laboratory  
 Starozuberská 1453, 756 54 Zubří

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
	generators, digital and mechanical stopwatches, timers, inspection instruments	10 ms		1,00	s	τ over 1 s	5·10 <sup>-10</sup>			
	Pulse signal duration / Time meters, counters, oscilloscopes, function generators, pulse generators, digital and mechanical stopwatches, timers, inspection instruments	5 ns	to	10 s		$U_{inp} \geq 1 \text{ V}$	1.0 ns	Direct measurement by a counter	MKF-t1, MKE-SRP.1	
		10 s	to	$10^5$ s			5·10 <sup>-10</sup>			
		$10^5$ s	to	$4 \cdot 10^5$ s			3.6·10 <sup>-8</sup>	Direct comparison with a calibrator	MKF-t1	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**MEROS, spol. s r.o.**  
CAB number 2249, MEROS Calibration Laboratory  
Starozuberská 1453, 756 54 Zubří

**CMC for the field of measured quantity: Humidity**

Ord. nr. <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
1*	Relative humidity / Humidity meters	5 % RH	to	10 % RH		Temperature (15 to 70) °C	2.4 % RH	Comparison with a humidity meter in a climatic chamber	MKRV.1	
		10 % RH	to	30 % RH		Temperature (15 to 70) °C	1.2 % RH			
		30 % RH	to	70 % RH		Temperature (15 to 70) °C	1.3 % RH			
		70 % RH	to	90 % RH		Temperature (15 to 70) °C	1.4 % RH			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).