

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

BMT Medical Technology s.r.o.
 CAB number 2386, Calibration Laboratory
 Cejl 157/50, Zábřovice, 602 00 Brno

CMC for the field of measured quantity: Pressure

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
1*	Electromechanical pressure gauges and pressure measuring chains	0 kPa		to	700 kPa	absolute pressure	0.3 kPa	Comparison with a standard digital manometer	QPP 11-16 QPP 11-19 QPP 11-23	
		-100 kPa		to	600 kPa	positive gauge pressure	0.3 kPa			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

BMT Medical Technology s.r.o.
 CAB number 2386, Calibration Laboratory
 Cejl 157/50, Zábřovice, 602 00 Brno

CMC for the field of measured quantity: Temperature

Ord. number ₁	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
1*	Electronic thermometers and temperature measuring chains	0 °C	to	180 °C		0.2 °C	Comparison with a standard digital thermometer	QPP 11-16 QPP 11-18		
		180 °C	to	300 °C		0.4 °C				

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).