

**Appendix is an integral part of
Certificate of Accreditation No.: 24/2022 of 13/01/2022**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

České vysoké učení technické v Praze
Laboratory for the Photovoltaic Systems Diagnostics (LPVSD)
Technická 2, 166 27 Praha 6 – Dejvice

The Laboratory provides expert opinions and interprets test results.

Tests:

Ordinal number¹	Test procedure/method name	Test procedure/method identification²	Tested object
1	Measurement of Volt-Ampere Characteristic	BZP_01_A (ČSN EN 60904-1 ed. 2, p. 7; ČSN EN IEC 61215-2 ed. 2, p. 4.2, 4.6, 4.7; ČSN EN 61215-2:2017, p. 4.2, 4.6, 4.7; ČSN EN 61215:1997, p. 10.2, 10.6, 10.7; ČSN EN 61215 ed. 2:2006, p. 10.2, 10.6, 10.7; ČSN EN 61646:1999, p. 10.2, 10.6, 10.7; ČSN EN 61646 ed. 2:2009, p. 10.2, 10.6, 10.7)	Photovoltaic Modules
2	Dielectric withstand test	BZP_03_A (ČSN EN IEC 61730-2 ed. 2, p. 10.13; ČSN EN IEC 61215-2 ed. 2, p. 4.3; ČSN EN 61215-2:2017, p. 4.3; ČSN EN 61215:1997, p. 10.3; ČSN EN 61215 ed. 2:2006, p. 10.3; ČSN EN 61646:1999, p. 10.3; ČSN EN 61646 ed. 2:2009, p. 10.3)	Photovoltaic Modules
3	Insulation test by applied voltage below 500 V	BZP_04_A (ČSN EN IEC 61215-2 ed. 2, p. 4.3; ČSN EN 61215-2:2017, p. 4.3; ČSN EN 61215:1997, p. 10.3; ČSN EN 61215 ed. 2:2006, p. 10.3; ČSN EN 61646:1999, p. 10.3; ČSN EN 61646 ed. 2:2009, p. 10.3)	Photovoltaic Modules

**Appendix is an integral part of
Certificate of Accreditation No.: 24/2022 of 13/01/2022**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

České vysoké učení technické v Praze
Laboratory for the Photovoltaic Systems Diagnostics (LPVSD)
Technická 2, 166 27 Praha 6 – Dejvice

Ordinal number¹	Test procedure/method name	Test procedure/method identification²	Tested object
4	Continuity test of equipotential bonding by applied current below 30 A	BZP_05_A (ČSN EN IEC 61730-2 ed. 2, p. 10.11)	Photovoltaic Modules
5*	Visual Inspection	BZP_06_A (ČSN EN IEC 61215-2 ed. 2, p. 4.1; ČSN EN 61215-2:2017, p. 4.1; ČSN EN 61215:1997, p. 10.1; ČSN EN 61215 ed. 2:2006, p. 10.1; ČSN EN 61646:1999, p. 10.1; ČSN EN 61646 ed. 2:2009, p. 10.1)	Photovoltaic Modules

¹ Asterisk at the ordinal number identifies the tests, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² If the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Explanations:

BZP_XX_A – LDFS Test Procedure