

**The Appendix is an integral part of
Certificate of Accreditation No. 264/2023 of 31/05/2023**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Circulos Group s.r.o.
CAB number 1593, TechLab Testing Laboratory
Služeb 3056/4, Strašnice, 108 00 Praha 10

Tests:

Ordinal number ¹	Test procedure / method name	Test procedure / method identification ²	Subject of the test	Degrees of freedom ³
1*	Surface temperature measurement with contact thermometers	SOP-8-01 (ČSN EN ISO 7726, Appendix F)	Instruments and equipment	-
2*	Measurement of temperature by MTAS system	SOP-8-02 (Manual GENO-tronics MTAS)	Instruments and equipment	-
3*	Measurement of environmental temperature	SOP-8-03 (ČSN EN ISO 17665-1; ČSN EN ISO 7726)	Instruments and equipment	-
4*	Measurement of pressure	SOP-8-04 (ČSN EN 285)	Instruments and equipment	-
5*	Enumeration of particles in air	SOP-8-05 (ČSN EN ISO 14644-1)	Instruments and equipment	-
6*	Determination of mean air flow velocity	SOP-8-06 (ČSN EN 12469)	Instruments and equipment	-
7*	Measurement of CO ₂ concentration by NDIR method	SOP-8-07 (Manual Vaisala CARBOCAP Hand-Held, Carbon Dioxide Metr GM70)	Instruments and equipment	-

¹ asterisk at the ordinal number identifies the tests, which the Laboratory is qualified to carry out outside the permanent laboratory premises

² if the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes)

³ the laboratory does not apply a flexible approach to the scope of accreditation

Explanations and abbreviations:

MTAS - Mobile Temperature Acquisition System

NDIR - Nondispersive Infrared Spectrometry

SOP - Standard Operating Procedure (Internal Test Procedure of the TechLab Testing Laboratory)