

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Vysoké učení technické v Brně  
CAB number 2395, CVVOZE Calibration Laboratory  
Technická 3082/12, 616 00 Brno

**CMC for the field of measured quantity: Mechanical motion**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
1*	Acceleration of linear harmonic mechanical vibrations / Vibrometers, seismometers, linear vibration generators, vibration testing systems <sup>4,5</sup>	0.01 m·s <sup>-2</sup>	to	1,000 m·s <sup>-2</sup>		0.2 Hz to 0.4 Hz 0.4 to 1 Hz 1 Hz to 1 kHz 1 to 5 kHz 5 to 10 kHz 10 to 15 kHz 15 to 20 kHz	1.5 % 1.0 % 0.5 % 0.7 % 1.5 % 2.0 % 3.0 %	Comparison with a reference sensor	KP-01 (ČSN ISO 16063-21), ČSN ISO 16063-44)	
2	Acceleration of linear harmonic mechanical vibrations / Vibrometers, seismometers, vibration transducer calibration systems, reference linear vibration generators <sup>4,5</sup>	0.01 m·s <sup>-2</sup>	to	1,000 m·s <sup>-2</sup>		0.2 Hz to 0.4 Hz 0.4 Hz to 1 Hz 1 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz 10 kHz to 15 kHz 15 kHz to 20 kHz	0.7 % 0.5 % 0.3 % 0.5 % 1.0 % 2.0 % 2.5 %	Direct measurement with a laser vibrometer	KP-01 (ČSN ISO 16063-11), ČSN ISO 16063-44)	

The Appendix is an integral part of  
Certificate of Accreditation No. 269/2023 of 31/05/2023

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Vysoké učení technické v Brně**  
CAB number 2395, CVVOZE Calibration Laboratory  
Technická 3082/12, 616 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
3	Phase shift of acceleration of linear harmonic mechanical vibrations / Vibrometers, seismometers, linear vibration generators, vibration testing systems <sup>4</sup>	0 °	to	360 °		0.2 Hz to 1 Hz 1 Hz to 5 kHz 5 kHz to 10 kHz 10 kHz to 15 kHz 15 kHz to 20 kHz	1.5 ° 0.7 ° 1.0 ° 2.0 ° 3.0 °	Comparison with a reference sensor	KP-01 (ČSN ISO 16063-21)	
4	Phase shift of acceleration of linear harmonic mechanical vibrations / Vibrometers, seismometers, vibration transducer calibration systems, reference linear vibration generators <sup>4</sup>	0 °	to	360 °		0.2 Hz to 0.4 Hz 0.4 Hz to 5 kHz 5 kHz to 10 kHz 10 kHz to 15 kHz 15 kHz to 20 kHz	0.7 ° 0.5 ° 1.0 ° 2.0 ° 3.0 °	Direct measurement with a laser vibrometer	KP-01 (ČSN ISO 16063-11)	
5*	Sensitivity / Vibration transducers, geophones, vibrometers and vibration measurement chains with an electric output <sup>4,5</sup>	0.01 mV/m·s <sup>-2</sup>	to	10,000 mV/m·s <sup>-2</sup>				Comparison with a reference sensor	KP-01 (ČSN ISO 16063-21)	
		0.01 pC/m·s <sup>-2</sup>	to	1,000 pC/m·s <sup>-2</sup>						

The Appendix is an integral part of  
Certificate of Accreditation No. 269/2023 of 31/05/2023

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Vysoké učení technické v Brně**  
CAB number 2395, CVVOZE Calibration Laboratory  
Technická 3082/12, 616 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
		0.1 V/m·s <sup>-1</sup>	to	10,000 V/m·s <sup>-1</sup>		0.2 Hz to 0.4 Hz	1.5 %			
		0.01 V/m	to	10,000 V/m		0.4 Hz to 1 Hz	1.0 %			
						1 Hz to 1 kHz	0.5 %			
						1 kHz to 5 kHz	0.7 %			
						5 kHz to 10 kHz	1.5 %			
						10 kHz to 15 kHz	2.0 %			
						15 kHz to 20 kHz	3.0 %			
6	Sensitivity / Vibration transducers, standard vibration transducers, geophones, vibrometers and vibration measurement chains with an electric output <sup>4,5</sup>	0.01 mV/m·s <sup>-2</sup>	to	10,000 mV/m·s <sup>-2</sup>				Direct measurement with a laser vibrometer	KP-01 (ČSN ISO 16063-11)	
		0.01 pC/m·s <sup>-2</sup>	to	1,000 pC/m·s <sup>-2</sup>						
		0.1 V/m·s <sup>-1</sup>	to	10,000 V/m·s <sup>-1</sup>		0.2 Hz to 0.4 Hz	0.7 %			
		0.01 V/m	to	10,000 V/m		0.4 Hz to 1 Hz	0.5 %			
						1 Hz to 1 kHz	0.3 %			
						1 kHz to 5 kHz	0.5 %			
						5 kHz to 10 kHz	1.0 %			
						10 kHz to 15 kHz	2.0 %			
						15 kHz to 20 kHz	2.5 %			
7	Sensitivity / Laser vibrometers with an electric output <sup>4,5</sup>	0.1 V/m·s <sup>-1</sup>	to	10,000 V/m·s <sup>-1</sup>				Direct measurement with a laser vibrometer	KP-01 (ČSN ISO 16063-41)	
		0.01 V/m	to	10,000 V/m		0,2 Hz až 5 kHz	0,25 %			

The Appendix is an integral part of  
Certificate of Accreditation No. 269/2023 of 31/05/2023

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Vysoké učení technické v Brně**  
CAB number 2395, CVVOZE Calibration Laboratory  
Technická 3082/12, 616 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
						5 kHz až 10 kHz 10 kHz až 15 kHz 15 kHz až 20 kHz	0,3 % 0,5 % 0,7 %			
8	Sensitivity phase shift / Vibration transducers, geophones, vibrometers and vibration measurement chains with an electric output	0 °	to	360 °		0.2 Hz to 1 Hz 1 Hz to 5 kHz 5 kHz to 10 kHz 10 kHz to 15 kHz 15 kHz to 20 kHz	1.5 ° 0.7 ° 1.0 ° 2.0 ° 3.0 °	Comparison with a reference sensor	KP-01 (ČSN ISO 16063-21)	
9	Sensitivity phase shift / Vibration transducers, standard vibration transducers, geophones, vibrometers and vibration measurement chains with an electric output	0 °	to	360 °		0.2 Hz to 0.4 Hz 0.4 Hz to 5 kHz 5 kHz to 10 kHz 10 kHz to 15 kHz 15 kHz to 20 kHz	0.7 ° 0.5 ° 1.0 ° 2.0 ° 3.0 °	Direct measurement with a laser vibrometer	KP-01 (ČSN ISO 16063-11)	

**The Appendix is an integral part of  
Certificate of Accreditation No. 269/2023 of 31/05/2023**

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Vysoké učení technické v Brně**  
CAB number 2395, CVVOZE Calibration Laboratory  
Technická 3082/12, 616 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
10	Sensitivity phase shift / Laser vibrometers with an electric output	0 °	to	360 °		0.2 Hz to 5 kHz 5 kHz to 10 kHz 10 kHz to 20 kHz	0,5 ° 0,7 ° 1,5 °	Direct measurement with a laser vibrometer	KP-01 (ČSN ISO 16063-41)	
11*	Frequency of linear harmonic mechanical vibrations / Vibrometers, seismometers, linear vibration generators, vibration testing systems	0.2 Hz	to	50 kHz			0.01 %	Measurement or comparison on a standard calibration device or with a reference frequency transducer	KP-01 (ČSN ISO 16063-21), ČSN ISO 16063-44)	
12	Frequency of linear harmonic mechanical vibrations / Vibrometers, seismometers, vibration transducer calibration systems, reference linear vibration generators	0.2 Hz	to	50 kHz			0.01 %	Measurement or comparison on a standard calibration device or with a standard frequency transducer	KP-01 (ČSN ISO 16063-11), ČSN ISO 16063-44)	
13*	Transfer / Charge amplifiers, voltage amplifiers, frequency filters, vibration test system controllers	0.001 mV/pC 0.001 V/V	to	1,000 mV/pC 1,000 V/V		0.1 Hz to 1 Hz 1 Hz to 20 kHz 20 kHz to 50 kHz	0.4 % 0.3 % 1.0 %	By simulated electrical signal	KP-02	

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Vysoké učení technické v Brně  
CAB number 2395, CVVOZE Calibration Laboratory  
Technická 3082/12, 616 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
14*	Transfer phase shift / Charge amplifiers, voltage amplifiers, frequency filters, vibration test system controllers	0 °	to	360 °		0.1 Hz to 1 Hz 1 Hz to 20 kHz	0.75 ° 0.5 °	By simulated electrical signal	KP-02	
15*	Acceleration of harmonic vibrations / Vibrometers without a transducer, measuring chains without a transducer with an output to a display or a scale, vibration test system controllers <sup>4,5</sup>	0.01 m·s <sup>-2</sup>	to	10,000 m·s <sup>-2</sup>		0.1 Hz to 1 Hz 1 Hz to 20 kHz	0.4 % 0.3 %	By simulated electrical signal	KP-02	
16	Maximum value of half-sine wave mechanical shock acceleration / shock meters	50 m·s <sup>-2</sup> 200 m·s <sup>-2</sup> 2 km·s <sup>-2</sup> 40 km·s <sup>-2</sup>	to	200 m·s <sup>-2</sup> 2 km·s <sup>-2</sup> 40 km·s <sup>-2</sup> 100 km·s <sup>-2</sup>			2.0 % 1.5 % 1.8 % 3.0 %	Direct measurement or comparison with a reference sensor	KP-01 (ČSN ISO 16063-22)	
17	Sensitivity of half-sine wave mechanical shock acceleration sensors / shock sensors, vibration sensors	0.000 0.000	1 mV/m·s <sup>-2</sup> 1 pC/m·s <sup>-2</sup>	to	100 mV/m·s <sup>-2</sup> 100 pC/m·s <sup>-2</sup>	50 m·s <sup>-2</sup> to 200 m·s <sup>-2</sup>	2.0 %	Comparison with a standard sensor	KP-01 (ČSN ISO 16063-22)	

The Appendix is an integral part of  
Certificate of Accreditation No. 269/2023 of 31/05/2023

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Vysoké učení technické v Brně  
CAB number 2395, CVVOZE Calibration Laboratory  
Technická 3082/12, 616 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
						200 m·s <sup>-2</sup> to 2 km·s <sup>-2</sup> 2 to 40 km·s <sup>-2</sup> 40 to 100 km·s <sup>-2</sup>	1.5 % 1.8 % 3.0 %			
18	Sensitivity of half-sine wave mechanical shock sensors / shock hammers	0.1 mV/N 0.1 pC/N	to	100 mV/N 100 pC/N			3.0 %	Comparison with a reference sensor using a known reaction mass	KP-01 (ČSN ISO 16063-22)	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

<sup>4</sup> The calibrated quantities may include also the velocity and deviation, if the calibration is carried out using a harmonic vibration signal of a known frequency.

<sup>5</sup> Acceleration can be specified also in g, sensor sensitivity in pC/g, resp. mV/g units, where 1 g = 9.807 m·s<sup>-2</sup>

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Vysoké učení technické v Brně  
CAB number 2395, CVVOZE Calibration Laboratory  
Technická 3082/12, 616 00 Brno

**CMC for the field of measured quantity: Electrical quantities**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min	unit	max	unit					
1*	DC Voltage / DC voltage sources and meters, voltmeters, multimeters, signal analyzers, vibration meters, vibration test system controllers							Direct measurement or comparison with a reference multimeter	KP-03	
		0 mV	to	100 mV			0.010 % + 0.010 mV			
		100 mV	to	1 V			0.010 % + 0.020 mV			
		1 V	to	10 V			0.010 % + 0.10 mV			
		10 V	to	100 V			0.010 % + 1.0 mV			
2*	DC current / DC current sources and meters, ammeters, multimeters							Direct measurement or comparison with a reference multimeter	KP-03	
		0 mA	to	1 mA			0.10 % + 0.10 uA			
		1 mA	to	10 mA			0.10 % + 5.0 uA			
		10 mA	to	100 mA			0.10 % + 10 uA			
		100 mA	to	1 A			0.20 % + 0.20 mA			
3*	DC Resistance / Resistors, resistance boxes, shunts, dividers							Direct measurement by a standard multimeter	KP-03	
		0 Ω	to	100 Ω			0.020 % + 0.010 Ω			
		100 Ω	to	1 kΩ			0.020 % + 0.020 Ω			
		1 kΩ	to	10 kΩ			0.020 % + 0.20 Ω			
		10 kΩ	to	100 kΩ			0.020 % + 2.0 Ω			
		100 kΩ	to	1 MΩ			0.025 % + 20 Ω			
		1 MΩ	to	10 MΩ			0.10 % + 200 Ω			

**The Appendix is an integral part of  
Certificate of Accreditation No. 269/2023 of 31/05/2023**

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Vysoké učení technické v Brně**  
CAB number 2395, CVVOZE Calibration Laboratory  
Technická 3082/12, 616 00 Brno

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min	unit	max	unit					
4*	AC Voltage / AC voltage sources and meters, AC voltmeters, multimeters, signal analyzers, vibration meters, vibration sensor calibration systems, signal generators, controllers for vibration test systems							Direct measurement or comparison with a reference multimeter	KP-03	
		1 mV	to	100 mV		10 Hz to 20 kHz	0.10 % + 0.050 mV			
						20 kHz to 50 kHz	0.20 % + 0.10 mV			
						50 kHz to 100 kHz	0.80 % + 0.15 mV			
		100 mV	to	1 V		10 Hz to 20 kHz	0.10 % + 0.50 mV			
						20 kHz to 50 kHz	0.20 % + 1.0 mV			
						50 kHz to 100 kHz	0.80 % + 1.5 mV			
		1 V	to	10 V		10 Hz to 20 kHz	0.10 % + 5.0 mV			
						20 kHz to 50 kHz	0.20 % + 10 mV			
						50 kHz to 100 kHz	0.80 % + 15 mV			
		10 V	to	100 V		10 Hz to 20 kHz	0.10 % + 50 mV			
						20 kHz to 50 kHz	0.20 % + 100 mV			
						50 kHz to 100 kHz	0.80 % + 150 mV			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Vysoké učení technické v Brně**  
CAB number 2395, CVVOZE Calibration Laboratory  
Technická 3082/12, 616 00 Brno

**CMC for the field of measured quantity: Time and frequency quantities**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded mesurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min	unit	max	unit					
1*	Frequency / Signal generators, counters, multimeters, signal analyzers, vibration meters, calibration systems, vibration test system controllers	0.1 Hz	to	1 Hz		250 mV to 50 V	$5.0 \cdot 10^{-3}$	Direct measurement or comparison with a reference counter	KP-03	
		1 Hz	to	10 MHz		30 mV to 50 V	$2.0 \cdot 10^{-5}$			
2*	Period / Signal generators, counters, multimeters, signal analyzers, vibration meters, calibration systems, vibration test system controllers	100 ns	to	1 s		30 mV to 50 V	$2.0 \cdot 10^{-5}$	Direct measurement or comparison with a reference counter	KP-03	
		1 s	to	10 s		250 mV to 50 V	$5.0 \cdot 10^{-3}$			
3*	Time interval / Signal generators, counters, multimeters, signal analyzers	0.1 ms	to	0.1 s		250 mV to 50 V	$1.0 \cdot 10^{-3}$	Direct measurement or comparison with a reference counter	KP-03	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).