

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Technický a zkušební ústav stavební Praha, s.p.**  
CAB number. 2275, TZÚS Praha, s.p. Calibration Laboratory – TIS Branch  
Prosecká 811/76a, Prosek, 190 00 Praha 9

**Calibration laboratory locations:**

- |    |   |  |
|----|---|--|
| 1. | Technický a zkušební ústav stavební Praha, s.p. – TIS Branch  | Prosecká 811/76a, 190 00 Praha 9       |
| 3. | Technický a zkušební ústav stavební Praha, s.p. – Branch 0200 | Nemanická 441, 370 10 České Budějovice |

**CMC for the field of measured quantity: Length**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
1	Dial indicators	0 mm	to	100 mm		9 μm	Measurement on a spiral microscope	IP 0960K003	1	
2	Slide gauges	0 mm	to	250 mm		0.015 mm	Measurement parallel gauge blocks	IP 0960K006	1	
		250 mm	to	450 mm		0.03 mm				
		450 mm	to	1,000 mm		0.05 mm				
3 to 7	Reserved									
8	Electronic length sensors	0 mm	to	250 mm		0,01 mm	Measurement of parallel gauge blocks	IP 0960K005	1	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Explanations and abbreviations:

L Measured length in m

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Technický a zkušební ústav stavební Praha, s.p.**  
CAB number. 2275, TZÚS Praha, s.p. Calibration Laboratory – TIS Branch  
Prosecká 811/76a, Prosek, 190 00 Praha 9

**CMC for the field of measured quantity: Flow**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
1	Flowmeters that can be calibrated with water	0.005 m <sup>3</sup> /h	to	20 m <sup>3</sup> /h		cold water	0.15 %	Mass method	IP 0960K011	1
						hot water	0.20 %			
		20 m <sup>3</sup> /h	to	200 m <sup>3</sup> /h		cold water	0.15 %			
						hot water	0.20 %			
2	Flowmeters that can be calibrated with water	0.005 m <sup>3</sup> /h	to	20 m <sup>3</sup> /h		cold water	0.20 %	Volume method	IP 0960K011	1
						hot water	0.25 %			
		20 m <sup>3</sup> /h	to	200 m <sup>3</sup> /h		cold water	0.20 %			
						hot water	0.25 %			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Technický a zkušební ústav stavební Praha, s.p.**  
CAB number. 2275, TZÚS Praha, s.p. Calibration Laboratory – TIS Branch  
Prosecká 811/76a, Prosek, 190 00 Praha 9

**CMC for the field of measured quantity: Force, mechanical tests**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Workplace
		min unit	max unit					
1*	Force / tensile testing machines and presses, working force gauges	1 N	to 30 N	Tension	0.080 %	Loading with weights  Comparison with a standard force-proving instrument	IP 0960K072	1
		30 N 200 N 20 kN	to 200 N to 20 kN up to 100 kN		0.040 % 0.025 % 0.060 %			
2	Hardness / Schmidt type hardness testers (for concrete and other materials)	1 N	to 30 N	Pressure	0.080 %	Loading with weights  Comparison with a standard force-proving instrument	IP 0960K001	1
		30 N 200 N 2 kN 100 kN 500 kN 1 MN	to 200 N to 2 kN to 100 kN to 500 kN to 1 MN to 5 MN		0.040 % 0.025 % 0.040 % 0.070 % 0.090 % 0.025 %			
		73 Rk	to 77 Rk	Type L	0.7 Rk	Check impacts on a standard anvil	IP 0960K001	1
		79 Rk	to 83 Rk	Type N	0.7 Rk			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Explanations and abbreviations:

R<sub>k</sub> Offset value (dimensionless quantity) corresponding to 0.5 division of a scale

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Technický a zkušební ústav stavební Praha, s.p.**  
 CAB number. 2275, TZÚS Praha, s.p. Calibration Laboratory – TIS Branch  
 Prosecká 811/76a, Prosek, 190 00 Praha 9

**CMC for the field of measured quantity: Pressure**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min	unit	max	unit					
1	Deformation pressure gauges, electromechanical pressure gauges with pressure transducer or digital indication	0.025 MPa		to	60 MPa	positive gauge pressure liquid	0.08 %	Comparison with a standard piston pressure gauge	IP 0960K018	1

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Technický a zkušební ústav stavební Praha, s.p.**  
CAB number. 2275, TZÚS Praha, s.p. Calibration Laboratory – TIS Branch  
Prosecká 811/76a, Prosek, 190 00 Praha 9

**CMC for the field of measured quantity: Temperature**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min	unit	max	unit					
1	Reserved									
2*	Direct-indicating thermometers, temperature measuring equipment	-40 °C	to	100 °C		0.3 °C	Comparison with a standard digital thermometer in a calibrated device	IP 0960K014	3	
		100 °C	to	250 °C		0.4 °C				
		250 °C	to	500 °C		1.4 °C				
		500 °C	to	900 °C		1.9 °C				
		900 °C	to	1,200 °C		2.7 °C				
3 - 4	Reserved									

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).