

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**Technické služby ochrany ovzduší Praha a.s.**

Emission Calibration Laboratory

Jenečská 146/44, 161 00 Praha 6

**CMC for the field of measured quantity: Emission**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min.	unit	max.	unit					
1*	Concentration of gaseous components							Comparison with a gas standard prepared in a calibrator in a volume-defined mixture of RM with inert gas	IP 101-03 (EN ISO 9169, ISO 11095)	
	NO	0 µmol/mol	to	6 µmol/mol			0.03 µmol/mol			
		6 µmol/mol	to	3,000 µmol/mol			0.50 %			
	CO	0 µmol/mol	to	42 µmol/mol			0.12 µmol/mol			
		42 µmol/mol	to	8,000 µmol/mol			0.28 %			
	SO <sub>2</sub>	0 µmol/mol	to	5 µmol/mol			0.03 µmol/mol			
		5 µmol/mol	to	2,000 µmol/mol			0.54 %			
	C <sub>3</sub> H <sub>8</sub>	0 µmol/mol	to	11 µmol/mol			0.05 µmol/mol			
		11 µmol/mol	to	5,000 µmol/mol			0.42 %			
	CH <sub>4</sub>	0 µmol/mol	to	10 µmol/mol			0.05 µmol/mol			
		10 µmol/mol	to	40,000 µmol/mol			0.46 %			
	NH <sub>3</sub>	0 µmol/mol	to	6 µmol/mol			0.012 µmol/mol			
		6 µmol/mol	to	2,000 µmol/mol			2.00 %			
	N <sub>2</sub> O	0 µmol/mol	to	6 µmol/mol			0.12 µmol/mol			
		6 µmol/mol	to	2,000 µmol/mol			2.00 %			
	NO <sub>2</sub>	0 µmol/mol	to	1.5 µmol/mol			0.03 µmol/mol			
		1.5 µmol/mol	to	2,000 µmol/mol			2.00 %			
	O <sub>2</sub>	0 mol/mol	to	0.002 mol/mol			0.00001 mol/mol			
		0.002 mol/mol	to	0.24 mol/mol			0.54 %			

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**Technické služby ochrany ovzduší Praha a.s.**

Emission Calibration Laboratory

Jenečská 146/44, 161 00 Praha 6

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min.	unit	max.	unit					
	CO <sub>2</sub>	0 mol/mol 0.002 mol/mol	to	0.24 mol/mol 0.24 mol/mol			0.00001 mol/mol 0.54 %			
2*	Converter efficiency NO <sub>2</sub> – NO (up to 500 µmol/mol NO)	0 %	to	100 %			0.4 % abs	Direct measurement	IP 102-01 (EN ISO 9169, EN 14792, US EPA 40 CFR Ch.I § 86.332 79)	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).