

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**VOP CZ, s.p.**  
CAB number 2399, Calibration Laboratory  
Dukelská 102, 742 42 Šenov u Nového Jičína

**CMC for the field of measured quantity: Length**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
1	Slide gauges	0 mm	to	250 mm			(11·L + 12) μm (14·L + 19) μm	Comparison with parallel gauge blocks	KP-GL-2.1	
2	Slide depth gauges	0 mm	to	250 mm			(11·L + 12) μm	Comparison with parallel gauge blocks	KP-GL-2.8	
3	Slide height gauges	0 mm	to	1,000 mm			(14·L + 19) μm	Comparison with parallel gauge blocks	KP-GL-2.9	
4	Weld gauges	0 mm	to	100 mm			(14·L + 120) μm	Comparison with parallel gauge blocks	KP-GL-2.7	
5	Micrometer calliper gauges	0 mm	to	200 mm			(11·L + 1.5) μm (14·L + 16) μm	Comparison with parallel gauge blocks	KP-GL-2.2	
6	Pasameters	0 mm	to	200 mm			(11·L + 1.5) μm	Comparison with parallel gauge blocks	KP-GL-2.3	
7	Dial indicators	0 mm	to	25 mm			(14·L + 1.5) μm	Direct measurement on a device for the calibration of dial indicators	KP-GL-2.4	
8	Two-contact internal gauges	0 mm	to	200 mm			(14·L + 1.3) μm	Direct measurement on a device for the calibration of dial indicators	KP-GL-2.4	
9	Thickness gauges	0 mm	to	1 mm			1.2 μm	Direct measurement on a length gauge	KP-GL-2.6	
10	Plain cylindrical gauges	0 mm	to	200 mm			(14·L + 2.4) μm	Direct measurement on a length gauge	KP-GL-1.2	
11	Male thread gauges for metric threads	3 mm	to	100 mm			(14·L + 2.8) μm	Direct measurement of a dimension across wires on a length gauge	KP-GL-4.1	
12	Tape measures	0 mm	to	10,00 0 mm			(0.042·L + 0.13) mm	Comparison with a linear scale	KP-GL-1.7	
13	Steel rules	0 mm	to	2,000 mm			(0.042·L + 0.13) mm	Comparison with a linear scale	KP-GL-1.6	
14	Feeler gauges	0 mm	to	10 mm			(14·L + 1.9) μm	Direct measurement on a length gauge	KP-GL-1.4	
15	Calibration foils	0 mm	to	10 mm			1.9 μm	Direct measurement on a length gauge	KP-GL-1.9	
16	Inside micrometer gauges	0 mm	to	250 mm			(14·L + 2.4) μm	Direct measurement on a length gauge	KP-GL-2.10	

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**VOP CZ, s.p.**  
CAB number 2399, Calibration Laboratory  
Dukelská 102, 742 42 Šenov u Nového Jičína

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
17	Surface rules	0 mm	to	1,100 mm			$(1.4 \cdot L + 7) \mu\text{m}$	Comparison with a linearity standard and parallel gauge blocks	KP-GL-3.1	
18	Two-contact and three-contact inside micrometers	0 mm	to	200 mm			$(14 \cdot L + 1.5) \mu\text{m}$	Direct measurement on a length gauge	KP-GL-2.11	
19	Setting and limit rings	22 mm	to	200 mm			$(14 \cdot L + 2.2) \mu\text{m}$	Comparative measurement by a distance meter	KP-GL-1.5	
20	90° Angles	0 mm	to	4 mm		length up to 630 length up to 1,000 mm	$(1.4 \cdot L + 9.2) \mu\text{m}$  $(2 \cdot L + 22) \mu\text{m}$	Direct measurement of parallel gauge blocks, with an indicator	KP-GU-1.1	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Explanatory notes:

L - nominal length (m)

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**VOP CZ, s.p.**  
CAB number 2399, Calibration Laboratory  
Dukelská 102, 742 42 Šenov u Nového Jičína

**CMC for the field of measured quantity: Plane angle**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
1	Angle gauges	0 °		to	360 °		4.8 ´	Comparison with angle gauges	KP-GU-2.1	
2	Clinometers	0 °		to	360 °		0.12 °	Comparison with a sine ruler	KP-GU-2.2	
3	Machinery and builder's levels (up to 2,000mm)	0 mm/m		to	87 mm/m		0.02 mm/m	Comparison with a sine ruler	KP-GU-2.3	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**VOP CZ, s.p.**  
CAB number 2399, Calibration Laboratory  
Dukelská 102, 742 42 Šenov u Nového Jičína

**CMC for the field of measured quantity: Force, mechanical tests**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min	jedn.	max	jedn.					
1	Torque wrenches and screwdrivers	1 Nm	to	1,000 Nm			0.6 %	Comparison with a torque sensor	KP-MS-3.1	
2	Hardness test plates and samples - Rockwell	10 HRBW	to	100 HRBW	HRBW		1.3 %	ČSN EN ISO 6508-3	KP-MS-2.3	
		20 HRC	to	70 HRC	HRC		1.2 %			
3	Hardness test plates and samples - Brinell	10 HBW	to	650 HBW	HBW 2.5		1.4 %	ČSN EN ISO 6506-3	KP-MS-2.4	
4	Hardness test plates and samples - Vickers	10 HV	to	3,000 HV	HV 10		1.6 %	ČSN EN ISO 6507-3	KP-MS-2.2	
					HV 30		1.2 %			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**VOP CZ, s.p.**  
CAB number 2399, Calibration Laboratory  
Dukelská 102, 742 42 Šenov u Nového Jičína

**CMC for the field of measured quantity: Pressure**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
1	Deformation pressure gauges	0 MPa	to	3 MPa		relative pressure	liquid (alcohol, oil)	4.8 kPa 32 kPa 65 kPa 190 kPa	Comparison with a standard digital pressure gauge	KP-MT-1.1
		3 MPa	to	20 MPa						
		20 MPa	to	50 MPa						
		50 MPa	to	140 MPa						
2	Digital pressure gauges	0 MPa	to	3 MPa		relative pressure	liquid (air)	3.8 kPa 25 kPa 37 kPa 104 kPa	Comparison with a standard digital pressure gauge	KP-MT-1.2
		3 MPa	to	20 MPa						
		20 MPa	to	50 MPa						
		50 MPa	to	140 MPa						

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**VOP CZ, s.p.**  
CAB number 2399, Calibration Laboratory  
Dukelská 102, 742 42 Šenov u Nového Jičína

**CMC for the field of measured quantity: Electrical quantities**

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
1	Direct-current voltage/ voltmeters, multimeters, analogue and digital	0 mV	to	20 mV			0.035 % + 10 µV	Direct measurement by a standard calibrator	KP-EM-2.1 KP-EM-2.2	
		20 mV	to	100 mV			0.095 %			
		100 mV	to	200 mV			0.029 %			
		0.2 V	to	1 V			0.012 %			
		1 V	to	2 V			0.0054 %			
		2 V	to	10 V			0.0090 %			
		10 V	to	20 V			0.0047 %			
		20 V	to	100 V			0.0090 %			
		100 V	to	240 V			0.0047 %			
		240 V	to	1,000 V			0.029 %			
2	Alternating-current voltage/ voltmeters, multimeters, analogue and digital	10 mV	to	20 mV		20 Hz to 10 kHz	0.58 %	Direct measurement by a standard calibrator	KP-EM-2.1 KP-EM-2.2	
		20 mV	to	100 mV		20 Hz to 10 kHz	0.56 %			
		100 mV	to	200 mV		20 Hz to 10 kHz	0.21 %			
		0.2 V	to	1 V		20 Hz to 10 kHz	0.066 %			
		1 V	to	2 V		20 Hz to 10 kHz	0.038 %			
		2 V	to	10 V		20 Hz to 10 kHz	0.066 %			
		10 V	to	20 V		20 Hz to 10 kHz	0.038 %			
		20 V	to	100 V		20 Hz to 10 kHz	0.58 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**VOP CZ, s.p.**  
CAB number 2399, Calibration Laboratory  
Dukelská 102, 742 42 Šenov u Nového Jičína

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
		100 V	to	240 V		20 Hz to 1 kHz	0.15 %			
		240 V	to	1,000 V		20 Hz to 1 kHz	0.13 %			
3	Direct current/ ammeters, multimeters, analogue and digital	1 µA	to	200 µA			0.058 % + 20 nA	Direct measurement by a standard calibrator	KP-EM-2.1 KP-EM-2.2	
		0.2 mA	to	1 mA			0.078 %			
		1 mA	to	2 mA			0.035 %			
		2 mA	to	10 mA			0.045 %			
		10 mA	to	20 mA			0.019 %			
		20 mA	to	100 mA			0.046 %			
		100 mA	to	200 mA			0.019 %			
		0.2 A	to	1 A			0.072 %			
		1 A	to	2 A			0.029 %			
		2 A	to	10 A			0.13 %			
		10 A	to	20 A			0.047 %			
4	Alternating current/ ammeters, multimeters, analogue and digital	10 µA	to	100 µA		20 Hz to 1 kHz	0.52 %	Direct measurement by a standard calibrator	KP-EM-2.1 KP-EM-2.2	
		100 µA	to	200 µA		20 Hz to 1 kHz	0.21 %			
		0.2 mA	to	1 mA		20 Hz to 1 kHz	0.24 %			
		1 mA	to	2 mA		20 Hz to 1 kHz	0.12 %			
		2 mA	to	10 mA		20 Hz to 1 kHz	0.18 %			
		10 mA	to	20 mA		20 Hz to 1 kHz	0.092 %			

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**VOP CZ, s.p.**  
CAB number 2399, Calibration Laboratory  
Dukelská 102, 742 42 Šenov u Nového Jičína

Ord. num- ber <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit	max.	unit					
		20 mA	to	100 mA		20 Hz to 1 kHz	0.17 %			
		100 mA	to	200 mA		20 Hz to 1 kHz	0.090 %			
		0.2 A	to	1 A		20 Hz to 1 kHz	0.21 %			
		1 A	to	2 A		20 Hz to 1 kHz	0.12 %			
		2 A	to	10 A		20 Hz to 1 kHz	0.49 %			
		10 A	to	20 A		20 Hz to 1 kHz	0.23 %			
5	DC resistance/ ohmmeters, multimeters, analogue and digital	0 Ω	to	100 Ω			0.036 % + 30 mΩ	Direct measurement with a standard calibrator or resistance standard	KP-EM-2.1 KP-EM-2.2	
		100 Ω	to	400 Ω			0.038 %			
		0.4 kΩ	to	2 kΩ			0.023 %			
		2 kΩ	to	10 kΩ			0.019 %			
		10 kΩ	to	200 kΩ			0.018 %			
		0.2 MΩ	to	1 MΩ			0.059 %			
		1 MΩ	to	4 MΩ			0.12 %			
		4 MΩ	to	20 MΩ			0.23 %			
		20 MΩ	to	50 MΩ			0.59 %			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).