

**The Appendix is an integral part of  
Certificate of Accreditation No. 47/2024 of 05/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**TÜV SÜD Czech s.r.o.**  
CAB number 2405, Calibration Laboratory  
Novodvorská 994, 142 21 Praha 4

**Calibration laboratory locations:**

1. UNO TECHNOLOGY PARK Bezděčín, Hala H1a, č.p. 108, 293 01 Mladá Boleslav, Czech Republic

**CMC for the field of measured quantity: Length**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Sensors of displacement, distance and position	0 mm	to	300 mm		0.04 mm	comparison with a height gauge	I540 – 068 – 10		
		300 mm	to	1,500 mm		0.20 mm				
2	Belt motion sensors	-400 mm	to	400 mm		(0.2·L + 0.40) mm	comparison with a calibration jig	I540 – 068 – 10		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher, depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

**Explanatory notes:**

L – measured length in metres

**The Appendix is an integral part of  
Certificate of Accreditation No. 47/2024 of 05/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**TÜV SÜD Czech s.r.o.**  
CAB number 2405, Calibration Laboratory  
Novodvorská 994, 142 21 Praha 4

**CMC for the field of measured quantity: Plane angle**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Sensors of position and inclinometers	0 °		to	360 °		0.02°	comparison with a digital inclinometer or rotary index table	I540 – 068 – 10	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher, depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

**The Appendix is an integral part of  
Certificate of Accreditation No. 47/2024 of 05/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**TÜV SÜD Czech s.r.o.**  
CAB number 2405, Calibration Laboratory  
Novodvorská 994, 142 21 Praha 4

**CMC for the field of measured quantity: Mechanical motion**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Meters and sensors of half-sine wave mechanical shock acceleration	100 m·s <sup>-2</sup>		to	1,471 m·s <sup>-2</sup>		1.0 %	comparison with a standard acceleration sensor	I 540 – 068 – 43	
		1,471 m·s <sup>-2</sup>		to	2,000 m·s <sup>-2</sup>		1.5 %			
		2,000 m·s <sup>-2</sup>		to	40,000 m·s <sup>-2</sup>		1.8 %			
2	Acceleration of linear mechanical vibrations of harmonic wave form <sup>4,5</sup> / Vibration sensors, vibrometers, vibration calibrators, vibration generators <sup>4,5</sup>	0.1 m·s <sup>-2</sup>		to	295 m·s <sup>-2</sup>	5 Hz up to 10 Hz	2.0 %	comparison with a standard acceleration sensor	I 540 – 068 – 45	
						10 Hz up to 20 Hz	1.0 %			
						20 Hz up to 80 Hz	0.75 %			
						80 Hz	0.5 %			
						80 Hz up to 1,000 Hz	0.75 %			
						1,000 Hz up to 5,000 Hz	1.0 %			
						5,000 Hz up to 10,000 Hz	2.0 %			
3	Sensitivity of vibration sensors and vibrometers <sup>4,5</sup>	0.01 mV/(m·s <sup>-2</sup> )		to	10,000 mV/(m·s <sup>-2</sup> )	5 Hz up to 10 Hz	2.0 %	comparison with a vibration standard	I 540 – 068 – 45	
		0.01 pC/(m·s <sup>-2</sup> )		to	1,000 pC/(m·s <sup>-2</sup> )	10 Hz up to 20 Hz	1.0 %			
		0.01 mV/(m·s <sup>-1</sup> )		to	10,000 mV/(m·s <sup>-1</sup> )	20 Hz up to 80 Hz	0.75 %			
		0.01 mV/m		to	10,000 mV/m	80 Hz	0.5 %			
						80 Hz up to 1,000 Hz	0.75 %			
						1,000 Hz up to 5,000 Hz	1.0 %			
						5,000 Hz up to 10,000 Hz	2.0 %			

**The Appendix is an integral part of  
Certificate of Accreditation No. 47/2024 of 05/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**TÜV SÜD Czech s.r.o.**  
CAB number 2405, Calibration Laboratory  
Novodvorská 994, 142 21 Praha 4

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
4	Transmission of amplifiers, filters and vibrometers	0.001 mV/pC to 10,000 mV/pC				0.2 Hz up to 1 Hz	0.5 %	direct measurement	I 540 – 068 – 45	
						1 Hz up to 5,000 Hz	0.4 %			
						5,000 Hz up to 10,000 Hz	0.4 %			
						10,000 Hz up to 20,000 Hz	0.6 %			
						20,000 Hz up to 50,000 Hz	1.0 %			
		0.001 V/V to 1,000 V/V				0.2 Hz up to 1 Hz	0.4 %			
						1 Hz up to 20,000 Hz	0.3 %			
						20,000 Hz up to 50,000 Hz	1.0 %			
5	Vibration measurement <sup>4,5/</sup> Vibration generators	0.1 m·s <sup>-2</sup> to 295 m·s <sup>-2</sup>				5 Hz up to 10 Hz	2.0 %	direct measurement	I 540 – 068 – 45	
						10 Hz up to 20 Hz	1.0 %			
						20 Hz up to 80 Hz	0.75 %			
						80 Hz	0.5 %			
						80 Hz up to 1,000 Hz	0.75 %			
						1,000 Hz up to 5,000 Hz	1.0 %			
						5,000 Hz up to 10,000 Hz	2.0 %			
6*	Speedometers, GPS and radar speedometers, speed sensors					distance standard measured distance	0.02 %	comparison with distance and time standard	I 540 – 068 – 29	
							0.04 %			
7*	Speedometers with a rolling wheel						0.1 % + 0.01 m·min <sup>-1</sup>	direct generation of circumferential velocity	I 540 – 068 – 05	

**The Appendix is an integral part of  
Certificate of Accreditation No. 47/2024 of 05/02/2024**

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**TÜV SÜD Czech s.r.o.**  
CAB number 2405, Calibration Laboratory  
Novodvorská 994, 142 21 Praha 4

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
8*	Revolution counters, rpm sensors, stroboscopes	500 min <sup>-1</sup> 10,000 min <sup>-1</sup>		to	10,000 min <sup>-1</sup> 100,000 min <sup>-1</sup>		0.006 min <sup>-1</sup> 0.06 min <sup>-1</sup>	direct measurement of an optical or electrical signal	I 540 – 068 – 05	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher, depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

<sup>4</sup> The measured quantity can also be velocity and displacement, assuming that a vibrational signal of the harmonic waveform is generated at a known frequency.

<sup>5</sup> It can also be given in the units g, pC/g or mV/g, where 1 g = 9.807 m·s<sup>-2</sup>

**The Appendix is an integral part of  
Certificate of Accreditation No. 47/2024 of 05/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**TÜV SÜD Czech s.r.o.**  
CAB number 2405, Calibration Laboratory  
Novodvorská 994, 142 21 Praha 4

**CMC for the field of measured quantity: Pressure, mechanical stress**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1	Deformation and electromechanical manometers					gas		comparison with a digital pressure gauge	I540 – 068 – 3 (ČSN EN 837– 1, ČSN EN 837– 3, EURAMET cg-17)	
		-95	kPa	to	0	kPa	0.1 kPa			
		0	MPa	to	0.7	MPa	0.03 % + 0.08 kPa			
		0.7	MPa	to	3.5	MPa	0.03 % + 0.4 kPa			
		3.5	MPa	to	7	MPa	0.03 % + 0.8 kPa			
		7	MPa	to	20	MPa	0.03 % + 2.3 kPa			
						oil		comparison with a piston pressure gauge		
		0	MPa	to	6	MPa	3.5 kPa			
		6	MPa	to	60	MPa	0.06 %			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher, depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

**The Appendix is an integral part of  
Certificate of Accreditation No. 47/2024 of 05/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**TÜV SÜD Czech s.r.o.**  
CAB number 2405, Calibration Laboratory  
Novodvorská 994, 142 21 Praha 4

**CMC for the field of measured quantity: Time, frequency**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Digital speed cameras			1,000 Hz			2.1·10 <sup>-4</sup> Hz	direct measurement with a standard counter	I 540 – 068 – 44	
2*	Time interval / stopwatches, timers and other chronometers	1 ms 1 s	to	1,000 ms 10,800 s			1.0·10 <sup>-5</sup> s 6.0·10 <sup>-4</sup> s	comparison with a standard counter	I 540 – 068 – 02	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher, depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).