The Appendix is an integral part of Certificate of Accreditation No. 500/2022 of 26/10/2022

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Carl Zeiss spol. s r. o. Calibration Laboratory Carl Zeiss IQS Radlická 3201/14, Smíchov, 150 00 Praha 5

CMC for the field of measured quantity: Length

Ord. num ber ¹	Calibrated quantity / Subject of calibration	Nominal range					Parameter(s)	Lowest expanded	Calibration principle	Calibration procedure	Work-
		min.	unit		max.	unit	quantity	uncertainty specified ²		identification ³	place
1*	Coordinate measuring machines Carl Zeiss with contacting probing	0) mm	to	3,030	mm		(0.23·L + 0.06) μm	Comparison using a step gauge and calibration ball (ČSN EN ISO 10360-2, ČSN EN ISO 10360-3, ČSN EN ISO 10360-4 and VDI/VDE 2617 Blatt 2.1, VDI/VDE 2617 Blatt 2.2, VDI/VDE 2617 Blatt 4)	KP001-ZEISS-10360 KP002-ZEISS-2617	

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

L length in m