

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

František Knížek
CAB number 2290, František Knížek - KALEX, Calibration Centre
A. Dvořáka 719, 533 41 Lázně Bohdaneč

Calibration laboratory locations:

1. **Workplace Lázně Bohdaneč** A. Dvořáka 719, 533 41 Lázně Bohdaneč
2. **Workplace Vlčí Habřina** Vlčí Habřina 122, 533 41 Lázně Bohdaneč

CMC for the field of measured quantity: Length

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max.	unit					
1	Parallel gauge blocks	0.5 mm	to	100 mm		$(2L + 0.2) \mu\text{m}$	Comparison with parallel gauge blocks in vertical position on a comparator Comparison with parallel gauge blocks in horizontal position on a length gauge Comparison with parallel gauge blocks in horizontal position on a length gauge	KPA-1.01	1	
		125 mm	to	500 mm		$(2.2L + 0.3) \mu\text{m}$			1, 2	
		500 mm	to	1,000 mm		$(2.2L + 0.3) \mu\text{m}$			2	
2*	Slide gauges, depth gauges, height gauges	0 mm	to	3,000 mm		$(8.7L + 11) \mu\text{m}$	Measurement using parallel gauge blocks	KPA-1.02	1	
3	Micrometers for external measurement Micrometer calliper gauges Pasameters Micropasameters	0 mm	to	500 mm		$(3L + 1) \mu\text{m}$	Measurement using parallel gauge blocks	KPA-1.03	1	
		0 mm	to	500 mm		$(3L + 1) \mu\text{m}$				
		0 mm	to	500 mm		$(3L + 1) \mu\text{m}$				
4	Micrometers for internal measurement Inside micrometer gauges Micrometer depth gauges Inside micrometers Micrometric heads	14 mm	to	500 mm		$(3L + 1) \mu\text{m}$	Measurement using parallel gauge blocks	KPA-1.04	1	
		14 mm	to	500 mm		$(2L + 1.1) \mu\text{m}$				
		14 mm	to	500 mm		$(2L + 1.1) \mu\text{m}$				
		0 mm	to	500 mm		$(3L + 1) \mu\text{m}$				

**The Appendix is an integral part of
Certificate of Accreditation No. 515/2023 of 02/10/2023**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

František Knížek
CAB number 2290, František Knížek - KALEX, Calibration Centre
A. Dvořáka 719, 533 41 Lázně Bohdaneč

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max.	unit					
5	Deviation meters Dial gauges Pupitasts Somcators Internal gauges	0 mm	to	100 mm		0.88 μm 1.2 μm 1.2 μm 1.2 μm	Measurement on a dial indicator calibration instrument	KPA-1.05	1	
6	Limit gauges for external measurement Micrometer calliper gauges Plain rings Threaded rings	1 mm	to	500 mm		(3L + 1) μm (3.6L + 1.2) μm (7.4L + 2.1) μm	Measurement on a length gauge	KPA-1.06	1	
7	Limit gauges for internal measurement cylinder, flat Thread gauges Feeler gauges Measuring wires Gauges for radius Gauges for threads Gauges for paint thickness	0 mm	to	500 mm		(5.3L + 0.75) μm (2.8L + 2.8) μm 3.6 μm 0.54 μm 4.0 μm 4.0 μm 1.4 μm	Measurement on a length gauge	KPA-1.07	1	
8*	Rules Steel rules Measuring magnifier Steel tape measures Tapes	0 mm	to	10,000 mm		(4.6L + 4.7) μm (4.6L + 4.7) μm (3.8L + 140) μm (0.06L + 0.3) mm	Measurement on a coordinate measuring machine Comparison with a steel gauge	KPA-1.08	1, 2	

**The Appendix is an integral part of
Certificate of Accreditation No. 515/2023 of 02/10/2023**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

František Knížek
CAB number 2290, František Knížek - KALEX, Calibration Centre
A. Dvořáka 719, 533 41 Lázně Bohdaneč

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max.	unit					
9*	Two-coordinate measuring machines, Measuring microscopes, Profile projectors	0 mm		to	1,000 mm		3.2 μm	Measuring with a glass ruler	KPA-1.09	1
10	Atypical length gauges	0 mm		to	250 mm		4.0 μm	Measurement on a coordinate measuring machine	KPA-1.10	1
11*	Surface plates, blocks, plates (longer side up to 5,000 mm) - flatness, straightness	0 m		to	10 m		3.8 μm	Measurement by an electronic level	KPA-1.13	1
12*	Length gauges	0 mm		to	1,000 mm		(2L + 0.25) μm	Comparison with parallel gauge blocks	KPA-1.14	1

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

František Knížek
CAB number 2290, František Knížek - KALEX, Calibration Centre
A. Dvořáka 719, 533 41 Lázně Bohdaneč

CMC for the field of measured quantity: Plane angle

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max.	unit					
1	Rigid angle gauges							KPA-1.11	1	
	check squares – angle between arms	0 °	to	180 °		32 µm/m	Measurement on a coordinate measuring machine			
	taper gauges	0 °	to	90 °		7''				
	centre squares	0 °	to	90 °		32 µm/m	Measurement on a small angle generator			
gauges for threads	0 °	to	90 °		7''					
2	Angle gauges	0 °	to	360 °		1.8'	Measurement using angle gauges	KPA-1.12	1	
3	Machinery levels – sensitivity measurement				Nominal sensitivity up to 0.02 mm/m		Measurement on a small angle generator	KPA-1.11	1	
		-1 °	to	1 °		5 µm/m				

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

František Knížek
CAB number 2290, František Knížek - KALEX, Calibration Centre
A. Dvořáka 719, 533 41 Lázně Bohdaneč

CMC for the field of measured quantity: Mass

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max.	unit					
1*	Scales with non-automatic function, mechanical, digital	0 kg	to	2 kg		$1.6 \cdot 10^{-6}$	Loading using a reference weight	KPA-2.01	1	
		2 kg	to	3 kg		$5 \cdot 10^{-6}$	class E2			
		3 kg	to	45 kg		$1.6 \cdot 10^{-5}$	class F1			
		45 kg	to	6,000 kg		$5 \cdot 10^{-5}$	class F2			
		6,000 kg	to	30,000 kg		$1.6 \cdot 10^{-4}$	class M1			
							class M1 with substitute load			
2	Weights and other objects	1 g	to	500 g		8.2 mg	Comparison with a standard weight	KPA-2.01	1, 2	
		0.5 kg	to	1 kg		8.6 mg				
		1 kg	to	2 kg		10 mg				
		2 kg	to	5 kg		16 mg				
		5 kg	to	20 kg		59 mg				

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

František Knížek
CAB number 2290, František Knížek - KALEX, Calibration Centre
A. Dvořáka 719, 533 41 Lázně Bohdaneč

CMC for the field of measured quantity: Force, torque

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max.	unit					
1*	Torque drivers	0.25 Nm	to	50 Nm		0.01	Measurement by a torque sensor	KPA-5.01	1	
	Torque wrenches, screwdrivers, Moment of force meters	0.25 Nm	to	0.5 Nm		0.01				
		0.5 Nm	to	200 Nm		0.005				
		200 Nm	to	500 Nm		0.005				
500 Nm	to	2,000 Nm		0.005						
2*	Force meters, force measuring devices	0 N	to	500 N	Tension, Pressure	0.001	Measurement by a reference force meter	KPA-5.02	1	
		500 N	to	10,000 N		0.003				
		10,000 N	to	100,000 N		0.005				

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

**The Appendix is an integral part of
Certificate of Accreditation No. 515/2023 of 02/10/2023**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

František Knížek
CAB number 2290, František Knížek - KALEX, Calibration Centre
A. Dvořáka 719, 533 41 Lázně Bohdaneč

CMC for the field of measured quantity: Pressure

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max	unit					
1*	Deformation manometers, Digital manometers, Pressure measuring chains, Pressure transducers with electrical output	-95 kPa	to	350 kPa		Underpressure / overpressure Gases	0.26 kPa	Comparison with a standard manometer	KPA-4.01, KPA-4.02	1
		350 kPa	to	1,000 kPa			0.58 kPa			
		1 MPa	to	3.5 MPa		Liquids	2.1 kPa			
		3.5 MPa	to	6 MPa			6.9 kPa			
		0 MPa	to	20 MPa			35 kPa			
		20 MPa	to	50 MPa			87 kPa			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

František Knížek
CAB number 2290, František Knížek - KALEX, Calibration Centre
A. Dvořáka 719, 533 41 Lázně Bohdaneč

CMC for the field of measured quantity: Temperature

Ord. number ₁	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max.	unit					
1	Glass thermometers	- 40 °C	to	200 °C	0 °C		0.07 °C 0.05 °C	Comparison with a standard thermometer in a liquid bath	KPA-3.01	2
2*	Direct indicating thermometers, temperature controllers	-40 °C	to	200 °C			0.08 °C	Comparison with a standard thermometer in a liquid bath	KPA-3.02	1
		200 °C	to	400 °C			0.44 °C	Comparison with a standard thermometer in a vertical furnace		
		400 °C	to	650 °C			1.5 °C			
		650 °C	to	900 °C			1.8 °C			
		900 °C	to	1,200 °C			2.4 °C			
3*	Infrared thermometers	50 °C	to	500 °C			3.2 °C	Comparison with a standard (black body)	KPA-3.03	1
4*	Contact thermometers	0 °C	to	50 °C			1.7 °C	Comparison with a standard thermometer	KPA-3.04	1
		50 °C	to	100 °C			1.9 °C			
		100 °C	to	200 °C			2.3 °C			
		200 °C	to	400 °C			2.6 °C			
		400 °C	to	600 °C			3.5 °C			
5*	Thermoelectric sensors and measuring chains Thermocouple sensors	-40 °C	to	200 °C		K, J, N	0.4 °C	Comparison with a standard thermometer in a liquid bath	KPA-3.05	1
		200 °C	to	400 °C			0.6 °C	Comparison with a standard thermometer in a vertical furnace		
		400 °C	to	650 °C			1.6 °C			

**The Appendix is an integral part of
Certificate of Accreditation No. 515/2023 of 02/10/2023**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

František Knížek
CAB number 2290, František Knížek - KALEX, Calibration Centre
A. Dvořáka 719, 533 41 Lázně Bohdaneč

Ord. number ₁	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max.	unit					
	Measuring chain without a sensor	650 °C	to	900 °C		2.3 °C	Direct measurement on a calibrator			
		900 °C	to	1,100 °C		2.6 °C				
		-100 °C	to	1,100 °C		0.3 °C				
6*	Resistance sensors and measuring chains						Comparison with a standard thermometer in a liquid bath Comparison with a standard thermometer in a vertical furnace Direct measurement on a calibrator	KPA-3.06	1	
	Resistance sensors	-40 °C	to	200 °C		0.2 °C				
		200 °C	to	400 °C		0.5 °C				
	Measuring chain without a sensor	-100 °C	to	400 °C		0.2 °C				

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies..

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

**The Appendix is an integral part of
Certificate of Accreditation No. 515/2023 of 02/10/2023**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

František Knížek
CAB number 2290, František Knížek - KALEX, Calibration Centre
A. Dvořáka 719, 533 41 Lázně Bohdaneč

CMC for the field of measured quantity: Humidity

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work place
		min	unit	max	unit					
1*	Instruments for measuring air humidity	10 % RH	to	90 % RH		Temperature range (15 to 60) °C	0.01 + 1.6 % RH	Comparison with a reference hygrometer	KPA-6.01	1

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value given here is based on the best laboratory conditions achievable; the uncertainty value of a particular calibration may be higher depending on the conditions of that calibration. For identical limit values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).