## The Appendix is an integral part of Certificate of Accreditation No. 547/2022 of 11/11/2022

## Accredited entity according to ČSN EN ISO/IEC 17025:2018:

## **NETTO Electronics s.r.o**

NETTO Electronics s.r.o. Calibration Laboratory Malešická 2777/45a, Žižkov, 130 00 Praha 3

## CMC for the field of measured quantity: Mass

Ord. number	Calibrated quantity / Subject of calibration	Nominal range			Parameter(s) of	Lowest expanded measurement		Calibration	Work-
		min uni	t	max unit	the meas. quantity	uncertainty specified <sup>2</sup>	Calibration principle	procedure identification <sup>3</sup>	place
1*	Electronic scales with non- automatic function	1 g	to	9.61 kg		1.9·10 <sup>-6</sup>	Load with class E2 reference weight according to OIML R111	KP 2.1.3	
		9.61 kg	to	75.72 kg		1.9·10 <sup>-5</sup>	Load with class F2 reference weight according to OIML R111		
		75.72 kg	to	3,536.72 kg		5.8·10 <sup>-5</sup>	Load with class M1 reference weight according to OIML R111		
		3,536.72 kg	to	6,000 kg	d = 0.5  kg	0.46 kg	Loading with class M1 reference weight using substitute load		

Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value given here is based on the best laboratory conditions achievable; the uncertainty value of a particular calibration may be higher depending on the conditions of that calibration. For identical limit values of adjacent ranges, the lower uncertainty value always applies.

<sup>&</sup>lt;sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).