

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**TESTIMA, spol. s r.o.**  
TESTIMA Calibration Laboratory  
Husova 353/6, 250 01 Brandýs nad Labem

**CMC for the field of measured quantity: Length**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min	unit	max	unit					
1	Ultrasonic thickness gauges	1 mm	to	100 mm		0.1 mm	Comparison with the value of the standard	KP 2		
2	Layer thickness gauges	8 μm	to	10200 μm		2.7 μm	Comparison with the value of the standard	KP 3		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value given here is based on the best laboratory conditions achievable; the uncertainty value of a particular calibration may be higher depending on the conditions of that calibration. For identical limit values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**TESTIMA, spol. s r.o.**  
TESTIMA Calibration Laboratory  
Husova 353/6, 250 01 Brandýs nad Labem

**CMC for the field of measured quantity: Testing of properties and defects of materials**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit	max	unit					
1	Analogue and digital ultrasonic defectoscopes							Signal display on the display scale	KP 1 (ČSN EN 12668-1:2010)	
	Stability after heating	5 % SH	to	100 % SH		2(16/V <sub>S</sub> mm) %				
		5 % SW	to	100 % SW		2(16/SW mm) %				
	Display instability	5 % SH	to	100 % SH		2(15.3/SH mm) %				
		5 % SW	to	100 % SW		2(15.3/SW mm) %				
	Stability at voltage fluctuation	5 % SH	to	100 % SH		2(16/V <sub>S</sub> mm) %				
		5 % SW	to	100 % SW		2(16/SW mm) %				
	Vertical linearity	5 % SH	to	100 % SH		2.3 % SH				
	Time base linearity	1 mm		9,999 mm		0.05 %				
	Time base linearity of analog. instruments	1 mm		9,999 mm		2(18/SW mm) %				
	Equivalent input noise level	2 V/√Hz	to	100 V/√Hz		3.5 %				
	Transmit pulse overshoot	2 V	to	500 V		7.9 %				
	Transmitting pulse rise time, duration	3 ns	to	500 ns		7.9 %				
Filter frequency upper and lower limit	0.2 MHz	to	30 MHz		2.8 %					
Filter medium frequency	0.2 MHz	to	30 MHz		3.3 %					
Decibel attenuator accuracy	0 dB	to	120 dB		0.2 dB					
Frequency filter bandwidth	0.2 MHz	to	30 MHz		2.8 %					

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**TESTIMA, spol. s r.o.**  
TESTIMA Calibration Laboratory  
Husova 353/6, 250 01 Brandýs nad Labem

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work place
		min	unit					
2	Digital ultrasonic defectoscopes							
	Display vertical linearity	5 % SH	to	100 % SH		2.3 % SH		
	Time base deviation	1 mm		9,999 mm		0.05 %		
	Noise level	2 V/√Hz	to	100 V/√Hz		3.5 %		
	Transmit pulse voltage	2 V	to	500 V		7.9 %		
	Transmitting pulse rise time, duration	3 ns	to	500 ns		7.9 %		
	Upper and lower limit frequency	0.2 MHz	to	30 MHz		2.8 %		
	Medium frequency	0.2 MHz	to	30 MHz		3.3 %		
	Gain linearity	0 dB	to	120 dB		0.2 dB		
	Bandwidth	0.2 MHz	to	30 MHz		2.8 %		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value given here is based on the best laboratory conditions achievable; the uncertainty value of a particular calibration may be higher depending on the conditions of that calibration. For identical limit values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

SH scale height on the display of the calibrated defectoscope in mm

SW scale width on the display of the calibrated defectoscope in mm