

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Length**

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
1	Parallel gauge blocks	0.5 mm	to	100 mm		(0.8L + 0.14) μm (1L + 0.16) μm	Comparison with parallel gauge blocks	Kp 01-013		
2*	Slide gauges / slide rules, depth gauges, height gauges, gear tooth calipers	0 mm	to	1,000 mm		14 μm 17 μm	Measurement of parallel gauge blocks	Kp 01-001		
3*	Linear height gauges	0 mm	to	1,000 mm		(1.2L + 0.5) μm	Measurement of parallel gauge blocks	Kp 01-001		
4*	Micrometer gauges / micrometers, pasameters, micropasameters, micrometer depth gauges	0 mm	to	25 mm		0.7 μm	Measurement of parallel gauge blocks	Kp 01-002		
		25 mm	to	100 mm		1.4 μm				
		100 mm	to	1,000 mm		2.2 μm				
		1,000 mm	to	1,500 mm		3.8 μm				
	Two-contact and three-contact internal gauges	3 mm	to	200 mm		1.6 μm	Comparison with setting rings			
	Inside micrometer gauges	10 mm	to	1,500 mm		3.7 μm	Comparison with parallel gauge blocks			
5	Micrometer gauges / inside micrometer gauges	10 mm	to	500 mm		(1L + 0.4) μm	Direct measurement by a length gauge	Kp 01-002		

The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
	Micrometric heads	0 mm	to 100 mm		1 μm	Direct measurement on a linear height gauge		
	Setting gauges for micrometer gauges	0 mm	to 500 mm		(1L + 0.4) μm			
		0 mm	to 950 mm		1.8 μm			
6	Indicators / direct, lever indicators and internal gauges with indicator	0 mm	to 100 mm		0.5 μm	Direct measurement by a length gauge	Kp 01-003	
7*	Indicators / direct, lever indicators and internal gauges with indicator	0 mm	to 50 mm		2.9 μm	Direct measurement on a portable measuring device	Kp 01-003	
8	Linear sensors	0 mm	to 100 mm		0.5 μm	Direct measurement by a length gauge	Kp 01-003	
9	Rules / precise gauges and measuring magnifiers	0 mm	to 100 mm		0.5 μm	Direct measurement by a length gauge	Kp 01-004	
		0 mm	to 200 mm		2.1 μm	Direct measurement by a microscope		
		200 mm	to 400 mm		2.7 μm			
	400 mm	to 600 mm		3.7 μm				
	Steel rules	0 mm	to 1,000 mm		42 μm	Comparison with a standard gauge		
		1,000 mm	to 2,000 mm		59 μm			
	Tape measures	0 m	to 8 m		0.17 mm	Comparison with a standard track		
		8 m	to 10 m		0.32 mm			

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
	Tape measures, wooden rulers and rulers of 2m length	0 m	to 100 m		(0.015L + 0.18) mm			
10*	Rules / steel gauges	0 mm	to 500 mm		0.12 mm	Measurement of parallel gauge blocks	Kp 01-004, Kp 01-015	
	Tape measures	0 m	to 10 m		(0.07L + 0.19) mm	Comparison with a standard gauge	Kp 01-004	
11	Setting rings and snap gauges	0.95 mm	to 10 mm		1 μm	Direct and comparative measurement on a length gauge	Kp 01-005	
		10 mm	to 275 mm		(4.3L + 0.7) μm			
	Cylindrical gauges	0 mm	to 100 mm		0.5 μm			
		100 mm	to 500 mm		1 μm			
	Slot gauges	0 mm	to 100 mm		0.5 μm			
		100 mm	to 500 mm		1 μm			
	Feeler gauges and wedges	0 mm	to 100 mm		0.5 μm			
	Cylindrical gauges	0 mm	to 100 mm		0.5 μm			
	Measuring wires	0 mm	to 10 mm		0.5 μm			
	Setting gauges for layer thickness gauges	0 mm	to 25 mm		0.8 μm			
Thread gauges – plug gauges	0 mm	to 300 mm		3.1 μm				
Thread gauges - female	5 mm	to 200 mm		4 μm	Direct measurement by a microscope			
Thread gauges - conical	0 mm	to 100 mm		5 μm	Direct measurement on a length gauge and linear height gauge			

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
	Efflux viscometers	0 mm	to 10 mm		4 μm	Direct measurement by a microscope		
12*	Cylindrical gauges	0 mm	to 100 mm		2 μm	Direct measurement with a micropasameter or micrometer	Kp 01-005	
	Slot gauges	0 mm	to 100 mm		2 μm			
	Feeler gauges and wedges	0 mm	to 30 mm		1.5 μm			
	Cylindrical gauges	0 mm	to 30 mm		1.5 μm			
	Thread gauges – plug gauges	0 mm	to 75 mm		5 μm			
13*	Thickness gauges and callipers	0 mm	to 500 mm	external measurement	2.2 μm	Comparison with parallel gauge blocks or thickness standard	Kp 01-010	
		3 mm	to 500 mm	internal measurement	3.7 μm	Comparison with parallel gauge blocks or setting rings		
14*	Dry layer thickness gauges	0 mm	to 25 mm	dry layers	1.4 μm	Comparative measurement by a thickness reference standard	Kp 01-009	
15	Wet layer thickness gauges	0 mm	to 15 mm	wet layers	1 μm	Direct measurement on a length gauge or a microscope	Kp 01-009	
16*	Surface plates / flatness	0 mm	to 5 mm	length up to 300 mm up to 1,000 mm up to 2,000 mm	3.7 μm 6.2 μm 34 μm	Comparison with parallel gauge blocks	Kp 01-008	
	Surface rules / straightness	0 mm	to 5 mm	length up to 1,000 mm up to 2,000 mm	5.1 μm 6.2 μm	Comparison with parallel gauge blocks		
	Blade rules / straightness	0 mm	to 5 mm	length up to 100 mm up to 300 mm	2.2 μm 2.5 μm	Comparison with parallel gauge blocks		

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
				up to 500 mm up to 1,000 mm	2.8 μm 5.1 μm			
17	Roller length gauges	0 m	to 100 m		(0.2L + 10) mm	Direct measurement by a special measuring device	Kp 01-014	
18	Laser distance meters	0.5 m	to 8 m		0.3 mm	Comparison with a standard track	Kp 01-014	
19	Levelling rods	0 m	to 7 m		0.3 mm	Comparison with a standard track or standard tape measure	Kp 01-014	
20	Telescopic length gauges	0 m	to 7 m		0.3 mm	Comparison with a standard track	Kp 01-014	
21	Weld gauges	0 mm	to 20 mm		10 μm	Comparison with parallel gauge blocks	Kp 01-015	
22	90° angles - perpendicularity	0 mm	to 5 mm	longer side up to 100 mm up to 1,000 mm	2.8 μm (8L + 6.5) μm	Comparison with parallel gauge blocks and perpendicularity standard	Kp 02-001	
	- straightness	0 mm	to 5 mm	longer side up to 100 mm up to 1,000 mm	2.2 μm 5.1 μm			
	- parallelity	0 mm	to 5 mm		2.9 μm			
23	Measuring jigs and profile gauges	0 mm	to 500 mm		(1L + 0.4) μm	Direct measurement on a length gauge	Kp 01-017	
		500 mm	to 950 mm		2 μm	Direct measurement on a linear height gauge		
24*	Measuring jigs and profile gauges	0 mm	to 100 mm		2 μm	Direct measurement with a micropasameter or micrometer	Kp 01-017	
		100 mm	to 300 mm		12 μm	Direct measurement with a slide gauge		

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
		300 mm	to 2,000 mm		15 μm	Comparison with parallel gauge blocks		
		2 m	to 10 m		0.5 mm	Direct measurement by a standard tape measure		
25*	Length gauges	0 mm	to 1,000 mm		(1L + 0.14) μm	Comparison with parallel gauge blocks	Kp 01-011	
26*	Measuring microscopes and profile projectors	0 mm	to 100 mm	axes X, Y	1.1 μm	Comparison with a standard gauge	Kp 01-019	
		100 mm	to 200 mm	axes X, Y	1.5 μm			
		200 mm	to 500 mm	axes X, Y	4 μm			
		0 mm	to 200 mm	Z-axis	2 μm	Comparison with parallel gauge blocks		
		0 mm	to 500 mm	perpendicularity of the X and Y axes	10 μm	Comparison with a standard square		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Plane angle**

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
1*	Angle gauges	0°		to	360°		2.4′	Comparison with angle gauges	Kp 02-002, Kp 01-015	
2*	Measuring jigs and profile gauges	0°		to	360°		0.9′	Direct measurement by a microscope	Kp 01-017, Kp 01-015	
3*	Measuring microscopes and profile projectors	0°		to	360°		0.6′	Comparison with angle gauges	Kp 01-019	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
 Calibration Laboratory  
 Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Force, mechanical tests**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
1*	Torque measuring devices, torque wrenches and screwdrivers, pneumatic and electric nutrunners	0.15 Nm	to	2 Nm		Clockwise	0.67 %	Comparison with a torque sensor	Kp 03-001	
		2 Nm	to	10 Nm			0.56 %			
		10 Nm	to	100 Nm			0.46 %			
		100 Nm	to	1,000 Nm			0.49 %			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Pressure, mechanical stress**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity		Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit						
1*	Deformation pressure gauges, digital pressure gauges, pressure transducers and pressure measuring chains	-95 kPa	to	0 kPa	negative gauge pressure	gas	0.1 % + 61 Pa	Comparison with a pressure gauge	Kp 05-001		
		0 kPa	to	100 kPa	positive gauge pressure	gas	0.1 % + 64 Pa 0.1 % + 480 Pa 0.1 % + 1.1 kPa				
		0.1 MPa	to	0.7 MPa							
		0.7 MPa	to	1.7 MPa							
		0 MPa	to	0.7 MPa	positive gauge pressure	liquid	0.1 % + 480 Pa 0.1 % + 1.1 kPa 0.1 % + 6.6 kPa 0.1 % + 46 kPa 0.72 MPa				
		0.7 MPa	to	1.7 MPa							
		1.7 MPa	to	7 MPa							
		7 MPa	to	70 MPa							
		70 MPa	to	100 MPa							

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Temperature**

Ord. number <sub>1</sub>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
1*	Indicating thermometers and temperature measuring chains	-20 °C	to	-5 °C			0.34 °C	Comparison with a thermometer in a calibrating oven	Kp 07-001	
		-5 °C	to	50 °C			0.27 °C			
		50 °C	to	100 °C			0.48 °C			
		100 °C	to	650 °C			0.64 °C			
2	Non-contact thermometers	30 °C	to	100 °C			1.7 °C	Comparison with a target black body	Kp 07-002	
		100 °C	to	200 °C			2.6 °C			
		200 °C	to	300 °C			2.7 °C			
		300 °C	to	400 °C			3.0 °C			
		400 °C	to	500 °C			3.3 °C			
3*	Simulation of temperature sensor signals / temperature sensor processing units	-210 °C	to	-100 °C	J		0.57 °C	Voltage calibrator simulation including cold junction compensation	Kp 04-001	
		-100 °C	to	150 °C			0.33 °C			
		150 °C	to	760 °C			0.38 °C			
		760 °C	to	1,200 °C			0.48 °C			
		-200 °C	to	-100 °C	K		0.66 °C			
		-100 °C	to	120 °C			0.40 °C			
		120 °C	to	1,370 °C			0.63 °C			
		-250 °C	to	-150 °C	T		1.5 °C			
		-150 °C	to	400 °C			0.37 °C			
		0 °C	to	250 °C	R		1.9 °C			
		250 °C	to	1,760 °C			1.2 °C			

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number 1	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work- place
		min.	unit					
		0 °C	to	250 °C	S	1.9 °C		
		250 °C	to	1,760 °C		1.2 °C		
		600 °C	to	1,820 °C	B	1.8 °C		
		-200 °C	to	-100 °C	N	1.0 °C		
		-100 °C	to	410 °C		0.51 °C		
		410 °C	to	1,300 °C		0.59 °C		
		-250 °C	to	-100 °C	E	1.3 °C		
		-100 °C	to	650 °C		0.47 °C		
		650 °C	to	1,000 °C		0.51 °C		
		-200 °C	to	900 °C	L	0.81 °C		
		-200 °C	to	600 °C	U	0.95 °C		
		0 °C	to	1,000 °C	C	0.72 °C		
		1,800 °C	to	2,310 °C		1.7 °C		
		-200 °C	to	0 °C	Pt 100	0.35 °C	Resistance calibrator simulation	
		0 °C	to	800 °C		0.58 °C		

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Electrical quantities**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
1*	DC voltage sources	0 mV	to	100 mV		0.0037 % + 13 μV	Direct measurement by a standard multimeter	Kp 04-001		
		0.1 V	to	1 V		0.0025 % + 37 μV				
		1 V	to	10 V		0.0024 % + 0.36 mV				
		10 V	to	100 V		0.0038 % + 3.6 mV				
		100 V	to	1,000 V		0.0041 % + 36 mV				
	DC voltage meters	0 mV	to	100 mV		0.008 % + 12 μV	Direct generation with a standard calibrator			
		0.1 V	to	1 V		0.008 % + 35 μV				
		1 V	to	10 V		0.008 % + 0.35 mV				
		10 V	to	100 V		0.008 % + 3.5 mV				
		100 V	to	1,000 V		0.008 % + 35 mV				
2*	Direct current sources	0 μA	to	100 μA		0.05 % + 37 nA	Direct measurement by a standard multimeter			
		0.1 mA	to	1 mA		0.05 % + 0.12 μA				
		1 mA	to	10 mA		0.05 % + 2.5 μA				
		10 mA	to	100 mA		0.05 % + 14 μA				
		100 mA	to	400 mA		0.05 % + 73 μA				
		0.4 A	to	1 A		0.05 % + 0.3 mA				
		1 A	to	3 A		0.10 % + 0.8 mA				
		3 A	to	10 A		0.15 % + 1.6 mA				
	Direct current meters	0 μA	to	100 μA		0.03 % + 35 nA	Direct generation with a standard calibrator			
		0.1 mA	to	1 mA		0.03 % + 0.12 μA				
		1 mA	to	10 mA		0.03 % + 1.2 μA				

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
		10 mA	to	100 mA		0.03 % + 12 μA				
		0.1 A	to	1 A		0.03 % + 0.18 mA				
		1 A	to	10 A		0.05 % + 2.4 mA				
		0 A	to	500 A		0.50 % + 0.49 A	Simulation using current coil			
3*	AC voltage sources	0.1 mV	to	100 mV	10 Hz to 20 kHz	0.06 % + 54 μV	Direct measurement by a standard multimeter	Kp 04-001		
		0.1 V	to	1 V	10 Hz to 20 kHz	0.06 % + 0.36 mV				
		1 V	to	10 V	10 Hz to 20 kHz	0.06 % + 3.6 mV				
		10 V	to	100 V	10 Hz to 20 kHz	0.06 % + 37 mV				
		100 V	to	1,000 V	10 Hz to 20 kHz	0.06 % + 0.29 V				
	AC voltage meters	0.1 mV	to	100 mV	10 Hz to 2 kHz	0.08 % + 43 μV	Direct generation with a standard calibrator			
		0.1 V	to	1 V	10 Hz to 2 kHz	0.08 % + 0.39 mV				
		1 V	to	10 V	10 Hz to 2 kHz	0.08 % + 4.0 mV				
		10 V	to	100 V	40 Hz to 1 kHz	0.08 % + 43 mV				
		100 V	to	1,000 V	40 Hz to 1 kHz	0.08 % + 0.60 V				
4*	Alternating current sources	0.1 μA	to	100 μA	10 Hz to 2 kHz	0.15 % + 80 nA	Direct measurement by a standard multimeter			
		0.1 mA	to	1 mA	10 Hz to 2 kHz	0.10 % + 0.54 μA				
		1 mA	to	10 mA	10 Hz to 2 kHz	0.15 % + 7.5 μA				
		10 mA	to	100 mA	10 Hz to 2 kHz	0.10 % + 56 μA				
		100 mA	to	400 mA	10 Hz to 1 kHz	0.10 % + 0.51 mA				
		0.4 A	to	1 A	10 Hz to 2 kHz	0.10 % + 0.8 mA				
		1 A	to	3 A	10 Hz to 2 kHz	0.15 % + 2.4 mA				
		3 A	to	10 A	10 Hz to 2 kHz	0.15 % + 16 mA				

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit					
	Alternating current meters	0.1 μA	to	100 μA	10 Hz to 2 kHz	0.1 % + 0.47 μA		
		0.1 mA	to	1 mA	10 Hz to 2 kHz	0.1 % + 0.96 μA		
		1 mA	to	10 mA	10 Hz to 2 kHz	0.1 % + 9.5 μA		
		10 mA	to	100 mA	10 Hz to 2 kHz	0.1 % + 96 μA		
		0.1 A	to	1 A	10 Hz to 2 kHz	0.1 % + 0.95 mA		
		1 A	to	10 A	10 Hz to 2 kHz	0.1 % + 20 mA		
		0 A	to	500 A	30 Hz to 60 Hz	0.34 % + 0.071 A	Simulation using current coil	
5*	DC resistance / resistors and resistance boxes	0 Ω	to	10 Ω		0.01 % + 12 mΩ	Kp 04-001 Kp 04-002	
		10 Ω	to	100 Ω		0.01 % + 58 mΩ		
		100 Ω	to	1 kΩ		0.01 % + 59 mΩ		
		1 kΩ	to	10 kΩ		0.01 % + 0.17 Ω		
		10 kΩ	to	100 kΩ		0.01 % + 2.1 Ω		
		0.1 MΩ	to	1 MΩ		0.01 % + 32 Ω		
		1 MΩ	to	10 MΩ		0.04 % + 1.3 kΩ		
		10 MΩ	to	100 MΩ		0.8 % + 35 kΩ		
		100 MΩ	to	1 GΩ		2.0 % + 0.21 MΩ		
			0.01 Ω	to	0.1 Ω			
		0.1 Ω	to	1 Ω		0.052 %		
		1 Ω	to	10 Ω		0.14 %		
	DC resistance meters	0 Ω	to	10 Ω		20 mΩ	Kp 04-001	
		10 Ω	to	100 Ω		40 mΩ		
		100 Ω	to	1 kΩ		0.27 Ω		
		1 kΩ	to	10 kΩ		2.4 Ω		
		10 kΩ	to	100 kΩ		24 Ω		

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
		0.1 MΩ	to	1 MΩ		0.24 Ω				
		1 MΩ	to	10 MΩ		5.8 kΩ				
		0.1 MΩ	to	0.5 MΩ		0.05 % + 10 Ω	Direct generation using a standard resistance box			
		0.6 MΩ	to	0.9 MΩ		0.05 % + 15 Ω				
		1 MΩ	to	5 MΩ		0.05 % + 0.25 kΩ				
		6 MΩ	to	9 MΩ		0.05 % + 0.35 kΩ				
		10 MΩ	to	50 MΩ		0.05 % + 2.5 kΩ				
		50 MΩ	to	100 MΩ		0.05 % + 9 kΩ				
6	Inspection equipment / insulation resistance meters	10 kΩ	to	100 kΩ	Measuring voltage up to 100 V	0.05 % + 13 Ω	Direct generation using a standard resistance box	Kp 04-003		
		0.1 MΩ	to	0.5 MΩ		0.05 % + 10 Ω				
		0.6 MΩ	to	0.9 MΩ		0.05 % + 15 Ω				
		1 MΩ	to	5 MΩ	Measuring voltage up to 500 V	0.05 % + 0.25 kΩ				
6 MΩ	to	9 MΩ	0.05 % + 0.35 kΩ							
		10 MΩ	to	50 MΩ	Measuring voltage up to 1,000 V	0.1 % + 2.5 kΩ				
		60 MΩ	to	100 MΩ		0.1 % + 9 kΩ				
	meters of transition resistance	0.1 Ω	to	1 Ω		0.2 % + 3.7 mΩ				
		1 Ω	to	10 Ω		0.1 % + 4.7 mΩ				
		10 Ω	to	100 Ω		0.05 % + 17 mΩ				
		100 Ω	to	1,000 Ω		0.05 % + 0.13 Ω				
		1 kΩ	to	10 kΩ		0.05 % + 1.3 Ω				

**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min.	unit	max.	unit					
	meters of leakage current	0.1 mA	to	1 mA	50 Hz to 60 Hz	0.10 % + 0.54 μA	Direct measurement by a standard ammeter			
		1 mA	to	10 mA	50 Hz to 60 Hz	0.15 % + 7.5 μA				
		10 mA	to	100 mA	50 Hz to 60 Hz	0.10 % + 56 μA				

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



**The Appendix is an integral part of  
Certificate of Accreditation No. 574/2021 of 8/11/2021**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**KZB-Kalibrace s.r.o.**  
Calibration Laboratory  
Mikoláše Alšed 2240, 434 01 Most

**CMC for the field of measured quantity: Time and frequency quantities**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Work-place
		min. unit	max. unit					
1*	Time interval / mechanical and digital stopwatch, timers and other time meters	5 s	to 3,600 s		11 ms	Comparison with a standard counter	Kp 06-001	

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).