

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

LESAK s.r.o.
Facility No. 2411, Calibration Laboratory
Božetěchova 2826/36, Královo Pole, 612 00 Brno

CMC for the field of measured quantity: Mass

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified ²	Calibration principle	Calibration procedure identification ³	Work-place
		min	unit	max	unit					
1*	Scales with non-automatic function with automatic indication (maximum number of divisions up to 1,200,000; <i>d</i> value greater than 1mg)	0 kg	to	6 kg		9.2·10 ⁻⁷	Loading using a reference weight of class E2	KP1		
		6 kg	to	30 kg		2.9·10 ⁻⁶	class F1			
		30 kg	to	70 kg		9.2·10 ⁻⁶	class F2			
		70 kg	to	10,000 kg		2.9·10 ⁻⁵	class M1			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. The uncertainty value given here is based on the best laboratory conditions achievable; the uncertainty value of a particular calibration may be higher depending on the conditions of that calibration. For identical limit values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).