

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

EG.D, a.s.
Calibration Laboratory
Čsl. armády 2114, 390 03 Tábor

CMC for the field of measured quantity: Electrical quantities

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified ²	Calibration principle	Calibration procedure identification ³	Work-place
		min. unit	max. unit					
1	AC electrical work / Single-phase and three-phase active and reactive, induction and static energy meters	3 Ws	to 500 kWh	60 V to 500 V / 0,005 A to 125 A Cos $\varphi = 1$ Cos $\varphi = 0.5$ Sin $\varphi = 1$ Sin $\varphi = 0.5$	0.062 % 0.064 % 0.12 % 0.13 %	Comparison with a reference energy meter	KPM 1	

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02:M, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).