



Český institut pro akreditaci, o.p.s.

Hájkova 2747/22, 130 00 Praha 3

Strana: 1/62

Zpracoval:

Ing. Martin Valenta

Datum zpracování:

20.10.2025

Plán standardizace – Program rozvoje metrologie 2025

Číslo úkolu: VII/05/25

Zpráva pro závěrečnou oponenturu úkolu

Principy kalibrace a další aspekty kalibrace v oborech mechanický pohyb,
vibrace a akustika

Příloha 2

Přehled příloh osvědčení o akreditaci zahraničních kalibračních laboratoří



OBSAH

OBSAH	2
1. A2LA (USA)	4
1.1. ELEMENT MATERIALS TECHNOLOGY HUNTSVILLE	4
1.2. ELEMENT U.S. SPACE & DEFENSE	5
2. ACCREDIA(Itálie).....	6
2.1. L.C.E. s.r.l.	6
2.2. Azienda USL TOSCANA SUD EST, U.F. Laboratorio di Sanità Pubblica.....	9
2.3. TRESICAL s.r.l.	10
3. NATA (Austrálie)	11
3.1. Noise and Vibration Measurement Systems Pty Ltd, Sydney Calib. Lab.....	11
3.2. National Measurement Institute (NMI), Acoustics, Ultrasonics and Vibration Project, Lindfield.....	12
3.3. ACU-VIB Electronics, ACU-VIB Electronics.....	13
4. COFRAC (Francie)	17
4.1. ACU-VIB Electronics, ACU-VIB Electronics.....	17
4.2. LABORATOIRE NATIONAL DE METROLOGIE ET D'ESSAIS.....	17
4.3. LABORATOIRE NATIONAL DE METROLOGIE ET D'ESSAIS.....	20
4.4. TRESICAL	21
4.5. HOTTINGER BRUEL & KJAER FRANCE SAS	23
5. DAkKS.....	25
5.1. SPEKTRA	25
5.2. Norsonic-Tippkemper GmbH	33
5.3. Trescal GmbH	38
5.4. Hottinger Brüel & Kjaer GmbH.....	40
5.5. Kistler Instrumente Gesellschaft mit beschränkter Haftung	43
6. ENAC (Španělsko).....	44
6.1. INSTITUTO NACIONAL DE TECNICA AEROESPACIAL “ESTEBAN TERRADAS”	44
6.2. HOTTINGER BRÜEL & KJAER IBÉRICA, S.L.....	45
6.3. TRESICAL IBERICA DE CALIBRACION, S.L.....	49



Český institut pro akreditaci, o.p.s.

Hájkova 2747/22, 130 00 Praha 3

Strana: 3/62

Zpracoval:

Ing. Martin Valenta

Datum zpracování:

20.10.2025

7.	PCA (Polsko)	51
7.1.	GLÓWNY INSTYTUT GÓRNICHTWA – PAŃSTWOWY INSTYTUT BADAWCZY	51
7.2.	OKRĘGOWY URZĄD MIAR w BIAŁYMSTOKU	53
8.	UKAS (Velká Británie)	54
8.1.	ACOEM EM	54
8.2.	BAE Systems Marine Limited	59
8.3.	University of Salford	59
8.1.	ACOEM EM	54



1. A2LA (USA)

1.1. ELEMENT MATERIALS TECHNOLOGY HUNTSVILLE

III. Mechanical

Acceleration – Measuring Equipment	(0.5 to 1) Hz	1.2 %	Standard low frequency accelerometer. Data acquisition card w/ low frequency long stroke shaker
	(> 1 to 5) Hz	1.1 %	
	(5 to 9) Hz	2.3 %	
	(10 to 99) Hz	1.6 %	
	100 Hz	1.2 %	
	(> 100 to 920) Hz	1.4 %	
	920 Hz to 5 kHz	1.8 %	
	(> 5 to 10) kHz	2.3 %	Standard accelerometer Data acquisition card w/ air bearing shaker TMS9155C
Acceleration (Shock) – Measuring Equipment	(20 to 10 000) g	2.0 %	TMS9155C



1.2. ELEMENT U.S. SPACE & DEFENSE

I. Mechanical

Parameter/Equipment	Frequency	CMC ^{2,3} (±)	Comments
Acceleration Sensitivity	(5 to 9.99) Hz (10 to 99) Hz 100 Hz (101 to 920) Hz (921 to 5000) Hz (5001 to 10 000) Hz	1.6 % 1.0 % 0.90 % 1.0 % 1.7 % 2.5 %	Back to back comparison

¹ This laboratory is not normally available for commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards of nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k=2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ In the statement of CMC, the value is defined as the percentage of reading.



2. ACCREDIA (Itálie)

2.1. L.C.E. s.r.l.

Area metrologica Metrological area		Accelerazione						
Settore / Calibration field (SAC-01) Catena accelerometrica								
Strumento Instrument	Misurando Measurand	Campo di misura Measurement range		Condizioni Additional parameters		Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Catene accelerometriche con trasduttori a singola e doppia faccia	Accelerazione (sensibilità in tensione o in carica)	9,81 m·s ⁻²		DC		1,2 %	Metodo interno. Taratura per confronto con accelerazione di gravità	A
		≥ 0,5 m·s ⁻²	≤ 300 m·s ⁻²	≥ 1 Hz	≤ 5 Hz	3,4 %		
				> 5 Hz	≤ 20 Hz	2,0 %		
				> 20 Hz	≤ 1 kHz	1,5 %		
				> 1 kHz	≤ 5 kHz	2,0 %		
				> 5 kHz	≤ 10 kHz	3,8 %		
	Velocità (sensibilità in tensione o in carica) (2)	≥ 0,005 m·s ⁻¹	≤ 0,08 m·s ⁻¹	≥ 1 Hz	≤ 5 Hz	3,4 %		
				> 5 Hz	≤ 20 Hz	2,0 %		
				> 20 Hz	≤ 1 kHz	1,5 %		
				> 1 kHz	≤ 5 kHz	2,0 %		
				> 5 kHz	≤ 10 kHz	3,8 %		
				> 5 kHz	≤ 10 kHz	3,4 %		
Spostamento (sensibilità in tensione o in carica) (2)	≥ 1 μm	≤ 12 mm	≥ 1 Hz	≤ 5 Hz	3,4 %			
			> 5 Hz	≤ 20 Hz	2,0 %			
			> 20 Hz	≤ 1 kHz	1,5 %			
			> 1 kHz	≤ 5 kHz	2,0 %			
			> 5 kHz	≤ 10 kHz	3,8 %			
			> 5 kHz	≤ 10 kHz	3,8 %			

(Continua) Area metrologica "Misure acustiche"

Settore / Calibration field (SAU-02) Sensibilità assoluta alla pressione acustica						
Strumento Instrument	Misurando Measurand	Condizioni Additional parameters	Campo di misura Measurement range	Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Microfoni	(14) Sensibilità alla pressione acustica	250 Hz	124 dB	0,09 dB	Metodo interno. Taratura per confronto in accoppiatore	A
			94 dB	da 0,11 dB a 0,22 dB		
			94 dB	da 0,11 dB a 0,22 dB		
	(15) Sensibilità alla pressione acustica (corretta per campo libero)	da 31,5 Hz a 16 kHz	94 dB	da 0,12 dB a 0,83 dB	Metodo interno. Taratura basata su IEC 61094-5:2016 e CEI EN 61094-5:2017	

Area metrologica Metrological area		Misure acustiche				
Settore / Calibration field (SAU-01) Livello di pressione acustica						
Strumento Instrument	Misurando Measurand	Condizioni Additional parameters	Campo di misura Measurement range	Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Pistonofoni	Livello di pressione acustica	da 160 Hz a 1,25 kHz	da 94 dB a 140 dB	0,10 dB	IEC EN 60942:2018 Annex B CEI EN 60942:2018 Annex B	A
	Frequenza	da 94 dB a 140 dB	da 160 Hz a 1,25 kHz	0,04 %		
Calibratori	Livello di pressione acustica	da 160 Hz a 1,25 kHz	da 90 dB a 125 dB	0,10 dB		
	Frequenza	da 94 dB a 140 dB	da 160 Hz a 1,25 kHz	0,04 %		
Calibratori multifrequenza	Livello di pressione acustica	da 31,5 Hz a 16 kHz	da 94 dB a 140 dB	da 0,10 dB a 0,49 dB	Metodo interno. Taratura mediante rapporto di livello di pressione acustica	
	Ponderazione "inversa A" (6)	da 31,5 Hz a 16 kHz	94 dB	0,15 dB		
	Correzioni pressione / campo libero microfoni (6)	da 31,5 Hz a 16 kHz	da 94 dB a 114 dB	0,12 dB		

(continua)



Český institut pro akreditaci, o.p.s.
Hájkova 2747/22, 130 00 Praha 3

Strana: 7/62
Zpracoval:
Ing. Martin Valenta
Datum zpracování:
20.10.2025

(Continua) Area metrologica "Accelerazione"

Settore / Calibration field		(SAC-02) Calibratori/Generatori						
Strumento Instrument	Misurando Measurand	Condizioni Additional parameters		Campo di misura Measurement range		Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Calibratore vibrometrico	Accelerazione	≥ 15 Hz	≤ 20 Hz	$\geq 0,5$ m·s ⁻²	≤ 15 m·s ⁻²	1,6 %	Metodo interno. Taratura per confronto con accelerometro tarato secondo la ISO 16063-21:2003	A
		> 20 Hz	≤ 1 kHz			1,3 %		
	Frequenza	$\geq 0,5$ m·s ⁻²	≤ 15 m·s ⁻²	≥ 15 Hz	≤ 1 kHz	0,08 %		

(Continua) Area metrologica "Accelerazione" – Settore "Catena accelerometrica" (SAC-01)

Strumento Instrument	Misurando Measurand	Campo di misura Measurement range		Condizioni Additional parameters		Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Analizzatori con trasduttore accoppiato	Accelerazione (sensibilità in tensione o in carica)	9,81 m·s ⁻²		DC		1,3 %	Metodo interno. Taratura per confronto con accelerazione di gravità ISO 16063-21:2003 (3)	A
		$\geq 0,5$ m·s ⁻²	≤ 300 m·s ⁻²	≥ 1 Hz	≤ 5 Hz	3,5 %		
				> 5 Hz	≤ 20 Hz	2,2 %		
				> 20 Hz	≤ 1 kHz	1,7 %		
				> 1 kHz	≤ 5 kHz	2,1 %		
				> 5 kHz	≤ 10 kHz	3,9 %		
	≥ 1 Hz			≤ 5 Hz	3,5 %			
	Velocità (sensibilità in tensione o in carica) (4)	$\geq 0,005$ m·s ⁻¹	$\leq 0,08$ m·s ⁻¹	> 5 Hz	≤ 20 Hz	2,2 %		
				> 20 Hz	≤ 1 kHz	1,7 %		
				> 1 kHz	≤ 5 kHz	2,1 %		
				> 5 kHz	≤ 10 kHz	3,9 %		
				≥ 1 Hz	≤ 5 Hz	3,5 %		
> 5 Hz				≤ 20 Hz	2,2 %			
Spostamento (sensibilità in tensione o in carica) (4)	≥ 1 μm	≤ 12 mm	> 20 Hz	≤ 1 kHz	1,7 %			
			> 1 kHz	≤ 5 kHz	2,1 %			
			> 5 kHz	≤ 10 kHz	3,9 %			
			≥ 1 Hz	≤ 5 Hz	3,5 %			
			> 5 Hz	≤ 20 Hz	2,2 %			
			> 20 Hz	≤ 1 kHz	1,7 %			
Analizzatori con trasduttore mano braccio	Accelerazione	≥ 1 m·s ⁻²	≤ 100 m·s ⁻²	≥ 10 Hz	≤ 800 Hz	$2,5 \cdot 10^{-2}$	ISO 8041-1:2017 §14	
Analizzatori con trasduttore corpo intero		$\geq 0,1$ m·s ⁻²	≤ 15 m·s ⁻²	≥ 2 Hz	≤ 80 Hz	$2,5 \cdot 10^{-2}$		

(Continua) Area metrologica "Misure acustiche" – Settore "Livello di pressione acustica" (SAU-01)

Strumento Instrument	Misurando Measurand	Condizioni Additional parameters		Campo di misura Measurement range		Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Pistonofoni	Livello di pressione acustica	da 160 Hz a 315 Hz		da 114 dB a 140 dB		0,10 dB	Metodo interno Taratura basata su IEC EN 60942:2003 Annex B CEI EN 60942:2004 Annex B	A
	Frequenza	da 114 dB a 140 dB		da 160 Hz a 315 Hz		0,04 %		
Calibratori	Livello di pressione acustica	da 160 Hz a 1,25 kHz		da 94 dB a 114 dB		0,10 dB		
	Frequenza	da 94 dB a 114 dB		da 160 Hz a 1,25 kHz		0,05 %		
Calibratori multifrequenza	Livello di pressione acustica	da 31,5 Hz a 16 kHz		da 94 dB a 114 dB		da 0,10 dB a 0,49 dB	Metodo interno. Taratura mediante rapporto di livello di pressione acustica	
	Ponderazione "inversa A" (8)	da 31,5 Hz a 16 kHz		94 dB		0,15 dB		
	Correzioni pressione / campo libero microfoni (8)	da 31,5 Hz a 16 kHz		da 94 dB a 114 dB		0,12 dB		

(continua)

¹ Con esclusione dello sfasamento. (S výjimkou fázového posunu.)

² I valori di campo di misura e di incertezza a fianco riportati sono determinati per calcolo a partire dai valori di accelerazione. (Hodnoty měřičiho rozsahu a nejistoty uvedené vedle jsou určeny výpočtem na základě hodnot zrychlení.)

³ Con esclusione dello sfasamento. (S výjimkou fázového posunu.)

⁴ I valori di campo di misura e di incertezza a fianco riportati sono determinati per calcolo a partire dai valori di accelerazione. (Hodnoty měřičiho rozsahu a nejistoty uvedené vedle jsou určeny výpočtem na základě hodnot zrychlení.)

⁵ Conformi alle norme IEC EN 60942:2018 e CEI EN 60942:2018. (V souladu s normami IEC EN 60942:2018 a CEI EN 60942:2018.)



(Continua) Area metrologica "Misure acustiche" – Settore "Livello di pressione acustica" (SAU-01)

Strumento Instrument	Misurando Measurand	Condizioni Additional parameters	Campo di misura Measurement range	Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
(continua)						
Fonometri	Livello di pressione acustica	da 31,5 Hz a 16 kHz	da 20 dB a 155 dB	da 0,14 dB a 0,84 dB	Metodo interno. Taratura basata su Guida CEI 29-30:1997	A
		da 63 Hz a 16 kHz	da 20 dB a 150 dB	da 0,07 dB a 0,45 dB	Metodo interno. Taratura basata su IEC 61672-3:2006 e CEI EN 61672-3:2007	
		da 63 Hz a 16 kHz	da 20 dB a 150 dB	da 0,07 dB a 0,45 dB	IEC 61672-3:2013; CEI EN 61672-3:2014	
		da 20 Hz a 20 kHz	da 20 dB a 150 dB	da 0,10 dB a 1,0 dB	Metodo interno. Taratura basata su IEC EN 61260:1995 e CEI EN 61260:1997	
		da 31,5 Hz a 8 kHz	da 20 dB a 150 dB	da 0,10 dB a 1,0 dB		
Filtri a bande di terzi di ottava		da 20 Hz a 20 kHz	da 20 dB a 150 dB	da 0,19 dB a 0,50 dB	IEC 61260-3:2016; CEI EN 61260-3:2017	
Filtri a bande di ottava		da 20 Hz a 20 kHz	da 20 dB a 150 dB	da 0,19 dB a 0,50 dB		
Filtri a bande di terzi di ottava		da 20 Hz a 20 kHz	da 20 dB a 150 dB	da 0,19 dB a 0,50 dB		
Filtri a bande di ottava		da 31,5 Hz a 16 kHz	da 20 dB a 150 dB	da 0,19 dB a 0,50 dB		

⁶ Verifica di specifiche per calibratori multifunzione non previste dalle norme IEC EN 60942:2018 e CEI EN 60942:2018. (Ověření specifikací pro multifunkční kalibrátory, které nejsou stanoveny normami IEC EN 60942:2018 a CEI EN 60942:2018.)

⁷ Conformi alle norme IEC EN 60942:2003 e CEI EN 60942:2004. (V souladu s normami IEC EN 60942:2003 a CEI EN 60942:2004.)

⁸ Verifica di specifiche per calibratori multifunzione non previste dalle norme IEC EN 60942:2003 e CEI EN 60942:2004. (Ověření specifikací pro multifunkční kalibrátory, které nejsou stanoveny normami IEC EN 60942:2003 a CEI EN 60942:2004.)

⁹ Conformi alle norme EN 60651:2001, IEC 60804:2000 e CEI EN 60804:2001. (V souladu s normami EN 60651:2001, IEC 60804:2000 a CEI EN 60804:2001.)

¹⁰ Conformi alle norme IEC 61672-1:2002 ed. 1 e CEI EN 61672-1:2003 ed. 1. (V souladu s normami IEC 61672-1:2002 ed. 1 a CEI EN 61672-1:2003 ed. 1.)

¹¹ Conformi alle norme IEC 61672-1:2013 ed. 2 e CEI EN 61672-1:2014 ed. 2. (V souladu s normami IEC 61672-1:2013 ed. 2 a CEI EN 61672-1:2014 ed. 2.)

¹² Conformi alle norme IEC 61260:1995 e CEI EN 61260:1997. (V souladu s normami IEC 61260:1995 a CEI EN 61260:1997.)

¹³ Conformi alle norme IEC 61260-1:2014 e CEI EN 61260-1:2017. (V souladu s normami IEC 61260-1:2014 a CEI EN 61260-1:2017.)

¹⁴ Conformi alle norme IEC 61094-1:2000 e CEI EN 61094-1:2001. (V souladu s normami IEC 61094-1:2000 a CEI EN 61094-1:2001.)

¹⁵ Conformi alle norme IEC 61094-4:1995 e CEI EN 61094-4:1997. (V souladu s normami IEC 61094-4:1995 a CEI EN 61094-4:1997.)



2.2. Azienda USL TOSCANA SUD EST, U.F. Laboratorio di Sanità Pubblica

Area metrologica Metrological area		Misure acustiche				
Settore / Calibration field		(SAU-01) Livello di pressione acustica				
Strumento Instrument	Misurando Measurand	Condizioni Additional parameters	Campo di misura Measurement range	Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Pistonofoni (1)	Livello di pressione acustica	250 Hz	124 dB	0,1 dB	IEC EN 60942:2018 Annex B CEI EN 60942:2018 Annex B	A
	Frequenza	124 dB	250 Hz	0,01 %		
Calibratori	Livello di pressione acustica	250 Hz, 1 kHz	da 94 dB a 114 dB	0,13 dB		
	Frequenza	da 94 dB a 114 dB	250 Hz, 1 kHz	0,01 %		
Pistonofoni (2)	Livello di pressione acustica	250 Hz	124 dB	0,1 dB	Metodo interno. Taratura basata su IEC EN 60942:2003 Annex B e CEI EN 60942:2004 Annex B	A
	Frequenza	124 dB	250 Hz	0,02 %		
Calibratori	Livello di pressione acustica	250 Hz, 1 kHz	da 94 dB a 114 dB	0,13 dB		
	Frequenza	da 94 dB a 114 dB	250 Hz, 1 kHz	0,02 %		

(continua)

(Continua) Area metrologica "Misure acustiche" – Settore "Livello di pressione acustica" (SAU-01)

Strumento Instrument	Misurando Measurand	Condizioni Additional parameters	Campo di misura Measurement range	Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Fonometri (3)	Livello di pressione acustica	da 31,5 Hz a 16 kHz	da 25 dB a 140 dB	da 0,15 dB a 1,5 dB	Guida CEI 29-30:1997	A
	Livello di pressione acustica (4)	da 31,5 Hz a 16 kHz	da 25 dB a 140 dB	da 0,15 dB a 1,52 dB	Metodo interno. Taratura basata su IEC 61672-3:2006 e CEI EN 61672-3:2007	
		Livello di pressione acustica (5)	da 31,5 Hz a 16 kHz	da 25 dB a 140 dB	da 0,08 dB a 0,8 dB	
Filtri a bande di terzi di ottava (6)	Livello di pressione acustica	da 20 Hz a 20 kHz	da 25 dB a 140 dB	da 0,2 dB a 0,5 dB	IEC 61260-1:2014 CEI EN 61260-1:2017	A
		da 31,5 Hz a 16 kHz	da 25 dB a 140 dB	da 0,2 dB a 0,5 dB		
Filtri a bande di terzi di ottava (7)	Livello di pressione acustica	da 20 Hz a 20 kHz	da 25 dB a 140 dB	da 0,2 dB a 2,0 dB	Metodo interno. Taratura basata su IEC 61260:1995 e CEI EN 61260:1997	A
		da 31,5 Hz a 16 kHz	da 25 dB a 140 dB	da 0,2 dB a 2,0 dB		

Settore / Calibration field		(SAU-02) Sensibilità assoluta alla pressione acustica				
Strumento Instrument	Misurando Measurand	Condizioni Additional parameters	Campo di misura Measurement range	Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Microfoni a condensatore (8)	Sensibilità alla pressione acustica	250 Hz	124 dB	0,12 dB	Metodo interno. Taratura per confronto in accoppiatore	A
		1 kHz		0,13 dB		
	Working Standard da ½" (9)	Risposta di frequenza in pressione	da 31,5 Hz a 16 kHz	94 dB		

¹ Conformi alle norme IEC EN 60942:2018 e CEI EN 60942:2018. (V souladu s normami IEC EN 60942:2018 a IEC EN 60942:2018.)

² Conformi alle norme IEC EN 60942:2003 e CEI EN 60942:2004. (Vyhovuje IEC EN 60942:2003 a IEC EN 60942:2004.)

³ Conformi alle norme CEI EN 60651:2001 / IEC 60651:1979, CEI EN 60804:2001 / IEC 60804:2000. (Vyhovuje IEC EN 60651:2001 / IEC 60651:1979, IEC EN 60804:2001 / IEC 60804:2000.)

⁴ Conformi alle norme IEC 61672-1:2002 ed. 1 e CEI EN 61672-1:2003 ed. 1. (Vyhovuje IEC 61672-1:2002 ed. 1 a IEC EN 61672-1:2003 ed. 1.)

⁵ Conformi alle norme IEC 61672-1:2013 ed. 2 e CEI EN 61672-1:2014 ed. 2. (Vyhovuje IEC 61672-1:2013 ed. 2 a IEC EN 61672-1:2014 ed. 2.)

⁶ Conformi alla norma IEC 61260-1:2014 e CEI EN 61260-1:2017. (Vyhovuje normě IEC 61260-1:2014 a ČSN EN 61260-1:2017.)

⁷ Conformi alla norma CEI EN 61260:1997 / IEC EN 61260:1995. (Vyhovuje normě IEC EN 61260:1997 / IEC EN 61260:1995.)



2.3. TRESICAL s.r.l.

Area metrologica Metrological area		Misure acustiche				
Settore / Calibration field		(SAU-01) Livello di pressione acustica				
Strumento Instrument	Misurando Measurand	Condizioni Additional parameters	Campo di misura Measurement range	Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Pistonofoni (17)	Livello di pressione acustica	250 Hz	124 dB	0,10 dB	Metodo interno. Taratura basata su IEC EN 60942:2003 Annex B CEI EN 60942:2004 Annex B	A
	Frequenza	124 dB	250 Hz	0,012 %		
Calibratori	Livello di pressione acustica	250 Hz, 1 kHz	da 94 dB a 114 dB	0,10 dB		
	Frequenza	da 94 dB a 114 dB	250 Hz, 1 kHz	0,012 %		
Pistonofoni (18)	Livello di pressione acustica	250 Hz	114 dB, 124 dB	0,10 dB	IEC 60942:2017 Annex B CEI EN IEC 60942:2018 Annex B	
	Frequenza	124 dB	250 Hz	0,012 %		
Calibratori	Livello di pressione acustica	250 Hz, 1 kHz	da 94 dB a 114 dB	0,10 dB		
	Frequenza	da 94 dB a 114 dB	250 Hz, 1 kHz	0,012 %		

(continua)

(Continua) Area metrologica "Misure acustiche" – Settore "Livello di pressione acustica" (SAU-01)

Strumento Instrument	Misurando Measurand	Condizioni Additional parameters	Campo di misura Measurement range	Incertezza Uncertainty	Metodo/Procedura Method / Procedure	Sede Location
Fonometri	Livello di pressione acustica	da 31,5 Hz a 16 kHz	da 25 dB a 140 dB	da 0,11 dB a 0,99 dB	Guida CEI 29-30:1997	A
		da 63 Hz a 16 kHz	da 25 dB a 140 dB	da 0,12 dB a 0,30 dB	Metodo interno. Taratura basata su IEC 61672-3:2006 e CEI EN 61672-3:2007	
		da 63 Hz a 16 kHz	da 25 dB a 140 dB	da 0,10 dB a 0,30 dB	IEC 61672-3:2013 CEI EN 61672-3:2014	



3. NATA (Austrálie)

3.1. Noise and Vibration Measurement Systems Pty Ltd, Sydney Calib. Lab.

Acoustical	Acoustic calibrators	Distortion;	Comparison with a	AS/IEC 60942 -
	Vibration measuring systems	Acceleration sensitivity; Frequency response	Comparison measurement against reference transducer	By the methods of ISO 16063 part 21

SERVICE	PRODUCT	DETERMINANT	TECHNIQUE	PROCEDURE	LIMITATIONS
---------	---------	-------------	-----------	-----------	-------------

Capability

With Calibration and Measurement Capability of-

Conditioning amplifiers 0.001 V/ms²

Data acquisition systems / Analyser electrical tests 0.004 dB

DC voltage 0.1 mV

1% from 5 Hz to 10 Hz

0.8% at greater than 10 Hz to 2 kHz

1.2% at greater than 2 kHz to 7 kHz

2% at greater than 7 kHz to 10 kHz

Vibration metrology - Vibration standards and reference equipment	Vibration calibrators	Acceleration; Distortion; Frequency	Direct measurement by frequency analyser	By the methods of ISO 16063-21
--	-----------------------	---	--	--------------------------------

Capability

with Calibration and Measurement Capability of -

At a nominal frequency of 160 Hz

Acceleration

0.8%

Distortion

0.25%

Frequency

0.01 Hz

Vibration metrology - Vibration equipment	Accelerometers	Acceleration sensitivity	Comparison measurement against reference transducer	By the methods of ISO 16063 part 21
--	----------------	--------------------------	---	-------------------------------------

Capability

Level response from 0.1 m/s² for transducers with a sensitivity above 0.1 pC/m/s² and 0.4 mV/m/s²

Up to 300 ms⁻² for accelerometers up to 50 g at 160 Hz

With Calibration and Measurement Capability of-

1% from 5 Hz to 10 Hz

0.8% at greater than 10 Hz to 2 kHz

1.2% at greater than 2 kHz to 7 kHz

2% at greater than 7 kHz to 10 kHz



3.2. National Measurement Institute (NMI), Acoustics, Ultrasonics and Vibration Project, Lindfield

SERVICE	PRODUCT	DETERMINANT	TECHNIQUE	PROCEDURE	LIMITATIONS
Acoustical metrology - Acoustic standards and reference equipment	Acoustic calibrators - Single frequency	Distortion; Frequency; Short term fluctuations; Sound pressure level	Comparison with a reference standard; Direct measurement against reference signal; Direct measurement by electrical input; Direct measurement by frequency analyser	IEC60942	
<p>Capability With Calibration and Measurement Capability of- 0.11 dB at nominal 1 kHz for sound pressure level 0.02 dB for short term level fluctuations 0.10 Hz for frequency 0.21% for total harmonic distortion and noise</p>					
	Sound level meters	Sound pressure level	Direct measurement against a reference instrument; Direct measurement by electrical input; Direct	AS 1259.1 - 1990 Clauses 8.9, 8.10, 10.1,10.2.2, 10.2.3, 10.3.2, 10.3.3, 10.3.4, 10.4.2,	

16.10.2025

Acoustic Calibration Laboratory - Accredited Organisation (Si... <https://nata.com.au/accredited-organisation/national-acoustic-calibr>

SERVICE	PRODUCT	DETERMINANT	TECHNIQUE	PROCEDURE	LIMITATIONS
			measurement by reference sound source; Pressure field comparison measurement	10.4.3, 10.4.4, 10.4.5 AS 1259.2 - 1990 Clauses 9.1, 9.2.1, 9.2.2, 9.2.3, 9.3.1, 9.3.2, 9.3.3, 9.3.5 IEC 60651:2001, IEC 60804:2000, IEC 61672-3	

Capability

With Calibration and Measurement Capability of-
 Acoustic tests using a multifunction acoustic calibrator

0.14 dB at 31.5 Hz

0.12 dB at 63 Hz

0.12 dB at 125 Hz to 500 Hz

0.11 dB at 1 kHz

0.13 dB at 2 kHz

0.14 dB at 4 kHz

0.15 dB at 8 kHz

0.21 dB at 12.5 kHz

0.37 dB at 16 kHz

Electrical tests

0.18 dB for frequency & time weighting

0.20 dB for level linearity

0.16 dB for toneburst response, peak C sound level & overload indication

0.09 dB for self-generated noise & high-level stability



3.3. ACU-VIB Electronics

SERVICE	PRODUCT	DETERMINANT	TECHNIQUE	PROCEDURE	LIMITATIONS
Acoustical metrology - Acoustic equipment	Band pass filters	Attenuation; Linearity	Direct measurement by electrical input	By the methods of AS/NZS 4476-1997, IEC 1260-1995 Clauses 5.3, 5.7, 5.8, 5.9; IEC 61260-3 and ACU-VIB procedure AVP06	
<p>Capability with Calibration and Measurement Capability of - For direct measurement of attenuation, A, (electronic filter) via voltage output to voltage input - 0.1 dB $-6 \text{ dB} \leq A \leq 40 \text{ dB}$ 0.3 dB $40 \text{ dB} < A \leq 80 \text{ dB}$ 0.11 dB for linearity 0.1 dB for mid-band relative attenuation 0.25 dB for lower limit of linearity range (noise)</p> <p>For recording of displayed attenuation (measuring instrument) -</p>					
	Microphones	Frequency response; Pressure sensitivity	Direct measurement against a reference instrument	Calibration by electrostatic actuator by the method of ACU-VIB Procedure AVP01; Calibration by pressure response in a small volume coupler by the method of ACU-VIB Procedure AVP11; Calibration by measurement of the open circuit sensitivity by the method of ACU-VIB procedure AVP04	
<p>Capability with Calibration and Measurement Capability of - Calibration by electrostatic actuator: 0.10 dB in the range 20 Hz to 20 kHz by electrostatic actuator</p> <p>Calibration by small volume coupler: 0.12 dB at 31.5 Hz, 63 Hz, 4 kHz and at 8 kHz 0.11 dB at 125 Hz and up to and including 2 kHz 0.08 dB at 1 kHz as the reference frequency 0.19 dB at 12.5 kHz</p>					



Dose meters for
sound

Electrical
response;
Frequency
response; Noise
exposure/dose

Direct
measurement by
electrical input;
Pressure field
comparison
measurement

AS/NZS 2399-1998
Clauses 6, 7, 8, 9, 10
and 11, IEC
61252 and ACU-VIB
procedure AVP07;
C weighted peak
according to IEC
61672-3 Clause19;
IEC 61672

Capability

Nominal octave frequencies from 63 Hz to 8 kHz;
with Calibration and Measurement Capability of -
0.03 Pa²h at a dose of 8 Pa²h (94 dBA 8h)
Acoustic tests in a small volume coupler at nominal octave frequencies
with Calibration and Measurement Capability of -
0.10 dB for frequencies from 63 Hz to below 125 Hz
0.09 dB for frequencies from 125 Hz to below 2 kHz
0.10 dB for frequencies from 2 kHz to below 4 kHz
0.11 dB for frequencies from 4 kHz to 8 kHz
0.09 dB at a reference frequency of 1 kHz

Signal
conditioning
amplifiers

Attenuation;
Frequency
response;
Linearity

Direct
measurement by
electrical input

By comparison with
calibrated
attenuator and
voltmeter by ACU-
VIB procedure
AVP09

Capability

At nominal frequencies from 10 Hz to 20 kHz and; Gain from 40 dB to -80 dB
Calibration of transducer preamplifiers and signal conditioning amplifiers
for Gain, Attenuation and Frequency response
with Calibration and Measurement Capability of -
0.05 dB for Gain from 40 dB to -40 dB;
0.08 dB for Gain from -40 dB to -80 dB;
0.01 dB for frequency response at nominal frequencies from 10 Hz to 20 kHz

Acoustical
metrology -
Acoustic
standards
and
reference
equipment

Acoustic
calibrators - Multi
frequency;
Acoustic
calibrators - Single
frequency

Distortion;
Frequency; Sound
pressure level

Comparison with
a reference
standard

By the methods of
IEC 60942 and
ACU-VIB procedure
AVP02 and AVP03

Capability

At nominal octave frequencies between 31.5 Hz and 16 kHz
with Calibration and Measurement Capability at nominal octave frequencies of -
0.10 dB for sound level pressure at 31.5 Hz
0.09 dB for sound level pressure at 63 Hz
0.08 dB for sound level pressure at 125 Hz up to 2 kHz
0.07 dB for sound level pressure at reference frequency of 1 kHz
0.11 dB for sound level pressure at 4 kHz to below 8 kHz
0.13 dB for sound level pressure at 8 kHz to below 12.5 kHz
0.18 dB for sound level pressure at 12.5 kHz to below 16 kHz
0.27 dB for sound level pressure at 16 kHz
0.05% for frequency
0.2% for total harmonic distortion and noise



Sound level meters	Electrical response; Sound pressure level	Direct measurement by electrical input; Pressure field comparison measurement	AS 1259.1 - 1990 Clauses 8.9, 8.10, 10.2.2, 10.2.3, 10.3.2, 10.3.3, 10.3.4, 10.4.1, 10.4.2, 10.4.3, 10.4.4 and 10.4.5; AS 1259.2 - 1990 Clauses 9.1, 9.2.1, 9.2.2, 9.2.3, 9.3.1, 9.3.2, 9.3.3, 9.3.4, 9.3.5; ACU-VIB procedure AVP05; IEC 60651, IEC 60804; IEC 61672-3; ACU-VIB procedure AVP10;
--------------------	---	---	--

Capability

From 31.5 Hz to 16 kHz for acoustic tests

From 20 Hz to 20 kHz for electrical tests

with Calibration and Measurement Capability of -

Electrical tests

0.1 dB (AS 1259-1, IEC 60651, AS 1259-2, IEC 60804)

0.09 dB (IEC 61672-3)

Acoustic tests in a small volume coupler at nominal octave frequencies

0.11 dB for frequencies from 31.5 Hz to below 63 Hz

0.10 dB for frequencies from 63 Hz to below 125 Hz

0.09 dB for frequencies from 125 Hz to below 2 kHz

0.10 dB for frequencies from 2 kHz to below 4 kHz

0.11 dB for frequencies from 4 kHz to below 8 kHz

0.16 dB for frequencies from 8 kHz to below 12.5 kHz

Vibration metrology - Vibration equipment	Accelerometers; Velocity transducers	Acceleration sensitivity; Frequency response; Velocity sensitivity	Comparison measurement against reference transducer	ISO 16063-21; ACU VIB Procedures AVP08, AVP12, AVP14, AVP15, AVP19	Maximum force of up to 1000 N or Displacement limit of 158 mm peak to peak
---	--------------------------------------	--	---	--	--

Capability

In the acceleration range (or equivalent) 0.01 m/s² up to 10 m/s² below 10 Hz,

up to 50 m/s² from 10 Hz,

up to 100 m/s² from 40 Hz,

up to 500 m/s² from 160 Hz,

Extended lower velocity range down to 0.1 mm/s from 0.5 Hz to below 16 Hz

For transducers of

Mass up to 5.0 kg at nominal frequencies from 0.5 Hz to 320 Hz;

Mass up to 500 grams for frequencies of 5kHz and below;

Mass up to 50 grams for frequencies of 10 kHz and below,

by comparison to the methods of ISO 16063-21

with calibration and measurement capability for sensitivity of

1.4% from 0.8 Hz to less than 1.6 Hz

0.6% from 1.6 Hz to less than 2500 Hz

1.0% from 2500 Hz to less than 5000 Hz

1.5% from 5000 Hz to less than 7000 Hz

2.3% from 7000 Hz up to 10000 Hz



Český institut pro akreditaci, o.p.s.

Hájkova 2747/22, 130 00 Praha 3

Strana: 16/62

Zpracoval:

Ing. Martin Valenta

Datum zpracování:

20.10.2025

systems; Vibration
recording systems

Electrical
response;
Frequency
response;
Velocity
sensitivity

against
reference
transducer;
Direct
measurement by
electrical input

Procedure AVP12,
AVP14, AVP15 and
AVP19

1000 N or
Displacement
limit of 158
mm peak to
peak

Capability

In the acceleration range (or equivalent) 0.01 m/s² up to 10 m/s² below 10 Hz,
up to 50 m/s² from 10 Hz,
up to 100 m/s² from 40 Hz,
up to 500 m/s² from 160 Hz,

Extended lower velocity range down to 0.1 mm/s from 0.5 Hz to below 16 Hz

For transducers of

Mass up to 5.0 kg at nominal frequencies from 0.5 Hz to 320 Hz;

Mass up to 500 grams for frequencies of 5kHz and below;

Mass up to 50 grams for frequencies of 10 kHz and below,

by comparison to the methods of ISO 16063-21 or by methods of ISO 8041

with calibration and measurement capability for level response of

1.4% from 0.8 Hz to less than 1.6 Hz

0.6% from 1.6 Hz to less than 2500 Hz

1.0% from 2500 Hz to less than 5000 Hz

1.5% from 5000 Hz to less than 7000 Hz

2.3% from 7000Hz up to 10000 Hz

Electrical tests

0.1% display linearity

0.25% Vector addition (PPV)

Vibration
metrology -
Vibration
standards
and
reference
equipment

Vibration
calibrators

Acceleration;
Distortion;
Frequency

Comparison
measurement
against
reference
transducer

ACU-VIB Procedure
AVP13, ISO
16063-21, ISO
16063-44, ISO 8041

Capability

For frequencies between 7 Hz and 10 kHz at levels between 1 ms⁻² and 100 ms⁻²

with a Calibration and Measurement Capability for acceleration at nominal frequencies of

1.2% from 7 Hz to below 16 Hz

1% from 16 Hz to below 80 Hz

0.6% from 80 Hz to 2 kHz

1% from 2 kHz to 10 kHz

with a Calibration and Measurement Capability of -

0.02 Hz for frequency

0.2% for total harmonic distortion and noise